

LB127698


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

49157



A= 0.0000e+000

B= 2.0051e-004

C= -8.4509e-002

Rho= 0.9988847

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	-0.015	-0.015	345	0.000	345	0				0/1)
0.2	0.200	0.191	-0.009	1374	0.0 %	1374					1
2.5	2.500	2.323	-0.177	12008	0.0 %	12008					5
5.0	5.000	5.271	0.271	26708	0.0 %	26708					5
7.5	7.500	7.659	0.159	38619	0.0 %	38619					2
10.0	10.000	9.772	-0.228	49157	0.0 %	49157					2

LB127698 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 10 Oct 2023 10:58:20

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	345	-	0.0000	7470A	PPB	10 Oct 2023 11:06:06	S	Std
0.2 - 1	56.2	1374	-	0.2000	7470A	PPB	10 Oct 2023 11:08:23	S	Std
2.5 - 1	52.5	12008	-	2.5000	7470A	PPB	10 Oct 2023 11:10:40	S	Std
5.0 - 1	55	26708	-	5.0000	7470A	PPB	10 Oct 2023 11:12:56	S	Std
7.5 - 1	57.5	38619	-	7.5000	7470A	PPB	10 Oct 2023 11:15:12	S	Std
10.0 - 1	510	49157	-	10.0000	7470A	PPB	10 Oct 2023 11:17:27	S	Std
ICV06 - 1	ICV06	20086	3.9429	-	7470A	PPB	10 Oct 2023 11:20:19	U	SMPL
ICB06 - 1	ICB06	212	-0.0420	-	7470A	PPB	10 Oct 2023 11:22:34	U	SMPL
CCV55 - 1	CCV55	26371	5.2030	-	7470A	PPB	10 Oct 2023 11:24:48	U	SMPL
CCB55 - 1	CCB55	218	-0.0408	-	7470A	PPB	10 Oct 2023 11:27:03	U	SMPL
CRA - 1	CRA	1319	0.1800	-	7470A	PPB	10 Oct 2023 11:29:18	U	SMPL
HighStd - 1	HighStd	49016	9.7435	-	7470A	PPB	10 Oct 2023 11:31:34	U	SMPL
ChkStd - 1	ChkStd	35522	7.0379	-	7470A	PPB	10 Oct 2023 11:33:48	U	SMPL
PB156268BL - 1	PBW	76	-0.0693	-	7470A	PPB	10 Oct 2023 11:36:04	U	SMPL
PB156268BS - 1	LCSW	17200	3.3642	-	7470A	PPB	10 Oct 2023 11:38:21	U	SMPL
O4770-02 - 1	1065	686	0.0530	-	7470A	PPB	10 Oct 2023 11:40:37	U	SMPL
O4770-04 - 1	1057	5165	0.9511	-	7470A	PPB	10 Oct 2023 11:42:53	U	SMPL
O4770-06 - 1	1058	4992	0.9164	-	7470A	PPB	10 Oct 2023 11:45:10	U	SMPL
O4770-06DUP - 1	1058DUP	5130	0.9441	-	7470A	PPB	10 Oct 2023 11:47:30	U	SMPL
O4770-06MS - 1	1058MS	21739	4.2743	-	7470A	PPB	10 Oct 2023 11:49:46	U	SMPL
CCV56 - 1	CCV56	25760	5.0805	-	7470A	PPB	10 Oct 2023 11:52:02	U	SMPL
CCB56 - 1	CCB56	197	-0.0450	-	7470A	PPB	10 Oct 2023 11:54:17	U	SMPL
O4770-06MSD - 1	1058MSD	21906	4.3078	-	7470A	PPB	10 Oct 2023 11:56:33	U	SMPL
O4778-10 - 1	NBS-GROUNDWATER	117	-0.0610	-	7470A	PPB	10 Oct 2023 11:58:48	U	SMPL
O4770-06LX5 - 1		1304	0.1770	-	7470A	PPB	10 Oct 2023 12:01:04	U	SMPL
O4770-06A - 1		26557	5.2403	-	7470A	PPB	10 Oct 2023 12:03:20	U	SMPL
CCV57 - 1	CCV57	25684	5.0653	-	7470A	PPB	10 Oct 2023 12:05:36	U	SMPL
CCB57 - 1	CCB57	198	-0.0448	-	7470A	PPB	10 Oct 2023 12:07:52	U	SMPL