



Analytical Summary Report

Analysis Method: 9040C

Parameter: Corrosivity

Run Number: LB127700

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	10/10/2023	09:10
2	CAL2	1	Water	NA	NA	20.6	7.01	10/10/2023	09:11
3	CAL3	1	Water	NA	NA	20.6	10.02	10/10/2023	09:15
4	ICV	1	Water	NA	NA	20.6	7.00	10/10/2023	09:20
5	CCV1	1	Water	NA	NA	20.7	2.01	10/10/2023	09:50
6	O4818-04	1	Water	NA	NA	20.6	7.10	10/10/2023	10:00
7	O4818-04DUP	1	Water	NA	NA	20.7	7.11	10/10/2023	10:01
8	O4818-05	1	Water	NA	NA	20.7	10.20	10/10/2023	10:10
9	CCV2	1	Water	NA	NA	20.6	12.02	10/10/2023	10:11

127700

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity 04818 WATER WorkList ID : 174430 Department : Wet-Chemistry Date : 10-10-2023 08:57:45

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4818-04	5881	Water	Corrosivity	Cool 4 deg C	PSEG03	E31	10/09/2023	9040C
O4818-05	5891	Water	Corrosivity	Cool 4 deg C	PSEG03	E31	10/09/2023	9040C

Date/Time 10-10-23 0911.00
Raw Sample Received by: JH (WC)
Raw Sample Relinquished by: JH SM

Date/Time 10-10-23 1211.00
Raw Sample Received by: JH SM
Raw Sample Relinquished by: JH (WC)