

LB 127708


245.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

42302



A= 0.0000e+000

B= 2.3626e-004

C= -1.7869e-002

Rho= 0.9999715

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	#/p)
0.0	0.000	0.019	0.019	157	0.000	157					
0.05	0.050	0.063	0.013	342	0.0 %	342	0				26
0.2	0.200	0.186	-0.014	862	0.0 %	862					-7
2.5	2.500	2.449	-0.051	10441	0.0 %	10441					-2
5.0	5.000	5.026	0.026	21349	0.0 %	21349					1
7.5	7.500	7.530	0.030	31949	0.0 %	31949					0
10.0	10.000	9.977	-0.023	42302	0.0 %	42302					0

LB127708 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 10 Oct 2023 12:40:24

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	157	-	0.0000	245.1	PPB	10 Oct 2023 13:06:06	S	Std
0.05 - 1	50.05	342	-	0.0500	245.1	PPB	10 Oct 2023 13:13:33	S	Std
0.2 - 1	50.2	862	-	0.2000	245.1	PPB	10 Oct 2023 13:15:50	S	Std
2.5 - 1	52.5	10441	-	2.5000	245.1	PPB	10 Oct 2023 13:18:06	S	Std
5.0 - 1	55	21349	-	5.0000	245.1	PPB	10 Oct 2023 13:20:22	S	Std
7.5 - 1	57.5	31949	-	7.5000	245.1	PPB	10 Oct 2023 13:22:38	S	Std
10.0 - 1	510	42302	-	10.0000	245.1	PPB	10 Oct 2023 13:31:24	S	Std
ICV07 - 1	ICV07	16782	3.9471	-245.1		PPB	10 Oct 2023 13:34:15	U	SMPL
ICB07 - 1	ICB07	-21	-0.0228	-245.1		PPB	10 Oct 2023 13:36:30	U	SMPL
CCV58 - 1	CCV58	21666	5.1010	-245.1		PPB	10 Oct 2023 13:38:45	U	SMPL
CCB58 - 1	CCB58	-12	-0.0207	-245.1		PPB	10 Oct 2023 13:41:02	U	SMPL
CRA - 1	CRA	296	0.0521	-245.1		PPB	10 Oct 2023 13:43:19	U	SMPL
HighStd - 1	HighStd	42790	10.0918	-245.1		PPB	10 Oct 2023 13:45:34	U	SMPL
ChkStd - 1	ChkStd	29657	6.9890	-245.1		PPB	10 Oct 2023 13:47:49	U	SMPL
PB156271BL - 1	PBW	-67	-0.0337	-245.1		PPB	10 Oct 2023 13:50:08	U	SMPL
PB156271BS - 1	LCSW	18468	4.3454	-245.1		PPB	10 Oct 2023 13:52:25	U	SMPL
O4717-01 - 1	402	272	0.0464	-245.1		PPB	10 Oct 2023 13:54:41	U	SMPL
O4717-01DUP - 1	402DUP	481	0.0958	-245.1		PPB	10 Oct 2023 13:56:57	U	SMPL
O4717-01MS - 1	402MS	18469	4.3457	-245.1		PPB	10 Oct 2023 13:59:13	U	SMPL
O4717-01MSD - 1	402MSD	20134	4.7390	-245.1		PPB	10 Oct 2023 14:01:31	U	SMPL
CCV59 - 1	CCV59	22502	5.2985	-245.1		PPB	10 Oct 2023 14:03:46	U	SMPL
CCB59 - 1	CCB59	-2	-0.0183	-245.1		PPB	10 Oct 2023 14:06:02	U	SMPL
O4717-01LX5 - 1		41	-0.0082	-245.1		PPB	10 Oct 2023 14:08:17	U	SMPL
O4717-01A - 1		16926	3.9811	-245.1		PPB	10 Oct 2023 14:10:32	U	SMPL
CCV60 - 1	CCV60	22218	5.2314	-245.1		PPB	10 Oct 2023 14:12:47	U	SMPL
CCB60 - 1	CCB60	-33	-0.0257	-245.1		PPB	10 Oct 2023 14:15:02	U	SMPL