



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB127710

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	10/10/2023	12:50
2	CAL2	1	Water	NA	NA	20.6	7.00	10/10/2023	12:51
3	CAL3	1	Water	NA	NA	20.6	10.02	10/10/2023	12:53
4	ICV	1	Water	NA	NA	20.6	7.00	10/10/2023	13:00
5	CCV1	1	Water	NA	NA	20.7	2.01	10/10/2023	13:05
6	O4821-01	1	Water	NA	NA	20.6	4.89	10/10/2023	13:10
7	O4821-01DUP	1	Water	NA	NA	20.7	4.90	10/10/2023	13:11
8	O4822-01	1	Water	NA	NA	20.1	5.00	10/10/2023	13:15
9	O4822-02	1	Water	NA	NA	20.1	4.91	10/10/2023	13:20
10	O4823-02	1	Water	NA	NA	20.4	5.43	10/10/2023	13:25
11	O4823-04	1	Water	NA	NA	20.5	5.49	10/10/2023	13:30
12	O4823-05	1	Water	NA	NA	20.6	5.95	10/10/2023	13:35
13	O4823-06	1	Water	NA	NA	20.1	5.82	10/10/2023	13:40
14	CCV2	1	Water	NA	NA	20.6	12.02	10/10/2023	13:45

WORKLIST(Hardcopy Internal Chain)

UB 127710

WorkList Name : ph 04823

WorkList ID : 174439

Department : Wet-Chemistry

Date : 10-10-2023 12:38:42

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4821-01	A CELL-B/C-POOL-WATER	Water	pH	Cool 4 deg C	STER07	E11	10/06/2023	9040C
O4822-01	A MAIN-CELL	Water	pH	Cool 4 deg C	STER16	E11	10/03/2023	9040C
O4822-02	A EX-CELL	Water	pH	Cool 4 deg C	STER16	E11	10/03/2023	9040C
O4823-02	A OILY-LIQUID	Water	pH	Cool 4 deg C	PSEG03	E31	10/10/2023	9040C
O4823-04	C TOTES-1	Water	pH	Cool 4 deg C	PSEG03	E31	10/10/2023	9040C
O4823-05	C TOTES-2	Water	pH	Cool 4 deg C	PSEG03	E31	10/10/2023	9040C
O4823-06	C TOTES-3	Water	pH	Cool 4 deg C	PSEG03	E31	10/10/2023	9040C

Date/Time

10.10.23 12:45

Raw Sample Received by:

JO (Wg)

Raw Sample Relinquished by:

JO (Wg)

Date/Time

10-10-23

Raw Sample Received by:

JO (Wg)

Raw Sample Relinquished by:

JO (Wg)