



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB127711

Slope : 97.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	10/10/2023	12:50
2	CAL2	1	Water	NA	NA	20.6	7.00	10/10/2023	12:51
3	CAL3	1	Water	NA	NA	20.6	10.02	10/10/2023	12:53
4	ICV	1	Water	NA	NA	20.6	7.00	10/10/2023	13:00
5	CCV1	1	Water	NA	NA	20.6	2.01	10/10/2023	13:50
6	O4823-01	1	Solid	20.02	20	20.4	6.11	10/10/2023	14:00
7	O4823-01DUP	1	Solid	20.02	20	20.5	6.12	10/10/2023	14:01
8	CCV2	1	Water	NA	NA	20.6	12.02	10/10/2023	14:10

WORKLIST(Hardcopy Internal Chain)

VB127711

WorkList Name : ph 04823 s

WorkList ID : 174441

Department : Wet-Chemistry

Date : 10-10-2023 13:02:44

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4823-01	+	OILY-DEBRIS-SOLID	Solid	pH	Cool 4 deg C	PSEG03 E31	10/10/2023	9045D

Date/Time 10-10-23 13:10

Raw Sample Received by: JH WY

Raw Sample Relinquished by: JH WY

Date/Time

10-10-23

16:00

Raw Sample Received by:

JH WY

Raw Sample Relinquished by:

JH WY