

L3127713


7471B

INSTRUMENT ID: CV2

Linear

μ Abs.:

7021E



A= 0.0000e+000

B= 1.4297e-004

C= -2.4862e-002

Rho= 0.9998833

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	(%)
0.0	0.000	0.013	0.013	262	0.000	262					-
0.2	0.200	0.193	-0.007	1521	0.0 %	1521					-4
2.5	2.500	2.561	0.061	18088	0.0 %	18088					2
5.0	5.000	4.883	-0.117	34330	0.0 %	34330					2
7.5	7.500	7.536	0.036	52883	0.0 %	52883					0
10.0	10.000	10.014	0.014	70218	0.0 %	70218					0

LB127713 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 10 Oct 2023 14:26:15

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	262	-	0.0000	7471B	PPB	10 Oct 2023 14:27:15	S	Std
0.2 - 1	50.2	1521	-	0.2000	7471B	PPB	10 Oct 2023 14:29:32	S	Std
2.5 - 1	52.5	18088	-	2.5000	7471B	PPB	10 Oct 2023 14:31:49	S	Std
5.0 - 1	55	34330	-	5.0000	7471B	PPB	10 Oct 2023 14:34:07	S	Std
7.5 - 1	57.5	52883	-	7.5000	7471B	PPB	10 Oct 2023 14:36:23	S	Std
10.0 - 1	51.4	70218	-	10.0000	7471B	PPB	10 Oct 2023 14:38:40	S	Std
ICV08 - 1	ICV08	28164	4.0018	-	7471B	PPB	10 Oct 2023 14:41:40	U	SMPL
ICB08 - 1	ICB08	141	-0.0047	-	7471B	PPB	10 Oct 2023 14:43:55	U	SMPL
CCV61 - 1	CCV61	34242	4.8708	-	7471B	PPB	10 Oct 2023 14:46:10	U	SMPL
CCB61 - 1	CCB61	140	-0.0048	-	7471B	PPB	10 Oct 2023 14:48:25	U	SMPL
CRA - 1	CRA	1483	0.1872	-	7471B	PPB	10 Oct 2023 14:50:41	U	SMPL
HighStd - 1	HighStd	69748	9.9471	-	7471B	PPB	10 Oct 2023 14:52:56	U	SMPL
ChkStd - 1	ChkStd	49184	7.0071	-	7471B	PPB	10 Oct 2023 14:55:10	U	SMPL
PB156278BL - 1	PBS	88	-0.0123	-	7471B	PPB	10 Oct 2023 14:57:30	U	SMPL
PB156278BS - 1	LCSS	31426	4.4682	-	7471B	PPB	10 Oct 2023 14:59:48	U	SMPL
O4791-01 - 1	TP-3	2492	0.3314	-	7471B	PPB	10 Oct 2023 15:02:04	U	SMPL
O4800-01 - 1	EO-01-100623	7243	1.0107	-	7471B	PPB	10 Oct 2023 15:04:21	U	SMPL
O4801-01 - 1	HR-2-100623	1744	0.2245	-	7471B	PPB	10 Oct 2023 15:06:39	U	SMPL
O4801-03 - 1	HR-4-100623	870	0.0995	-	7471B	PPB	10 Oct 2023 15:08:59	U	SMPL
O4804-01 - 1	HD-02-100623	1176	0.1433	-	7471B	PPB	10 Oct 2023 15:11:15	U	SMPL
CCV62 - 1	CCV62	33939	4.8275	-	7471B	PPB	10 Oct 2023 15:13:30	U	SMPL
CCB62 - 1	CCB62	187	0.0019	-	7471B	PPB	10 Oct 2023 15:15:45	U	SMPL
O4805-01 - 1	B-1(0-5)	32522	4.6249	-	7471B	PPB	10 Oct 2023 15:18:01	U	SMPL
O4805-01DUP - 1	B-1(0-5)DUP	35153	5.0010	-	7471B	PPB	10 Oct 2023 15:25:42	U	SMPL
O4805-02 - 1	O4805-01MS	66721	9.5144	-	7471B	PPB	10 Oct 2023 15:27:56	U	SMPL
O4805-03 - 1	O4805-01MSD	69801	9.9547	-	7471B	PPB	10 Oct 2023 15:30:13	U	SMPL
O4805-04 - 1	H-1(0-5)	38419	5.4680	-	7471B	PPB	10 Oct 2023 15:32:29	U	SMPL
O4805-05 - 1	M-2(0-5)	23426	3.3244	-	7471B	PPB	10 Oct 2023 15:34:47	U	SMPL
O4805-06 - 1	M-1(0-5)	30212	4.2946	-	7471B	PPB	10 Oct 2023 15:37:04	U	SMPL
O4805-07 - 1	DUP-02	11165	1.5714	-	7471B	PPB	10 Oct 2023 15:39:21	U	SMPL
O4808-01 - 1	SU-03-100623	2666	0.3563	-	7471B	PPB	10 Oct 2023 15:41:38	U	SMPL
O4812-01 - 1	VNJ-223	6019	0.8357	-	7471B	PPB	10 Oct 2023 15:43:53	U	SMPL
CCV63 - 1	CCV63	34578	4.9188	-	7471B	PPB	10 Oct 2023 15:46:08	U	SMPL
CCB63 - 1	CCB63	213	0.0056	-	7471B	PPB	10 Oct 2023 15:48:23	U	SMPL
O4817-01 - 1	72-11930	2755	0.3690	-	7471B	PPB	10 Oct 2023 15:50:41	U	SMPL
O4817-04 - 1	VNJ-241	2240	0.2954	-	7471B	PPB	10 Oct 2023 15:52:56	U	SMPL
O4805-01LX5 - 1		7309	1.0201	-	7471B	PPB	10 Oct 2023 15:55:12	U	SMPL
O4805-01A - 1		61439	8.7592	-	7471B	PPB	10 Oct 2023 15:57:28	U	SMPL
CCV64 - 1	CCV64	34474	4.9039	-	7471B	PPB	10 Oct 2023 16:02:07	U	SMPL
CCB64 - 1	CCB64	222	0.0069	-	7471B	PPB	10 Oct 2023 16:04:24	U	SMPL