

LB127716


SFAM01.1

INSTRUMENT ID: CV2

Linear

μ Abs.:

56340



A= 0.0000e+000

B= 1.7887e-004 510%

C= -1.5699e-002 $r^2 = 0.9999$

Rho= 0.9999224

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	(%)
0.00	0.000	-0.002	-0.002	75	0.000	75					2.2
0.05	0.050	0.061	0.011	428	0.0 %	428	0				4
0.20	0.200	0.208	0.008	1253	0.0 %	1253					0
2.50	2.500	2.495	-0.005	14037	0.0 %	14037					1
5.00	5.000	5.027	0.027	28195	0.0 %	28195					1
7.50	7.500	7.399	-0.101	41454	0.0 %	41454					1
10.0	10.000	10.062	0.062	56340	0.0 %	56340					1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	75	-		0 SFAM01.1	PPB	10/10/2023 16:25	Std
	0.05 S01	428	-		0.05 SFAM01.1	PPB	10/10/2023 16:27	Std
	0.2 S02	1253	-		0.2 SFAM01.1	PPB	10/10/2023 16:29	Std
	2.5 S03	14037	-		2.5 SFAM01.1	PPB	10/10/2023 16:34	Std
	5 S04	28195	-		5 SFAM01.1	PPB	10/10/2023 16:36	Std
	7.5 S05	41454	-		7.5 SFAM01.1	PPB	10/10/2023 16:38	Std
	10 S06	56340	-		10 SFAM01.1	PPB	10/10/2023 16:46	Std
ICV009	ICV009	22815	4.0651 -		SFAM01.1	PPB	10/10/2023 16:49	SMPL
ICB009	ICB009	12	-0.0136 -		SFAM01.1	PPB	10/10/2023 16:51	SMPL
CCV065	CCV065	28856	5.1456 -		SFAM01.1	PPB	10/10/2023 16:58	SMPL
CCB065	CCB065	-30	-0.0211 -		SFAM01.1	PPB	10/10/2023 17:00	SMPL
PB156279BL	PBS279	0	-0.0157 -		SFAM01.1	PPB	10/10/2023 17:02	SMPL
O4565-01	MJKW93	689	0.1075 -		SFAM01.1	PPB	10/10/2023 17:04	SMPL
O4565-02	MJKW94	556	0.0838 -		SFAM01.1	PPB	10/10/2023 17:07	SMPL
O4565-03	MJKW95	395	0.055 -		SFAM01.1	PPB	10/10/2023 17:09	SMPL
O4565-04	MJKW96	432	0.0616 -		SFAM01.1	PPB	10/10/2023 17:11	SMPL
O4565-05	MJKW97	251	0.0292 -		SFAM01.1	PPB	10/10/2023 17:14	SMPL
O4565-06	MJKW98	382	0.0526 -		SFAM01.1	PPB	10/10/2023 17:16	SMPL
O4565-07	MJKW99	662	0.1027 -		SFAM01.1	PPB	10/10/2023 17:18	SMPL
O4565-08	MJKWA0	791	0.1258 -		SFAM01.1	PPB	10/10/2023 17:20	SMPL
O4565-09	MJKWA1	1122	0.185 -		SFAM01.1	PPB	10/10/2023 17:23	SMPL
O4565-10	MJKWA2	977	0.1591 -		SFAM01.1	PPB	10/10/2023 17:25	SMPL
O4565-11	MJKWA3	678	0.1056 -		SFAM01.1	PPB	10/10/2023 17:27	SMPL
O4565-12	MJKWA4	1067	0.1752 -		SFAM01.1	PPB	10/10/2023 17:29	SMPL
O4565-13	MJKWA5	567	0.0857 -		SFAM01.1	PPB	10/10/2023 17:32	SMPL
O4565-14	MJKWA6	642	0.0991 -		SFAM01.1	PPB	10/10/2023 17:34	SMPL
O4565-15	MJKWB7	7808	1.3809 -		SFAM01.1	PPB	10/10/2023 17:36	SMPL
O4565-16	MJKWB8	3084	0.5359 -		SFAM01.1	PPB	10/10/2023 17:38	SMPL
O4565-17	MJKWB9	1156	0.1911 -		SFAM01.1	PPB	10/10/2023 17:41	SMPL
O4565-18	MJKWB9D	756	0.1195 -		SFAM01.1	PPB	10/10/2023 17:43	SMPL
CCV066	CCV066	28665	5.1115 -		SFAM01.1	PPB	10/10/2023 17:48	SMPL
CCB066	CCB066	38	-0.0089 -		SFAM01.1	PPB	10/10/2023 17:50	SMPL
O4565-20	MJKWC0	561	0.0846 -		SFAM01.1	PPB	10/10/2023 17:52	SMPL
O4565-21	MJKWC1	576	0.0873 -		SFAM01.1	PPB	10/10/2023 17:54	SMPL
O4565-22	MJKWC3	525	0.0782 -		SFAM01.1	PPB	10/10/2023 17:57	SMPL
PB156280BL	PBS280	22	-0.0118 -		SFAM01.1	PPB	10/10/2023 17:59	SMPL
O4584-01	MJKWB4	1197	0.1984 -		SFAM01.1	PPB	10/10/2023 18:01	SMPL
O4584-02	MJKWB5	552	0.083 -		SFAM01.1	PPB	10/10/2023 18:03	SMPL
O4584-03	MJKWB6	1463	0.246 -		SFAM01.1	PPB	10/10/2023 18:06	SMPL
O4584-04	MJKWC4	716	0.1124 -		SFAM01.1	PPB	10/10/2023 18:08	SMPL
O4584-05	MJKWC5	1041	0.1705 -		SFAM01.1	PPB	10/10/2023 18:10	SMPL
O4584-06	MJKWC6	2526	0.4361 -		SFAM01.1	PPB	10/10/2023 18:12	SMPL
O4584-07	MJKWC7	1363	0.2281 -		SFAM01.1	PPB	10/10/2023 18:15	SMPL
O4584-08	MJKWC8	708	0.1109 -		SFAM01.1	PPB	10/10/2023 18:17	SMPL
O4584-09	MJKWC9	609	0.0932 -		SFAM01.1	PPB	10/10/2023 18:19	SMPL
O4584-10	MJKWD0	509	0.0753 -		SFAM01.1	PPB	10/10/2023 18:21	SMPL

INSTRUMENT ID : CV1

O4584-11	MJKWD1	1096	0.1803 -	SFAM01.1	PPB	10/10/2023 18:24 SMPL
O4584-12	MJKWD2	942	0.1528 -	SFAM01.1	PPB	10/10/2023 18:26 SMPL
O4584-13	MJKWD3	493	0.0725 -	SFAM01.1	PPB	10/10/2023 18:28 SMPL
O4584-14	MJKWD4	417	0.0589 -	SFAM01.1	PPB	10/10/2023 18:30 SMPL
O4584-15	MJKWD4D	622	0.0956 -	SFAM01.1	PPB	10/10/2023 18:33 SMPL
O4584-16	MJKWD4S	17033	3.0309 -	SFAM01.1	PPB	10/10/2023 18:35 SMPL
CCV067	CCV067	27840	4.9639 -	SFAM01.1	PPB	10/10/2023 18:37 SMPL
CCB067	CCB067	31	-0.0102 -	SFAM01.1	PPB	10/10/2023 18:40 SMPL
O4584-17	MJKWD5	3772	0.659 -	SFAM01.1	PPB	10/10/2023 18:42 SMPL
O4584-18	MJKWD6	672	0.1045 -	SFAM01.1	PPB	10/10/2023 18:44 SMPL
O4584-19	MJKWD7	546	0.082 -	SFAM01.1	PPB	10/10/2023 18:46 SMPL
O4584-20	MJKWD8	292	0.0365 -	SFAM01.1	PPB	10/10/2023 18:49 SMPL
O4565-19		16102	2.8644 -	SFAM01.1	PPB	10/10/2023 18:51 SMPL
CCV068	CCV068	28276	5.0419 -	SFAM01.1	PPB	10/10/2023 18:53 SMPL
CCB068	CCB068	23	-0.0116 -	SFAM01.1	PPB	10/10/2023 18:55 SMPL