


LB127734

7470A

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0 D
0.0	0.000	0.040	0.040	100	0.000	100					
0.2	0.200	0.208	0.008	912	0.0 %	912	0				4
2.5	2.500	2.485	-0.015	11885	0.0 %	11885					-1
5.0	5.000	4.842	-0.158	23243	0.0 %	23243					-3
7.5	7.500	7.696	0.196	37000	0.0 %	37000					3
10.0	10.000	9.929	-0.071	47761	0.0 %	47761					-1

LB127734

Method: 7470A

Operator: Admin

Date of Analysis: 11 Oct 2023 11:51:29

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	100	-	0.0000	7470A	PPB	11 Oct 2023 11:58:57	S	Std
0.2 - 1	50.2	912	-	0.2000	7470A	PPB	11 Oct 2023 12:01:14	S	Std
2.5 - 1	52.5	11885	-	2.5000	7470A	PPB	11 Oct 2023 12:03:31	S	Std
5.0 - 1	55	23243	-	5.0000	7470A	PPB	11 Oct 2023 12:05:48	S	Std
7.5 - 1	57.5	37000	-	7.5000	7470A	PPB	11 Oct 2023 12:08:04	S	Std
10.0 - 1	59	47761	-	10.0000	7470A	PPB	11 Oct 2023 12:10:20	S	Std
ICV10 - 1	ICV10	19018	3.9651	-	7470A	PPB	11 Oct 2023 12:13:01	U	SMPL
ICB10 - 1	ICB10	16	0.0226	-	7470A	PPB	11 Oct 2023 12:15:16	U	SMPL
CCV69 - 1	CCV69	24296	5.0602	-	7470A	PPB	11 Oct 2023 12:17:31	U	SMPL
CCB69 - 1	CCB69	22	0.0238	-	7470A	PPB	11 Oct 2023 12:19:45	U	SMPL
CRA - 1	CRA	890	0.2039	-	7470A	PPB	11 Oct 2023 12:22:00	U	SMPL
HighStd - 1	HighStd	46780	9.7251	-	7470A	PPB	11 Oct 2023 12:24:15	U	SMPL
ChkStd - 1	ChkStd	32921	6.8497	-	7470A	PPB	11 Oct 2023 12:26:30	U	SMPL
PB156281BL - 1	PBW	-54	0.0081	-	7470A	PPB	11 Oct 2023 12:28:46	U	SMPL
O4809-01 - 1	M-092023-103-01-A	56	0.0309	-	7470A	PPB	11 Oct 2023 12:33:18	U	SMPL
O4809-02 - 1	M-092023-103-02-A	64	0.0325	-	7470A	PPB	11 Oct 2023 12:35:34	U	SMPL
O4809-03 - 1	M-092023-103-01-B	81	0.0361	-	7470A	PPB	11 Oct 2023 12:37:51	U	SMPL
O4809-04 - 1	M-092023-103-02-B	67	0.0332	-	7470A	PPB	11 Oct 2023 12:40:08	U	SMPL
O4812-02 - 1	VNJ-223	48	0.0292	-	7470A	PPB	11 Oct 2023 12:42:23	U	SMPL
CCV70 - 1	CCV70	24926	5.1909	-	7470A	PPB	11 Oct 2023 12:44:38	U	SMPL
CCB70 - 1	CCB70	6	0.0205	-	7470A	PPB	11 Oct 2023 12:46:53	U	SMPL
O4813-01 - 1	SB-15	39	0.0274	-	7470A	PPB	11 Oct 2023 12:49:09	U	SMPL
O4813-02 - 1	SB-16	37	0.0269	-	7470A	PPB	11 Oct 2023 12:51:24	U	SMPL
O4813-03 - 1	SB-17	42	0.0280	-	7470A	PPB	11 Oct 2023 12:53:39	U	SMPL
O4813-04 - 1	SB-18	22	0.0238	-	7470A	PPB	11 Oct 2023 12:55:55	U	SMPL
O4813-05 - 1	SB-19	44	0.0284	-	7470A	PPB	11 Oct 2023 12:58:11	U	SMPL
O4813-06 - 1	SB-20	50	0.0296	-	7470A	PPB	11 Oct 2023 13:00:27	U	SMPL
O4813-07 - 1	S-9	42	0.0280	-	7470A	PPB	11 Oct 2023 13:02:43	U	SMPL
O4813-08 - 1	S-10	34	0.0263	-	7470A	PPB	11 Oct 2023 13:05:00	U	SMPL
O4815-01 - 1	CHRT-26333	37	0.0269	-	7470A	PPB	11 Oct 2023 13:07:17	U	SMPL
O4815-02 - 1	CHRT-20416	52	0.0301	-	7470A	PPB	11 Oct 2023 13:09:33	U	SMPL
CCV71 - 1	CCV71	24724	5.1490	-	7470A	PPB	11 Oct 2023 13:11:48	U	SMPL
CCB71 - 1	CCB71	9	0.0211	-	7470A	PPB	11 Oct 2023 13:14:04	U	SMPL
O4817-03 - 1	72-11930	34	0.0263	-	7470A	PPB	11 Oct 2023 13:16:19	U	SMPL
O4817-06 - 1	VNJ-241	34	0.0263	-	7470A	PPB	11 Oct 2023 13:18:35	U	SMPL
O4817-06DUP - 1	VNJ-241DUP	42	0.0280	-	7470A	PPB	11 Oct 2023 13:20:51	U	SMPL
O4817-06MS - 1	VNJ-241MS	18958	3.9527	-	7470A	PPB	11 Oct 2023 13:23:08	U	SMPL
O4817-06MSD - 1	VNJ-241MSD	20167	4.2035	-	7470A	PPB	11 Oct 2023 13:25:25	U	SMPL
PB156282BL - 1	PBW	-44	0.0101	-	7470A	PPB	11 Oct 2023 13:27:41	U	SMPL
PB156282BS - 1	LCSW	21293	4.4371	-	7470A	PPB	11 Oct 2023 13:29:56	U	SMPL
O4818-04 - 1	5881	-14	0.0164	-	7470A	PPB	11 Oct 2023 13:32:13	U	SMPL
O4818-04DUP - 1	5881DUP	56	0.0309	-	7470A	PPB	11 Oct 2023 13:34:30	U	SMPL
O4818-04MS - 1	5881MS	19529	4.0711	-	7470A	PPB	11 Oct 2023 13:36:46	U	SMPL
CCV72 - 1	CCV72	24753	5.1550	-	7470A	PPB	11 Oct 2023 13:39:01	U	SMPL
CCB72 - 1	CCB72	-51	0.0087	-	7470A	PPB	11 Oct 2023 13:41:17	U	SMPL
O4818-04MSD - 1	5881MSD	19149	3.9923	-	7470A	PPB	11 Oct 2023 13:43:32	U	SMPL
O4818-05 - 1	5891	63	0.0323	-	7470A	PPB	11 Oct 2023 13:45:47	U	SMPL
PB156249TB - 1	PB156249TB	2	0.0197	-	7470A	PPB	11 Oct 2023 13:48:03	U	SMPL
PB156263TB - 1	PB156263TB	35	0.0265	-	7470A	PPB	11 Oct 2023 13:50:18	U	SMPL
O4817-06LX5 - 1		35	0.0265	-	7470A	PPB	11 Oct 2023 13:52:34	U	SMPL
O4817-06A - 1		19183	3.9993	-	7470A	PPB	11 Oct 2023 13:54:50	U	SMPL
O4818-04LX5 - 1		-11	0.0170	-	7470A	PPB	11 Oct 2023 13:57:07	U	SMPL
O4818-04A - 1		19830	4.1336	-	7470A	PPB	11 Oct 2023 13:59:25	U	SMPL
PB156281BS - 1	LCSW	18399	3.8367	-	7470A	PPB	11 Oct 2023 14:01:42	U	SMPL
CCV73 - 1	CCV73	24368	5.0751	-	7470A	PPB	11 Oct 2023 14:03:57	U	SMPL
CCB73 - 1	CCB73	-44	0.0101	-	7470A	PPB	11 Oct 2023 14:06:12	U	SMPL