



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB127735

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	10/11/2023	09:25
2	CAL2	1	Water	NA	NA	20.3	6.99	10/11/2023	09:26
3	CAL3	1	Water	NA	NA	20.4	10.02	10/11/2023	09:30
4	ICV	1	Water	NA	NA	20.5	7.01	10/11/2023	09:35
5	CCV1	1	Water	NA	NA	20.6	2.01	10/11/2023	09:37
6	O4836-04	1	Water	NA	NA	20.6	4.72	10/11/2023	09:45
7	O4836-04DUP	1	Water	NA	NA	20.7	4.73	10/11/2023	09:46
8	CCV2	1	Water	NA	NA	20.6	12.02	10/11/2023	09:50

12-1-23
N3

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH O4836 WorkList ID : 174474 Department : Wet-Chemistry Date : 10-11-2023 09:11:36

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4836-04	DW-1	Water	pH	Cool 4 deg C	PSEG03	E11	10/10/2023	9040C

Date/Time 10/11/23 091220
Raw Sample Received by: 20 USC1
Raw Sample Relinquished by: 20 USC1

Date/Time 10-11-23 11:00
Raw Sample Received by: 20 USC1
Raw Sample Relinquished by: 20 USC1