



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB127736

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	10/11/2023	11:37
2	CAL2	1	Water	NA	NA	20.4	7.00	10/11/2023	11:38
3	CAL3	1	Water	NA	NA	20.4	10.02	10/11/2023	11:39
4	ICV	1	Water	NA	NA	20.7	7.01	10/11/2023	11:40
5	CCV1	1	Water	NA	NA	20.6	2.01	10/11/2023	11:45
6	O4813-09	1	Solid	20.02	20	20.7	6.15	10/11/2023	11:50
7	O4813-09DUP	1	Solid	20.03	20	20.8	6.16	10/11/2023	11:51
8	CCV2	1	Water	NA	NA	20.7	12.02	10/11/2023	12:00

VB127736

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH O4813

WorkList ID : 174482

Department : Wet-Chemistry

Date : 10-11-2023 11:19:05

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4813-09	50126-50127-COMP	Solid	pH	Cool 4 deg C	PSEG03	E31	10/09/2023	9045D

Date/Time

10-11-23

Raw Sample Received by:

10/10/23

Raw Sample Relinquished by:

10/10/23

Date/Time

10-11-23

Raw Sample Received by:

10/10/23

Raw Sample Relinquished by:

10/10/23