



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Iwona

Parameter: COD

SUPERVISOR REVIEW BY: Sohil

Run Number: LB127739

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP104201
COD calibration std. 100 ppm	WP104198
COD calibration std. 10 ppm	WP104196
COD calibration std. 150 ppm	WP104199
COD calibration std. 50 ppm	WP104197
COD calibration std. 0 ppm	WP104194
COD Digestion Vials Low Level 0-150Mg/L	W3026
COD ICV-LCS std, 50ppm	WP104612
COD CCV std, 50ppm	WP104611
COD LOD std, 5ppm	WP104613

Temp In (C): 149	Date In: 10/11/2023	Time In: 09:35
Temp Out (C): 151	Date Out: 10/11/2023	Time Out: 11:35

Intercept: 0.039

Slope: 0.9994

Regression: 0.9999

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	-0.039		09/15/2023	13:20
2	CAL2	10	1	Water	9.000	8.966	-10.3	09/15/2023	13:20
3	CAL3	50	1	Water	51.000	50.992	2	09/15/2023	13:21
4	CAL4	100	1	Water	101.000	101.022	1	09/15/2023	13:21
5	CAL5	150	1	Water	149.000	149.05	-0.6	09/15/2023	13:22



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Iwona

Parameter: COD

SUPERVISOR REVIEW BY: Sohil

Run Number: LB127739

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50	NA	NA	1	Water	51.000	50.992	09/15/2023	13:22
2	ICB		NA	NA	1	Water	0.000	-0.039	09/15/2023	13:23
3	CCV1	50	NA	NA	1	Water	49.000	48.990	10/11/2023	14:00
4	CCB1		NA	NA	1	Water	1.000	0.962	10/11/2023	14:00
5	LB127739BL		1	50	1	Solid	1.000	0.962	10/11/2023	14:01
6	LB127739BS	50	1	50	1	Solid	51.000	50.992	10/11/2023	14:01
7	O4699-03		1	50	1	Solid	6.000	5.965	10/11/2023	14:02
8	CCV2	50	NA	NA	1	Water	50.000	49.991	10/11/2023	14:02
9	CCB2		NA	NA	1	Water	1.000	0.962	10/11/2023	14:03