


CB127340

7471B

INSTRUMENT ID: CV1



A= 0.0000e+000

B= 1.4594e-004

C= 1.3728e-002

Rho= 0.9999424

Accept = Accepted

| Std ID | Conc.  | Calc. | Dev.   | Mean  | SD or %RSD | Rep 1 | Rep 2 | Rep 3 | Rep 4 | Rep 5 |  |
|--------|--------|-------|--------|-------|------------|-------|-------|-------|-------|-------|--|
| 0.0    | 0.000  | 0.022 | 0.022  | 55    | 0.000      | 55    |       |       |       |       |  |
| 0.2    | 0.200  | 0.204 | 0.004  | 1306  | 0.0 %      | 1306  |       |       |       |       |  |
| 2.5    | 2.500  | 2.450 | -0.050 | 16695 | 0.0 %      | 16695 |       |       |       |       |  |
| 5.0    | 5.000  | 4.989 | -0.011 | 34095 | 0.0 %      | 34095 |       |       |       |       |  |
| 7.5    | 7.500  | 7.571 | 0.071  | 51783 | 0.0 %      | 51783 |       |       |       |       |  |
| 10.0   | 10.000 | 9.964 | -0.036 | 68181 | 0.0 %      | 68181 |       |       |       |       |  |

9/01  
1  
2  
2  
0  
1  
0

## LB127740M INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 11 Oct 2023 14:32:36

| Sample ID        | Extended ID               | $\mu$ Abs | Conc    | Std Conc | Method | Units | Date                 | Type | Type |
|------------------|---------------------------|-----------|---------|----------|--------|-------|----------------------|------|------|
| 0.0 - 1          | 50                        | 55        | -       | 0.0000   | 7471B  | PPB   | 11 Oct 2023 14:35:37 | S    | Std  |
| 0.2 - 1          | 50-2                      | 1306      | -       | 0.2000   | 7471B  | PPB   | 11 Oct 2023 14:37:55 | S    | Std  |
| 2.5 - 1          | 52-5                      | 16695     | -       | 2.5000   | 7471B  | PPB   | 11 Oct 2023 14:40:12 | S    | Std  |
| 5.0 - 1          | 55                        | 34095     | -       | 5.0000   | 7471B  | PPB   | 11 Oct 2023 14:42:28 | S    | Std  |
| 7.5 - 1          | 57-5                      | 51783     | -       | 7.5000   | 7471B  | PPB   | 11 Oct 2023 14:44:44 | S    | Std  |
| 10.0 - 1         | 516                       | 68181     | -       | 10.0000  | 7471B  | PPB   | 11 Oct 2023 14:47:01 | S    | Std  |
| ICV11 - 1        | ICV11                     | 28282     | 4.1411  | -        | 7471B  | PPB   | 11 Oct 2023 14:49:49 | U    | SMPL |
| ICB11 - 1        | ICB11                     | -47       | 0.0069  | -        | 7471B  | PPB   | 11 Oct 2023 14:52:03 | U    | SMPL |
| CCV74 - 1        | CCV74                     | 34594     | 5.0622  | -        | 7471B  | PPB   | 11 Oct 2023 14:54:18 | U    | SMPL |
| CCB74 - 1        | CCB74                     | -48       | 0.0067  | -        | 7471B  | PPB   | 11 Oct 2023 14:56:34 | U    | SMPL |
| CRA - 1          | CRA                       | 1309      | 0.2048  | -        | 7471B  | PPB   | 11 Oct 2023 14:58:49 | U    | SMPL |
| HighStd - 1      | HighStd                   | 67979     | 9.9343  | -        | 7471B  | PPB   | 11 Oct 2023 15:01:04 | U    | SMPL |
| ChkStd - 1       | ChkStd                    | 48157     | 7.0415  | -        | 7471B  | PPB   | 11 Oct 2023 15:03:19 | U    | SMPL |
| PB156308BL - 1   | PBS                       | -102      | -0.0012 | -        | 7471B  | PPB   | 11 Oct 2023 15:05:36 | U    | SMPL |
| PB156308BS - 1   | LCSS                      | 30517     | 4.4672  | -        | 7471B  | PPB   | 11 Oct 2023 15:07:53 | U    | SMPL |
| O4798-01 - 1     | FRIABLE-ASBESTOS-SAMPLE   | 181320    | 26.4747 | -        | 7471B  | PPB   | 11 Oct 2023 15:10:09 | U    | SMPL |
| O4825-01 - 1     | SU-02-101023              | 3485      | 0.5223  | -        | 7471B  | PPB   | 11 Oct 2023 15:12:25 | U    | SMPL |
| O4830-01 - 1     | RT-2327                   | 15281     | 2.2438  | -        | 7471B  | PPB   | 11 Oct 2023 15:14:45 | U    | SMPL |
| O4837-05 - 1     | ETGI-290                  | 2403      | 0.3644  | -        | 7471B  | PPB   | 11 Oct 2023 15:17:01 | U    | SMPL |
| O4837-07 - 1     | ETGI-284                  | 3855      | 0.5763  | -        | 7471B  | PPB   | 11 Oct 2023 15:19:16 | U    | SMPL |
| CCV75 - 1        | CCV75                     | 34464     | 5.0432  | -        | 7471B  | PPB   | 11 Oct 2023 15:21:31 | U    | SMPL |
| CCB75 - 1        | CCB75                     | -5        | 0.0130  | -        | 7471B  | PPB   | 11 Oct 2023 15:23:46 | U    | SMPL |
| O4837-09 - 1     | SOIL-COMP                 | 3270      | 0.4909  | -        | 7471B  | PPB   | 11 Oct 2023 15:26:01 | U    | SMPL |
| O4839-01 - 1     | HD-02-101023              | 1116      | 0.1766  | -        | 7471B  | PPB   | 11 Oct 2023 15:33:03 | U    | SMPL |
| O4839-01DUP - 1  | HD-02-101023DUP           | 1033      | 0.1645  | -        | 7471B  | PPB   | 11 Oct 2023 15:35:17 | U    | SMPL |
| O4839-01MS - 1   | HD-02-101023MS            | 28652     | 4.1951  | -        | 7471B  | PPB   | 11 Oct 2023 15:37:33 | U    | SMPL |
| O4839-01MSD - 1  | HD-02-101023MSD           | 28777     | 4.2133  | -        | 7471B  | PPB   | 11 Oct 2023 15:39:48 | U    | SMPL |
| O4840-01 - 1     | ASPHALT-FLOOR-FRONT-RIGHT | 145       | 0.0349  | -        | 7471B  | PPB   | 11 Oct 2023 15:42:04 | U    | SMPL |
| O4840-02 - 1     | ASPHALT-FLOOR-FRONT-LEFT  | 148       | 0.0353  | -        | 7471B  | PPB   | 11 Oct 2023 15:44:21 | U    | SMPL |
| O4840-03 - 1     | ASPHALT-FLOOR-CENTER-AREA | 187       | 0.0410  | -        | 7471B  | PPB   | 11 Oct 2023 15:46:38 | U    | SMPL |
| O4840-04 - 1     | ASPHALT-FLOOR-REAR-LEFT   | 233       | 0.0477  | -        | 7471B  | PPB   | 11 Oct 2023 15:48:54 | U    | SMPL |
| CCV76 - 1        | CCV76                     | 35107     | 5.1371  | -        | 7471B  | PPB   | 11 Oct 2023 15:51:10 | U    | SMPL |
| CCB76 - 1        | CCB76                     | 11        | 0.0153  | -        | 7471B  | PPB   | 11 Oct 2023 15:53:24 | U    | SMPL |
| O4839-01LX5 - 1  |                           | 221       | 0.0460  | -        | 7471B  | PPB   | 11 Oct 2023 15:55:41 | U    | SMPL |
| O4839-01A - 1    |                           | 27695     | 4.0554  | -        | 7471B  | PPB   | 11 Oct 2023 15:57:57 | U    | SMPL |
| O4798-01DLX5 - 1 |                           | 45909     | 6.7135  | -        | 7471B  | PPB   | 11 Oct 2023 16:00:12 | U    | SMPL |
| CCV77 - 1        | CCV77                     | 35209     | 5.1520  | -        | 7471B  | PPB   | 11 Oct 2023 16:02:28 | U    | SMPL |
| CCB77 - 1        | CCB77                     | -21       | 0.0107  | -        | 7471B  | PPB   | 11 Oct 2023 16:04:45 | U    | SMPL |