



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/12/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 16:10
In Date: 10/11/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:11
Out Date: 10/12/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB127743

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O4850-01	YC630	1	1.15	8.40	9.55	8.55	88.1	
O4850-02	YC631	2	1.15	8.35	9.5	8.06	82.8	
O4850-03	YC640	3	1.16	8.71	9.87	8.86	88.4	
O4850-04	YC641	4	1.18	8.43	9.61	8.76	89.9	
O4850-05	YC642	5	1.16	8.61	9.77	9.12	92.5	
O4850-06	YC642D	6	1.16	8.61	9.77	9.12	92.5	
O4850-07	YC642S	7	1.16	8.61	9.77	9.12	92.5	
O4850-08	YC654	8	1.18	8.50	9.68	8.51	86.2	
O4850-09	YC660	9	1.15	8.39	9.54	9.06	94.3	
O4850-10	YC678	10	1.16	8.54	9.7	9.18	93.9	
O4850-11	YC680	11	1.18	8.40	9.58	8.49	87.0	
O4850-12	YC681	12	1.19	8.64	9.83	8.85	88.7	
O4850-13	YC5T5	13	1.19	8.32	9.51	9.08	94.8	
O4850-14	YC643	14	1.14	8.55	9.69	8.73	88.8	
O4850-15	YC644	15	1.19	8.55	9.74	9.00	91.3	
O4850-16	YC645	16	1.13	8.80	9.93	9.26	92.4	
O4850-17	YC673	17	1.15	8.37	9.52	8.72	90.4	
O4850-18	YC674	18	1.16	8.33	9.49	8.89	92.8	
O4850-19	YC675	19	1.17	8.59	9.76	9.04	91.6	
O4850-20	YC676	20	1.13	8.40	9.53	8.81	91.4	
O4850-21	YC677	21	1.19	8.42	9.61	9.24	95.6	
O4850-22	YC679	22	1.18	8.51	9.69	8.9	90.7	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$