

SFAM01.1

INSTRUMENT ID: CV1

Linear

$$\mu \text{ Abs.} :$$

5735C

A= 0.0000e+000

B= 1.7592e-004 slope

C = -3.7603e-002

Rho= 0.9998644

Accept=Accepted

Max. Conc:—

10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	1/10
0.00	0.000	-0.022	-0.022	91	0.000	91					1
0.05	0.050	0.040	-0.010	439	0.0 %	439	0				20
0.20	0.200	0.183	-0.017	1253	0.0 %	1253					9
2.50	2.500	2.595	0.095	14963	0.0 %	14963					4
5.00	5.000	5.018	0.018	28740	0.0 %	28740					0
7.50	7.500	7.385	-0.115	42190	0.0 %	42190					2
10.0	10.000	10.052	0.052	57350	0.0 %	57350					1

LB127746

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	91	-		0 SFAM01.1	PPB	10/11/2023 16:16	Std
	0.05 S01	439	-	0.05	SFAM01.1	PPB	10/11/2023 16:18	Std
	0.2 S02	1253	-	0.2	SFAM01.1	PPB	10/11/2023 16:21	Std
	2.5 S03	14963	-	2.5	SFAM01.1	PPB	10/11/2023 16:23	Std
	5 S04	28740	-	5	SFAM01.1	PPB	10/11/2023 16:25	Std
	7.5 S05	42190	-	7.5	SFAM01.1	PPB	10/11/2023 16:28	Std
	10 S06	0	-	10	SFAM01.1	PPB	10/11/2023 16:33	Std
ICV012	ICV012	22736	3.9622 -		SFAM01.1	PPB	10/11/2023 16:34	SMPL
ICB012	ICB012	7	-0.0364 -		SFAM01.1	PPB	10/11/2023 16:36	SMPL
CCV078	CCV078	30576	5.3414 -		SFAM01.1	PPB	10/11/2023 16:38	SMPL
CCB078	CCB078	3	-0.0371 -		SFAM01.1	PPB	10/11/2023 16:40	SMPL
PB156320BL	PBS320	20	-0.0341 -		SFAM01.1	PPB	10/11/2023 16:43	SMPL
O4617-01	MJKWD9	850	0.1119 -		SFAM01.1	PPB	10/11/2023 16:45	SMPL
O4617-02	MJKWE1	1191	0.1719 -		SFAM01.1	PPB	10/11/2023 16:47	SMPL
O4617-03	MJKWE2	403	0.0333 -		SFAM01.1	PPB	10/11/2023 16:49	SMPL
O4617-04	MJKWE3	356	0.025 -		SFAM01.1	PPB	10/11/2023 16:52	SMPL
O4617-05	MJKWE4	380	0.0292 -		SFAM01.1	PPB	10/11/2023 16:54	SMPL
O4617-06	MJKWE5	603	0.0685 -		SFAM01.1	PPB	10/11/2023 16:56	SMPL
O4617-07	MJKWE6	2019	0.3176 -		SFAM01.1	PPB	10/11/2023 16:58	SMPL
O4617-08	MJKWE7	1430	0.214 -		SFAM01.1	PPB	10/11/2023 17:01	SMPL
O4617-09	MJKWE7D	1325	0.1955 -		SFAM01.1	PPB	10/11/2023 17:03	SMPL
O4617-10	MJKWE7S	17620	3.0621 -		SFAM01.1	PPB	10/11/2023 17:06	SMPL
O4617-11	MJKWE8	587	0.0657 -		SFAM01.1	PPB	10/11/2023 17:08	SMPL
O4617-12	MJKWE9	11718	2.0238 -		SFAM01.1	PPB	10/11/2023 17:10	SMPL
O4617-13	MJKWF0	1349	0.1997 -		SFAM01.1	PPB	10/11/2023 17:12	SMPL
O4617-14	MJKWF1	1311	0.193 -		SFAM01.1	PPB	10/11/2023 17:15	SMPL
O4617-15	MJKWF2	534	0.0563 -		SFAM01.1	PPB	10/11/2023 17:17	SMPL
O4617-16	MJKWF3	1313	0.1934 -		SFAM01.1	PPB	10/11/2023 17:19	SMPL
O4617-17	MJKWF4	605	0.0688 -		SFAM01.1	PPB	10/11/2023 17:21	SMPL
O4617-18	MJKWF5	557	0.0604 -		SFAM01.1	PPB	10/11/2023 17:24	SMPL
O4617-19	MJKWF7	512	0.0525 -		SFAM01.1	PPB	10/11/2023 17:26	SMPL
CCV079	CCV079	30226	5.2798 -		SFAM01.1	PPB	10/11/2023 17:28	SMPL
CCB079	CCB079	48	-0.0292 -		SFAM01.1	PPB	10/11/2023 17:30	SMPL
O4617-20	MJKWF8	618	0.0711 -		SFAM01.1	PPB	10/11/2023 17:33	SMPL
O4617-21	MJKWF9	551	0.0593 -		SFAM01.1	PPB	10/11/2023 17:35	SMPL
O4617-22	MJKWG0	1824	0.2833 -		SFAM01.1	PPB	10/11/2023 17:37	SMPL
PB156321BL	PBS321	22	-0.0337 -		SFAM01.1	PPB	10/11/2023 17:39	SMPL
O4644-01	MJKWG1	2283	0.364 -		SFAM01.1	PPB	10/11/2023 17:42	SMPL
O4644-02	MJKWG2	18976	3.3007 -		SFAM01.1	PPB	10/11/2023 17:44	SMPL
O4644-03	MJKWG3	460	0.0433 -		SFAM01.1	PPB	10/11/2023 17:46	SMPL
O4644-04	MJKWG4	557	0.0604 -		SFAM01.1	PPB	10/11/2023 17:49	SMPL
O4644-05	MJKWG6	590	0.0662 -		SFAM01.1	PPB	10/11/2023 17:51	SMPL
O4644-06	MJKWG7	1001	0.1385 -		SFAM01.1	PPB	10/11/2023 17:53	SMPL
O4644-07	MJKWG8	1403	0.2092 -		SFAM01.1	PPB	10/11/2023 17:55	SMPL
O4644-08	MJKWG9	944	0.1285 -		SFAM01.1	PPB	10/11/2023 17:58	SMPL
O4644-09	MJKWH0	1431	0.2141 -		SFAM01.1	PPB	10/11/2023 18:00	SMPL

LB127746

INSTRUMENT ID : CV1

O4644-10	MJKWG5	715	0.0882 -	SFAM01.1	PPB	10/11/2023 18:02	SMPL
O4644-11	MJKWG5D	752	0.0947 -	SFAM01.1	PPB	10/11/2023 18:04	SMPL
O4644-12	MJKWG5S	14864	2.5773 -	SFAM01.1	PPB	10/11/2023 18:07	SMPL
O4644-13	MJKWJ0	4340	0.7259 -	SFAM01.1	PPB	10/11/2023 18:09	SMPL
O4644-14	MJKWJ1	509	0.0519 -	SFAM01.1	PPB	10/11/2023 18:11	SMPL
O4644-15	MJKWJ2	593	0.0667 -	SFAM01.1	PPB	10/11/2023 18:13	SMPL
O4644-16	MJKWJ3	583	0.065 -	SFAM01.1	PPB	10/11/2023 18:16	SMPL
CCV080	CCV080	28139	4.9127 -	SFAM01.1	PPB	10/11/2023 18:18	SMPL
CCB080	CCB080	36	-0.0313 -	SFAM01.1	PPB	10/11/2023 18:20	SMPL
O4644-17	MJKWJ4	580	0.0644 -	SFAM01.1	PPB	10/11/2023 18:22	SMPL
O4644-18	MJKWJ5	401	0.0329 -	SFAM01.1	PPB	10/11/2023 18:25	SMPL
O4644-19	MJKWJ6	583	0.065 -	SFAM01.1	PPB	10/11/2023 18:27	SMPL
O4644-20	MJKWJ7	1112	0.158 -	SFAM01.1	PPB	10/11/2023 18:29	SMPL
O4644-21	MJKWJ8	6347	1.079 -	SFAM01.1	PPB	10/11/2023 18:32	SMPL
O4644-22	MJKWJ9	2029	0.3193 -	SFAM01.1	PPB	10/11/2023 18:34	SMPL
CCV081	CCV081	28232	4.929 -	SFAM01.1	PPB	10/11/2023 18:36	SMPL
CCB081	CCB081	34	-0.0316 -	SFAM01.1	PPB	10/11/2023 18:38	SMPL