



TOTAL VOLATILE SOLIDS SM2540 G

TEMP1 IN: 104 °C 10/11/2023 17:30 TEMP1 OUT: 103 °C 10/12/2023 07:30
TEMP2 IN: 103 °C 10/12/2023 08:00 TEMP2 OUT: 103 °C 10/12/2023 09:30
TEMP3 IN: 540 °C 10/12/2023 10:00 TEMP3 OUT: 540 °C 10/12/2023 11:30
TEMP4 IN: 560 °C 10/12/2023 12:00 TEMP4 OUT: 560 °C 10/12/2023 13:30

Run Number: LB127750

SUPERVISOR: Iwona

ANALYST: jignesh

BalanceID: WC SC-6

OvenID: WC OVEN-1

Dish #	Lab ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish + SampleWt Drying @103-@105°C (g)	Final Dish + SampleWt Drying @103-@105°C (g)	Dish + Samplewt Drying @550 (±50) °C (g)	Final Dish + Samplewt Drying @550 (±50) °C (g)	Weight Diff (g)	Result (mg/L)
1	LB127750BL	96.8743	96.8743	100	96.8743	96.8743	96.8743	96.8743	0.0000	0
2	O4859-01	100.7700	100.7700	100	100.8480	100.848	100.8419	100.8419	0.0061	61
3	O4859-01DUP	75.5520	75.5520	100	75.6299	75.6299	75.6235	75.6235	0.0064	64

A = Sample Volume (ml)

B = Final Dish + Samplewt Drying @550 (±50) °C (g)

C = Final Dish + SampleWt Drying @103-@105°C (g)

D = Weight (g)

E = Final Empty Dish Weight (g)

F = Final Dish + SampleWt Drying @103-@105°C (g)

Weight D = C - B

Result % = $\frac{D}{A} \star 1000000$

WORKLIST(Hardcopy Internal Chain)

127150

WorkList Name : tvs waterv 04849

WorkList ID : 174500

Department : Wet-Chemistry

Date : 10-11-2023 16:39:59

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4859-01	SVS-GROUNDWATER	Water	TVS	Cool 4 deg C	PSEG03	E31	10/11/2023	SM2540 G

Date/Time

10-11-23 16:40

Raw Sample Received by:

SV (QC)

Raw Sample Relinquished by:

SV SM

Date/Time

10-11-23

Raw Sample Received by:

SV SM

Raw Sample Relinquished by:

SV (QC)