

L3127765


SFAM01.1

INSTRUMENT ID: CY1

Linear

μ Abs.:

56645



A= 0.0000e+000

B= 1.7514e-004 slope

C= -1.1269e-001 y-intercept

Rho= 0.9992525

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0/0
0.00	0.000	-0.074	-0.074	223	0.000	223					-
0.05	0.050					439	0				-
0.20	0.200	0.169	-0.031	1608	0.0 %	1608					-16
2.50	2.500	2.452	-0.048	14643	0.0 %	14643					-2
5.00	5.000	5.265	0.265	30704	0.0 %	30704					5
7.50	7.500	7.580	0.080	43925	0.0 %	43925					1
10.0	10.000	9.808	-0.192	56645	0.0 %	56645					-2

LB127765

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Stnd Conc	Method	Units	Date	Type
	0 S0	223	-		0 SFAM01.1	PPB	10/12/2023 12:24	Std
	0.2 S01	1608	-		0.2 SFAM01.1	PPB	10/12/2023 12:28	Std
	2.5 S02	14643	-		2.5 SFAM01.1	PPB	10/12/2023 12:31	Std
	5 S03	30704	-		5 SFAM01.1	PPB	10/12/2023 12:33	Std
	7.5 S04	43925	-		7.5 SFAM01.1	PPB	10/12/2023 12:35	Std
	10 S05	56645	-		10 SFAM01.1	PPB	10/12/2023 12:37	Std
ICV013	ICV013	24671	4.2081 -		SFAM01.1	PPB	10/12/2023 12:40	SMPL
ICB013	ICB013	136	-0.0889 -		SFAM01.1	PPB	10/12/2023 12:43	SMPL
CCV082	CCV082	30943	5.3066 -		SFAM01.1	PPB	10/12/2023 12:45	SMPL
CCB082	CCB082	128	-0.0903 -		SFAM01.1	PPB	10/12/2023 12:47	SMPL
PB156342BL	PBW342	243	-0.0701 -		SFAM01.1	PPB	10/12/2023 12:49	SMPL
O4659-12	EW506	285	-0.0628 -		SFAM01.1	PPB	10/12/2023 12:51	SMPL
O4675-02	MJWKJ0	255	-0.068 -		SFAM01.1	PPB	10/12/2023 12:54	SMPL
CCV083	CCV083	30915	5.3017 -		SFAM01.1	PPB	10/12/2023 12:56	SMPL
CCB083	CCB083	205	-0.0768 -		SFAM01.1	PPB	10/12/2023 12:58	SMPL