

LB127767


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

57541



A= 0.0000e+000

B= 1.7580e-004 slope

C= -3.9368e-002 y-intercept

Rho= 0.9997694

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.00	0.000	-0.026	-0.026	76	0.000	76				
0.05	0.050	0.044	-0.006	473	0.0 %	473	0			
0.20	0.200	0.185	-0.015	1275	0.0 %	1275				
2.50	2.500	2.602	0.102	15025	0.0 %	15025				
5.00	5.000	5.031	0.031	28842	0.0 %	28842				
7.50	7.500	7.338	-0.162	41963	0.0 %	41963				
10.0	10.000	10.076	0.076	57541	0.0 %	57541				

128
 10/12/23
 0/0/0
 7
 -12 -12
 -8
 4
 1
 -2
 1

LB127767

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0		76	-		0 SFAM01.1	PPB	10/12/2023 13:41	Std
0.05		473	-	0.05	SFAM01.1	PPB	10/12/2023 13:43	Std
0.2		1275	-	0.2	SFAM01.1	PPB	10/12/2023 13:45	Std
2.5		15025	-	2.5	SFAM01.1	PPB	10/12/2023 13:48	Std
5		28842	-	5	SFAM01.1	PPB	10/12/2023 13:52	Std
7.5		41963	-	7.5	SFAM01.1	PPB	10/12/2023 13:54	Std
10		57541	-	10	SFAM01.1	PPB	10/12/2023 13:56	Std
ICV014	ICV014	22521	3.9199 -		SFAM01.1	PPB	10/12/2023 13:59	SMPL
ICB014	ICB014	10	-0.0376 -		SFAM01.1	PPB	10/12/2023 14:02	SMPL
CCV084	CCV084	28860	5.0343 -		SFAM01.1	PPB	10/12/2023 14:04	SMPL
CCB084	CCB084	14	-0.0369 -		SFAM01.1	PPB	10/12/2023 14:06	SMPL
PB156346BL	PBS346	315	0.016 -		SFAM01.1	PPB	10/12/2023 14:08	SMPL
O4674-02	MJUBH4	2962	0.4814 -		SFAM01.1	PPB	10/12/2023 14:11	SMPL
O4674-03	MJUBH5	13128	2.2686 -		SFAM01.1	PPB	10/12/2023 14:13	SMPL
O4674-04	MJUBH6	2553	0.4095 -		SFAM01.1	PPB	10/12/2023 14:15	SMPL
O4674-05	MJUBH7	6874	1.1691 -		SFAM01.1	PPB	10/12/2023 14:17	SMPL
O4674-06	MJUBH8	2700	0.4353 -		SFAM01.1	PPB	10/12/2023 14:20	SMPL
O4674-07	MJUBH9	2365	0.3764 -		SFAM01.1	PPB	10/12/2023 14:22	SMPL
O4674-08	MJUBJ0	5280	0.8889 -		SFAM01.1	PPB	10/12/2023 14:24	SMPL
O4674-09	MJUBJ0D	6120	1.0365 -		SFAM01.1	PPB	10/12/2023 14:26	SMPL
O4674-10	MJUBJ0S	22577	3.9297 -		SFAM01.1	PPB	10/12/2023 14:29	SMPL
O4674-11	MJUBJ1	6434	1.0917 -		SFAM01.1	PPB	10/12/2023 14:31	SMPL
O4674-12	MJUBJ2	1836	0.2834 -		SFAM01.1	PPB	10/12/2023 14:33	SMPL
O4674-13	MJUBJ3	528	0.0535 -		SFAM01.1	PPB	10/12/2023 14:35	SMPL
O4674-14	MJUBJ4	885	0.1162 -		SFAM01.1	PPB	10/12/2023 14:38	SMPL
O4674-15	MJUBJ5	605	0.067 -		SFAM01.1	PPB	10/12/2023 14:40	SMPL
O4674-16	MJUBJ6	1098	0.1537 -		SFAM01.1	PPB	10/12/2023 14:42	SMPL
O4674-17	MJUBJ7	757	0.0937 -		SFAM01.1	PPB	10/12/2023 14:44	SMPL
O4674-18	MJUBJ8	653	0.0754 -		SFAM01.1	PPB	10/12/2023 14:47	SMPL
O4674-19	MJUBJ9	596	0.0654 -		SFAM01.1	PPB	10/12/2023 14:49	SMPL
O4674-20	MJUBK0	918	0.122 -		SFAM01.1	PPB	10/12/2023 14:51	SMPL
CCV085	CCV085	29346	5.1197 -		SFAM01.1	PPB	10/12/2023 14:54	SMPL
CCB085	CCB085	28	-0.0344 -		SFAM01.1	PPB	10/12/2023 14:56	SMPL
O4674-21	MJUBK1	377	0.0269 -		SFAM01.1	PPB	10/12/2023 14:58	SMPL
O4674-22	MJUBK2	527	0.0533 -		SFAM01.1	PPB	10/12/2023 15:00	SMPL
PB156347BL	PBS347	307	0.0146 -		SFAM01.1	PPB	10/12/2023 15:03	SMPL
O4675-01	MJKWK0	715	0.0863 -		SFAM01.1	PPB	10/12/2023 15:05	SMPL
O4675-03	MJKWK0D	705	0.0846 -		SFAM01.1	PPB	10/12/2023 15:07	SMPL
O4675-04	MJKWK0S	15524	2.6898 -		SFAM01.1	PPB	10/12/2023 15:09	SMPL
CCV086	CCV086	29536	5.1531 -		SFAM01.1	PPB	10/12/2023 15:12	SMPL
CCB086	CCB086	31	-0.0339 -		SFAM01.1	PPB	10/12/2023 15:14	SMPL

LB127767

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0		76	-		0 SFAM01.1	PPB	10/12/2023 13:41	Std
0.05		473	-	0.05	SFAM01.1	PPB	10/12/2023 13:43	Std
0.2		1275	-	0.2	SFAM01.1	PPB	10/12/2023 13:45	Std
2.5		15025	-	2.5	SFAM01.1	PPB	10/12/2023 13:48	Std
5		28842	-	5	SFAM01.1	PPB	10/12/2023 13:52	Std
7.5		41963	-	7.5	SFAM01.1	PPB	10/12/2023 13:54	Std
10		57541	-	10	SFAM01.1	PPB	10/12/2023 13:56	Std
ICV014	ICV014	22521	3.9199 -		SFAM01.1	PPB	10/12/2023 13:59	SMPL
ICB014	ICB014	10	-0.0376 -		SFAM01.1	PPB	10/12/2023 14:02	SMPL
CCV084	CCV084	28860	5.0343 -		SFAM01.1	PPB	10/12/2023 14:04	SMPL
CCB084	CCB084	14	-0.0369 -		SFAM01.1	PPB	10/12/2023 14:06	SMPL
PB156346BL	PBS346	315	0.016 -		SFAM01.1	PPB	10/12/2023 14:08	SMPL
O4674-02	MJUBH4	2962	0.4814 -		SFAM01.1	PPB	10/12/2023 14:11	SMPL
O4674-03	MJUBH5	13128	2.2686 -		SFAM01.1	PPB	10/12/2023 14:13	SMPL
O4674-04	MJUBH6	2553	0.4095 -		SFAM01.1	PPB	10/12/2023 14:15	SMPL
O4674-05	MJUBH7	6874	1.1691 -		SFAM01.1	PPB	10/12/2023 14:17	SMPL
O4674-06	MJUBH8	2700	0.4353 -		SFAM01.1	PPB	10/12/2023 14:20	SMPL
O4674-07	MJUBH9	2365	0.3764 -		SFAM01.1	PPB	10/12/2023 14:22	SMPL
O4674-08	MJUBJ0	5280	0.8889 -		SFAM01.1	PPB	10/12/2023 14:24	SMPL
O4674-09	MJUBJ0D	6120	1.0365 -		SFAM01.1	PPB	10/12/2023 14:26	SMPL
O4674-10	MJUBJ0S	22577	3.9297 -		SFAM01.1	PPB	10/12/2023 14:29	SMPL
O4674-11	MJUBJ1	6434	1.0917 -		SFAM01.1	PPB	10/12/2023 14:31	SMPL
O4674-12	MJUBJ2	1836	0.2834 -		SFAM01.1	PPB	10/12/2023 14:33	SMPL
O4674-13	MJUBJ3	528	0.0535 -		SFAM01.1	PPB	10/12/2023 14:35	SMPL
O4674-14	MJUBJ4	885	0.1162 -		SFAM01.1	PPB	10/12/2023 14:38	SMPL
O4674-15	MJUBJ5	605	0.067 -		SFAM01.1	PPB	10/12/2023 14:40	SMPL
O4674-16	MJUBJ6	1098	0.1537 -		SFAM01.1	PPB	10/12/2023 14:42	SMPL
O4674-17	MJUBJ7	757	0.0937 -		SFAM01.1	PPB	10/12/2023 14:44	SMPL
O4674-18	MJUBJ8	653	0.0754 -		SFAM01.1	PPB	10/12/2023 14:47	SMPL
O4674-19	MJUBJ9	596	0.0654 -		SFAM01.1	PPB	10/12/2023 14:49	SMPL
O4674-20	MJUBK0	918	0.122 -		SFAM01.1	PPB	10/12/2023 14:51	SMPL
CCV085	CCV085	29346	5.1197 -		SFAM01.1	PPB	10/12/2023 14:54	SMPL
CCB085	CCB085	28	-0.0344 -		SFAM01.1	PPB	10/12/2023 14:56	SMPL
O4674-21	MJUBK1	377	0.0269 -		SFAM01.1	PPB	10/12/2023 14:58	SMPL
O4674-22	MJUBK2	527	0.0533 -		SFAM01.1	PPB	10/12/2023 15:00	SMPL
PB156347BL	PBS347	307	0.0146 -		SFAM01.1	PPB	10/12/2023 15:03	SMPL
O4675-01	MJKWK0	715	0.0863 -		SFAM01.1	PPB	10/12/2023 15:05	SMPL
O4675-03	MJKWK0D	705	0.0846 -		SFAM01.1	PPB	10/12/2023 15:07	SMPL
O4675-04	MJKWK0S	15524	2.6898 -		SFAM01.1	PPB	10/12/2023 15:09	SMPL
CCV086	CCV086	29536	5.1531 -		SFAM01.1	PPB	10/12/2023 15:12	SMPL
CCB086	CCB086	31	-0.0339 -		SFAM01.1	PPB	10/12/2023 15:14	SMPL