

LB127771


SFAM01.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

60095



A= 0.0000e+000
 B= 1.6917e-004 *slope*
 C= -5.1936e-003 *y-intercept*
 Rho= 0.9991802
 Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>o/p</i>
0.00	0.000	0.005	0.005	58	0.000	58					
0.05	0.050					473	0				
0.20	0.200	0.204	0.004	1238	0.0 %	1238					
2.50	2.500	2.547	0.047	15086	0.0 %	15086					
5.00	5.000	5.095	0.095	30146	0.0 %	30146					
7.50	7.500	7.189	-0.311	42526	0.0 %	42526					
10.0	10.000	10.161	0.161	60095	0.0 %	60095					

o/p
 =
 2
 2
 2
 2
 - 4
 2

LB127771
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 SS0	58	-		0 SFAM01.1	PPB	10/12/2023 15:46	Std
	0.2 S01	1238	-		0.2 SFAM01.1	PPB	10/12/2023 15:48	Std
	2.5 S02	15086	-		2.5 SFAM01.1	PPB	10/12/2023 15:50	Std
	5 S03	30146	-		5 SFAM01.1	PPB	10/12/2023 15:53	Std
	7.5 S04	42526	-		7.5 SFAM01.1	PPB	10/12/2023 15:55	Std
	10 S05	60095	-		10 SFAM01.1	PPB	10/12/2023 16:00	Std
ICV015	ICV015	23018	3.8887 -		SFAM01.1	PPB	10/12/2023 16:05	SMPL
ICB015	ICB015	18	-0.0021 -		SFAM01.1	PPB	10/12/2023 16:07	SMPL
CCV087	CCV087	29847	5.0439 -		SFAM01.1	PPB	10/12/2023 16:09	SMPL
CCB087	CCB087	17	-0.0023 -		SFAM01.1	PPB	10/12/2023 16:11	SMPL
PB156351BL	PBS351	54	0.0039 -		SFAM01.1	PPB	10/12/2023 16:14	SMPL
O4659-07	EW501	3687	0.6185 -		SFAM01.1	PPB	10/12/2023 16:16	SMPL
O4659-08	EW501D	3195	0.5353 -		SFAM01.1	PPB	10/12/2023 16:18	SMPL
O4659-09	EW501S	16727	2.8245 -		SFAM01.1	PPB	10/12/2023 16:20	SMPL
O4659-10	EW502	6322	1.0643 -		SFAM01.1	PPB	10/12/2023 16:23	SMPL
O4659-11	EW504	632	0.1017 -		SFAM01.1	PPB	10/12/2023 16:25	SMPL
O4659-13	EW503	6132	1.0321 -		SFAM01.1	PPB	10/12/2023 16:27	SMPL
CCV088	CCV088	30033	5.0754 -		SFAM01.1	PPB	10/12/2023 16:29	SMPL
CCB088	CCB088	42	0.0019 -		SFAM01.1	PPB	10/12/2023 16:32	SMPL