



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/13/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 10/12/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:15
Out Date: 10/13/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB127777

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O4862-01	RPTP-14A	1	1.16	8.81	9.97	8.56	84.0	
O4862-02	RPTP-13A	2	1.15	8.81	9.96	9.54	95.2	
O4862-03	RPTP-12A	3	1.15	8.43	9.58	8.3	84.8	
O4862-04	RPTP-11A	4	1.18	8.54	9.72	8.32	83.6	
O4862-05	RPTP-10A	5	1.18	8.81	9.99	8.72	85.6	
O4862-06	RPTP-9A	6	1.18	8.47	9.65	9.22	94.9	
O4862-07	RPTP-8A	7	1.17	8.80	9.97	8.58	84.2	
O4862-08	RPTP-7A	8	1.15	8.41	9.56	9.21	95.8	
O4862-09	RPTP-6A	9	1.16	8.36	9.52	8.03	82.2	
O4862-10	RPTP-5A	10	1.14	8.83	9.97	9.55	95.2	
O4862-11	RPTP-4A	11	1.13	8.83	9.96	8.59	84.5	
O4862-12	RPTP-3A	12	1.19	8.63	9.82	9.42	95.4	
O4862-13	RPTP-2A	13	1.19	8.48	9.67	8.28	83.6	
O4862-14	RPTP-1A	14	1.19	8.62	9.81	9.38	95.0	
O4862-15	TP-1-4A	15	1.16	8.57	9.73	9.35	95.6	
O4876-01	30103	16	1.18	8.63	9.81	9.75	99.3	
O4876-02	30104	17	1.13	8.49	9.62	9.57	99.4	
O4878-01	72-11995	18	1.16	8.57	9.73	8.73	88.3	
O4878-02	72-11995-E2	19	1.13	8.44	9.57	8.42	86.4	
O4878-03	222	20	1.12	8.56	9.68	8.82	90.0	
O4878-04	222-E2	21	1.19	8.50	9.69	8.02	80.4	
O4878-05	283	22	1.19	8.53	9.72	7.68	76.1	
O4878-06	283-E2	23	1.15	8.81	9.96	7.86	76.2	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$