

LB127783


7471B

INSTRUMENT ID: CV1

Linear

μ Abs.:

59575



A= 0.0000e+000

B= 1.7043e-004

C= -2.7008e-002

Rho= 0.9996483

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	0.001	0.001	166	0.000	166					
0.2	0.200	0.204	0.004	1357	0.0 %	1357					
2.5	2.500	2.597	0.097	15394	0.0 %	15394					
5.0	5.000	4.940	-0.060	29143	0.0 %	29143					
7.5	7.500	7.331	-0.169	43174	0.0 %	43174					
10.0	10.000	10.127	0.127	59575	0.0 %	59575					

400
1
2
4
1
1

LB127783 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 13 Oct 2023 08:38:26

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1		166	-	0.0000	7471B	PPB	13 Oct 2023 08:44:10	S	Std
0.2 - 1		1357	-	0.2000	7471B	PPB	13 Oct 2023 08:46:28	S	Std
2.5 - 1		15394	-	2.5000	7471B	PPB	13 Oct 2023 08:48:45	S	Std
5.0 - 1		29143	-	5.0000	7471B	PPB	13 Oct 2023 08:51:02	S	Std
7.5 - 1		43174	-	7.5000	7471B	PPB	13 Oct 2023 08:53:18	S	Std
10.0 - 1		59575	-	10.0000	7471B	PPB	13 Oct 2023 08:58:10	S	Std
ICV16 - 1	ICV16	23412	3.9632	-	7471B	PPB	13 Oct 2023 09:03:33	U	SMPL
ICB16 - 1	ICB16	46	-0.0192	-	7471B	PPB	13 Oct 2023 09:05:47	U	SMPL
CCV89 - 1	CCV89	29679	5.0313	-	7471B	PPB	13 Oct 2023 09:08:03	U	SMPL
CCB89 - 1	CCB89	38	-0.0205	-	7471B	PPB	13 Oct 2023 09:10:17	U	SMPL
CRA - 1	CRA	1348	0.2027	-	7471B	PPB	13 Oct 2023 09:12:34	U	SMPL
HighStd - 1	HighStd	59336	10.0858	-	7471B	PPB	13 Oct 2023 09:14:49	U	SMPL
ChkStd - 1	ChkStd	41564	7.0569	-	7471B	PPB	13 Oct 2023 09:17:04	U	SMPL
PB156354BL - 1	PBS	-7	-0.0282	-	7471B	PPB	13 Oct 2023 09:19:22	U	SMPL
PB156354BS - 1	LCSS	26910	4.5594	-	7471B	PPB	13 Oct 2023 09:21:39	U	SMPL
O4841-01 - 1	G-1-(5-10)	21571	3.6494	-	7471B	PPB	13 Oct 2023 09:23:55	U	SMPL
O4841-02 - 1	M-3(0-5)	6153	1.0217	-	7471B	PPB	13 Oct 2023 09:26:12	U	SMPL
O4844-01 - 1	OR-603-COMP-05	721	0.0959	-	7471B	PPB	13 Oct 2023 09:28:30	U	SMPL
O4844-01DUP - 1	OR-603-COMP-05DUP	503	0.0587	-	7471B	PPB	13 Oct 2023 09:30:47	U	SMPL
CCV90 - 1	CCV90	29511	5.0027	-	7471B	PPB	13 Oct 2023 09:35:17	U	SMPL
CCB90 - 1	CCB90	26	-0.0226	-	7471B	PPB	13 Oct 2023 09:37:33	U	SMPL
O4844-01MSD - 1	OR-603-COMP-05MSD	26851	4.5493	-	7471B	PPB	13 Oct 2023 09:39:49	U	SMPL
O4844-05 - 1	OR-603-COMP-06	410	0.0429	-	7471B	PPB	13 Oct 2023 09:42:04	U	SMPL
O4844-09 - 1	OR-603-COMP-07	441	0.0482	-	7471B	PPB	13 Oct 2023 09:44:20	U	SMPL
O4844-13 - 1	OR-603-COMP-08	434	0.0470	-	7471B	PPB	13 Oct 2023 09:46:36	U	SMPL
O4846-01 - 1	WASTE	12257	2.0620	-	7471B	PPB	13 Oct 2023 09:48:51	U	SMPL
O4854-01 - 1	TR-05-101123	2562	0.4096	-	7471B	PPB	13 Oct 2023 09:51:07	U	SMPL
O4858-01 - 1	WC-1	2283	0.3621	-	7471B	PPB	13 Oct 2023 09:53:23	U	SMPL
O4858-04 - 1	WC-2	963	0.1371	-	7471B	PPB	13 Oct 2023 09:55:40	U	SMPL
O4858-07 - 1	WC-3	4201	0.6890	-	7471B	PPB	13 Oct 2023 09:57:57	U	SMPL
O4844-01MS - 1	OR-603-COMP-05MS	25456	4.3115	-	7471B	PPB	13 Oct 2023 10:00:13	U	SMPL
CCV91 - 1	CCV91	29648	5.0260	-	7471B	PPB	13 Oct 2023 10:02:28	U	SMPL
CCB91 - 1	CCB91	40	-0.0202	-	7471B	PPB	13 Oct 2023 10:04:43	U	SMPL
O4844-01LX5 - 1		186	0.0047	-	7471B	PPB	13 Oct 2023 10:07:00	U	SMPL
O4844-01A - 1		26462	4.4830	-	7471B	PPB	13 Oct 2023 10:09:15	U	SMPL
CCV92 - 1	CCV92	29816	5.0546	-	7471B	PPB	13 Oct 2023 10:11:31	U	SMPL
CCB92 - 1	CCB92	29	-0.0221	-	7471B	PPB	13 Oct 2023 10:13:47	U	SMPL