


LB127784

245.1

INSTRUMENT ID: CV1

 μ Abs.:

42510



Linear

A= 0.0000e+000

B= 2.3761e-004

C= -1.4765e-003

Rho= 0.9997783

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	g/100
0.0	0.000	0.018	0.018	82	0.000	82					
0.05	0.050	0.062	0.012	269	0.0 %	269	0				24
0.2	0.200	0.210	0.010	888	0.0 %	888					5
2.5	2.500	2.470	-0.030	10400	0.0 %	10400					-1
5.0	5.000	5.058	0.058	21294	0.0 %	21294					1
7.5	7.500	7.333	-0.167	30866	0.0 %	30866					-2
10.0	10.000	10.099	0.099	42510	0.0 %	42510					1

LB127784 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 13 Oct 2023 11:42:27

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	5.0	82	-	0.0000	245.1	PPB	13 Oct 2023 11:42:55	S	Std
0.05 - 1	50.05	269	-	0.0500	245.1	PPB	13 Oct 2023 11:45:13	S	Std
0.2 - 1	50.2	888	-	0.2000	245.1	PPB	13 Oct 2023 11:47:30	S	Std
2.5 - 1	52.5	10400	-	2.5000	245.1	PPB	13 Oct 2023 11:49:46	S	Std
5.0 - 1	55	21294	-	5.0000	245.1	PPB	13 Oct 2023 11:52:02	S	Std
7.5 - 1	57.5	30868	-	7.5000	245.1	PPB	13 Oct 2023 11:54:18	S	Std
10.0 - 1	58	42510	-	10.0000	245.1	PPB	13 Oct 2023 11:56:34	S	Std
ICV17 - 1	ICV17	16217	3.8519	-245.1		PPB	13 Oct 2023 12:00:01	U	SMPL
ICB17 - 1	ICB17	-54	-0.0143	-245.1		PPB	13 Oct 2023 12:02:16	U	SMPL
CCV93 - 1	CCV93	21505	5.1084	-245.1		PPB	13 Oct 2023 12:04:32	U	SMPL
CCB93 - 1	CCB93	-68	-0.0176	-245.1		PPB	13 Oct 2023 12:06:47	U	SMPL
CRA - 1	CRA	242	0.0560	-245.1		PPB	13 Oct 2023 12:09:03	U	SMPL
HighStd - 1	HighStd	44129	10.4841	-245.1		PPB	13 Oct 2023 12:11:18	U	SMPL
ChkStd - 1	ChkStd	29368	6.9767	-245.1		PPB	13 Oct 2023 12:13:33	U	SMPL
PB156378BL - 1	PBW	14	0.0019	-245.1		PPB	13 Oct 2023 12:15:50	U	SMPL
PB156378BS - 1	LCSW	16996	4.0370	-245.1		PPB	13 Oct 2023 12:18:07	U	SMPL
O4478-02 - 1	EFF-WW	-21	-0.0065	-245.1		PPB	13 Oct 2023 12:20:23	U	SMPL
O4722-01 - 1	EFFLUENT-DAY-1-METALS	476	0.1116	-245.1		PPB	13 Oct 2023 12:30:42	U	SMPL
O4722-01DUP - 1	EFFLUENT-DAY-1-METALS DUP	352	0.0822	-245.1		PPB	13 Oct 2023 12:35:41	U	SMPL
O4722-01MS - 1	EFFLUENT-DAY-1-METALS MS	16036	3.8089	-245.1		PPB	13 Oct 2023 12:37:57	U	SMPL
O4722-01MSD - 1	EFFLUENT-DAY-1-METALS MSD	17821	4.2330	-245.1		PPB	13 Oct 2023 12:40:13	U	SMPL
CCV94 - 1	CCV94	20665	4.9088	-245.1		PPB	13 Oct 2023 12:50:06	U	SMPL
CCB94 - 1	CCB94	-28	-0.0081	-245.1		PPB	13 Oct 2023 12:52:21	U	SMPL
O4722-01LX5 - 1		1	-0.0012	-245.1		PPB	13 Oct 2023 12:54:37	U	SMPL
O4722-01A - 1		14622	3.4729	-245.1		PPB	13 Oct 2023 12:59:00	U	SMPL
CCV95 - 1	CCV95	20048	4.7622	-245.1		PPB	13 Oct 2023 13:01:16	U	SMPL
CCB95 - 1	CCB95	-182	-0.0447	-245.1		PPB	13 Oct 2023 13:03:32	U	SMPL