

LB127796

7470A

Linear

 μ Abs.:

43661

A= 0.0000e+000

B= 2.3138e-004

C= 3.7196e-002

Rho= 0.9996945

Accept=Accepted

Max. Conc.: — 10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.038	0.038	5	0.000	5	0				
0.2	0.200	0.234	0.034	849	0.0 %	849					17
2.5	2.500	2.500	-0.000	10643	0.0 %	10643					0
5.0	5.000	4.947	-0.053	21218	0.0 %	21218					-1
7.5	7.500	7.342	-0.158	31570	0.0 %	31570					-2
10.0	10.000	10.140	0.140	43661	0.0 %	43661					1

Sample ID	Extended ID	μ Abs.	Conc.	Stdnd Conc	Method	Units	Date	Type
	0	5	-		0 7470A	PPB	10/13/2023 13:59	Std
	0.2	849	-		0.2 7470A	PPB	10/13/2023 14:01	Std
	2.5	10643	-		2.5 7470A	PPB	10/13/2023 14:04	Std
	5	21218	-		5 7470A	PPB	10/13/2023 14:06	Std
	7.5	31570	-		7.5 7470A	PPB	10/13/2023 14:11	Std
	10	43661	-		10 7470A	PPB	10/13/2023 14:16	Std
ICV18	ICV18	16030	3.7463 -		7470A	PPB	10/13/2023 14:19	SMPL
ICB18	ICB18	-20	0.0326 -		7470A	PPB	10/13/2023 14:24	SMPL
CCV01	CCV01	21132	4.9268 -		7470A	PPB	10/13/2023 14:26	SMPL
CCB01	CCB01	-528	-0.085 -		7470A	PPB	10/13/2023 14:29	SMPL
CRA	CRA	631	0.1832 -		7470A	PPB	10/13/2023 14:31	SMPL
HighStd	HighStd	43220	10.0376 -		7470A	PPB	10/13/2023 14:33	SMPL
ChkStd	ChkStd	30451	7.083 -		7470A	PPB	10/13/2023 14:35	SMPL
PB156355BL	PBW	-180	-0.0045 -		7470A	PPB	10/13/2023 14:40	SMPL
PB156355BS	LCSW	16285	3.8053 -		7470A	PPB	10/13/2023 14:43	SMPL
O4818-01	COMP-1	-474	-0.0725 -		7470A	PPB	10/13/2023 14:45	SMPL
O4818-01DUP	COMP-1DUP	-226	-0.0151 -		7470A	PPB	10/13/2023 14:47	SMPL
O4818-01MSD	COMP-1MSD	16278	3.8036 -		7470A	PPB	10/13/2023 14:52	SMPL
O4818-01MS	COMP-1MS	13933	3.261 -		7470A	PPB	10/13/2023 14:56	SMPL
O4818-02	5887	-271	-0.0255 -		7470A	PPB	10/13/2023 14:59	SMPL
PB156356BS	LCSW	0	0.0372 -		7470A	PPB	10/13/2023 15:32	SMPL
PB156356BS	LCSW	16379	3.827 -		7470A	PPB	10/13/2023 15:35	SMPL
O4844-02	OR-603-COMP-05	0	0.0372 -		7470A	PPB	10/13/2023 15:37	SMPL
CCV02	CCV02	20516	4.7842 -		7470A	PPB	10/13/2023 15:42	SMPL
CCB02	CCB02	-534	-0.0864 -		7470A	PPB	10/13/2023 15:45	SMPL
O4844-14	OR-603-COMP-08	-222	-0.0142 -		7470A	PPB	10/13/2023 15:47	SMPL
O4858-03	WC-1	-153	0.0018 -		7470A	PPB	10/13/2023 15:49	SMPL
O4858-06	WC-2	-113	0.011 -		7470A	PPB	10/13/2023 15:52	SMPL
O4858-09	WC-3	-104	0.0131 -		7470A	PPB	10/13/2023 15:54	SMPL
O4858-09DUP	WC-3DUP	-92	0.0159 -		7470A	PPB	10/13/2023 15:56	SMPL
O4858-09MS	WC-3MS	17658	4.1229 -		7470A	PPB	10/13/2023 15:58	SMPL
O4858-09MSD	WC-3MSD	17887	4.1759 -		7470A	PPB	10/13/2023 16:01	SMPL
PB156357BL	PBW	-39	0.0282 -		7470A	PPB	10/13/2023 16:03	SMPL
PB156357BS	LCSW	17887	4.1759 -		7470A	PPB	10/13/2023 16:05	SMPL
O4836-04	DW-1	10975	2.5766 -		7470A	PPB	10/13/2023 16:07	SMPL
CCV03	CCV03	21321	4.9705 -		7470A	PPB	10/13/2023 16:10	SMPL
CCB03	CCB03	-27	0.0309 -		7470A	PPB	10/13/2023 16:12	SMPL
O4859-01	SVS-GROUNDWA1	0	0.0372 -		7470A	PPB	10/13/2023 16:14	SMPL
O4836-04	DW-1	0	0.0372 -		7470A	PPB	10/13/2023 16:20	SMPL
O4859-01	SVS-GROUNDWA1	-81	0.0185 -		7470A	PPB	10/13/2023 16:22	SMPL
O4859-01DUP	SVS-GROUNDWA1	-8	0.0353 -		7470A	PPB	10/13/2023 16:24	SMPL
O4859-01MS	SVS-GROUNDWA1	14246	3.3335 -		7470A	PPB	10/13/2023 16:27	SMPL
O4859-01MSD	SVS-GROUNDWA1	12961	3.0361 -		7470A	PPB	10/13/2023 16:29	SMPL
PB156377BL	PBW	-135	0.006 -		7470A	PPB	10/13/2023 16:31	SMPL
PB156377BS	LCSW	18311	4.274 -		7470A	PPB	10/13/2023 16:33	SMPL
O4832-01	D3254	-108	0.0122 -		7470A	PPB	10/13/2023 16:36	SMPL

INSTRUMENT ID : CV1

O4837-01	2934-2940	-117	0.0101 -	7470A	PPB	10/13/2023 16:38 SMPL
O4837-01DUP	2934-2940DUP	-63	0.0226 -	7470A	PPB	10/13/2023 16:40 SMPL
O4837-01MS	2934-2940MS	14123	3.305 -	7470A	PPB	10/13/2023 16:42 SMPL
CCV04	CCV04	20972	4.8898 -	7470A	PPB	10/13/2023 16:45 SMPL
CCB04	CCB04	12	0.04 -	7470A	PPB	10/13/2023 16:47 SMPL
O4837-01MSD	2934-2940MSD	14699	3.4383 -	7470A	PPB	10/13/2023 16:49 SMPL
O4837-03	2904	-256	-0.022 -	7470A	PPB	10/13/2023 16:51 SMPL
PB156270TB		-99	0.0143 -	7470A	PPB	10/13/2023 16:54 SMPL
PB156318TB		-67	0.0217 -	7470A	PPB	10/13/2023 16:56 SMPL
PB156283TB		-60	0.0233 -	7470A	PPB	10/13/2023 16:58 SMPL
CCV05	CCV05	21985	5.1241 -	7470A	PPB	10/13/2023 17:00 SMPL
CCB05	CCB05	19	0.0416 -	7470A	PPB	10/13/2023 17:03 SMPL
O4818-01LX5		0	0.0372 -	7470A	PPB	10/13/2023 17:05 SMPL
O4818-03	5880	173	0.0772 -	7470A	PPB	10/13/2023 17:06 SMPL
O4820-01	M-092023-103-03	-69	0.0212 -	7470A	PPB	10/13/2023 17:08 SMPL
O4829-01	374	-65	0.0222 -	7470A	PPB	10/13/2023 17:10 SMPL
O4837-02	2941-2942	-41	0.0277 -	7470A	PPB	10/13/2023 17:12 SMPL
O4837-04	2919	-56	0.0242 -	7470A	PPB	10/13/2023 17:15 SMPL
PB156356BL	PBW	28	0.0437 -	7470A	PPB	10/13/2023 17:17 SMPL
PB156356BS	LCSW	18119	4.2296 -	7470A	PPB	10/13/2023 17:19 SMPL
O4844-02	OR-603-COMP-05	-226	-0.0151 -	7470A	PPB	10/13/2023 17:21 SMPL
O4844-06	OR-603-COMP-06	-96	0.015 -	7470A	PPB	10/13/2023 17:24 SMPL
O4844-10	OR-603-COMP-07	-72	0.0205 -	7470A	PPB	10/13/2023 17:26 SMPL
CCV06	CCV06	21693	5.0566 -	7470A	PPB	10/13/2023 17:28 SMPL
CCB06	CCB06	-199	-0.0088 -	7470A	PPB	10/13/2023 17:30 SMPL
O4844-14	OR-603-COMP-08	0	0.0372 -	7470A	PPB	10/13/2023 17:33 SMPL
O4818-01LX5		20	0.0418 -	7470A	PPB	10/13/2023 17:33 SMPL
O4818-01A		17692	4.1308 -	7470A	PPB	10/13/2023 17:35 SMPL
O4858-09LX5		-117	0.0101 -	7470A	PPB	10/13/2023 17:38 SMPL
O4858-09A		18076	4.2197 -	7470A	PPB	10/13/2023 17:40 SMPL
O4859-01LX5		-105	0.0129 -	7470A	PPB	10/13/2023 17:42 SMPL
O4859-01A		18389	4.2921 -	7470A	PPB	10/13/2023 17:45 SMPL
O4837-01LX5		-152	0.002 -	7470A	PPB	10/13/2023 17:47 SMPL
O4837-01A		18559	4.3314 -	7470A	PPB	10/13/2023 17:49 SMPL
CCV07	CCV07	22358	5.2104 -	7470A	PPB	10/13/2023 17:51 SMPL
CCB07	CCB07	626	0.182 -	7470A	PPB	10/13/2023 17:54 SMPL
CCB07	CCB07	-2	0.0367 -	7470A	PPB	10/13/2023 17:56 SMPL