



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB127795

Analyst By : jignesh

Supervisor Review By : Sohil

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	10/13/2023	16:10
2	CAL2	1	Water	NA	NA	20.6	7.00	10/13/2023	16:11
3	CAL3	1	Water	NA	NA	20.7	10.02	10/13/2023	16:15
4	ICV	1	Water	NA	NA	20.6	7.00	10/13/2023	16:20
5	CCV1	1	Water	NA	NA	20.6	2.01	10/13/2023	16:22
6	O4859-01	1	Water	NA	NA	20.1	7.19	10/13/2023	16:25
7	O4884-01	1	Water	NA	NA	20.1	6.46	10/13/2023	16:30
8	O4884-01DUP	1	Water	NA	NA	20.2	6.47	10/13/2023	16:31
9	CCV2	1	Water	NA	NA	20.7	12.02	10/13/2023	16:35

JB127795

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH O4869

WorkList ID : 174572

Department : Wet-Chemistry

Date : 10-13-2023 15:57:11

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4859-01	SVS-GROUNDWATER	Water	pH	Cool 4 deg C	PSEG03	E31	10/11/2023	9040C
O4884-01	026-OUTFALL	Water	pH	Cool 4 deg C	MINO01	E31	10/12/2023	9040C

Date/Time 10-13-23 16:00
Raw Sample Received by: JD (CSC)
Raw Sample Relinquished by: JT (CSC)

Date/Time 10-13-23 17:30
Raw Sample Received by: JT (CSC)
Raw Sample Relinquished by: JD (CSC)