



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Sohil

Run Number: LB127796

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	10/13/2023	16:10
2	CAL2	1	Water	NA	NA	20.6	7.00	10/13/2023	16:11
3	CAL3	1	Water	NA	NA	20.7	10.02	10/13/2023	16:15
4	ICV	1	Water	NA	NA	20.6	7.00	10/13/2023	16:20
5	CCV1	1	Water	NA	NA	20.7	2.01	10/13/2023	16:40
6	O4881-01	1	Solid	20.02	20	20.6	10.83	10/13/2023	16:45
7	O4881-01DUP	1	Solid	20.03	20	20.7	10.84	10/13/2023	16:46
8	CCV2	1	Water	NA	NA	20.7	12.02	10/13/2023	16:50

WORKLIST(Hardcopy Internal Chain)

VB 23716

WorkList Name : PH 04881

WorkList ID : 174571

Department : Wet-Chemistry

Date : 10-13-2023 15:55:24

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4881-01	A455-SLUDGE	Solid	pH	Cool 4 deg C	PSEG03	E31	10/13/2023	9045D

Date/Time 10.13.23 16:00

Raw Sample Received by: JO (CSC)

Raw Sample Relinquished by: JT (CSC)

Date/Time 10.13.23 17:30

Raw Sample Received by: JT (CSC)

Raw Sample Relinquished by: JO (CSC)