



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB127797

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	10/13/2023	09:30
2	O4859-01		1	12	100.00	>212.0	>212.0	10/13/2023	10:00
3	O4859-01DUP		1	12	100.00	>212.0	>212.0	10/13/2023	10:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6127797

WorkList Name : FP-W-10-12 WorkList ID : 174538 Department : Wet-Chemistry Date : 10-12-2023 15:31:44

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4859-01 P	SVS-GROUNDWATER	Water	Flash Point	Cool 4 deg C	PSEG03	E31	10/11/2023	1010B

Date/Time 10/13/2023 09:20
 Raw Sample Received by: Pri CWC
 Raw Sample Relinquished by: Pri CWC

Date/Time 10/13/2023 10:40
 Raw Sample Received by: Pri CWC
 Raw Sample Relinquished by: Pri CWC