

LB127815


SFAM01.1

INSTRUMENT ID: CV 1

Linear

 μ Abs.:

53215



A= 0.0000e+000

B= 1.8682e-004 slope

C= 1.6336e-002 y-intercept

Rho= 0.9997391

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.023	0.023	38	0.000	38					
0.05	0.050					473	0				
0.20	0.200	0.232	0.032	1154	0.0 %	1154					16
2.50	2.500	2.471	-0.029	13140	0.0 %	13140					11
5.00	5.000	4.869	-0.131	25976	0.0 %	25976					3
7.50	7.500	7.646	0.146	40840	0.0 %	40840					2
10.0	10.000	9.958	-0.042	53215	0.0 %	53215					0

LB127815

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	38	-		0 SFAM01.1	PPB	10/16/2023 13:42	Std
	0.2 S01	1154	-		0.2 SFAM01.1	PPB	10/16/2023 13:44	Std
	2.5 S02	13140	-		2.5 SFAM01.1	PPB	10/16/2023 13:46	Std
	5 S03	25976	-		5 SFAM01.1	PPB	10/16/2023 13:51	Std
	7.5 S04	40840	-		7.5 SFAM01.1	PPB	10/16/2023 13:54	Std
	10 S05	53215	-		10 SFAM01.1	PPB	10/16/2023 13:56	Std
ICV019	ICV019	20753	3.8935 -		SFAM01.1	PPB	10/16/2023 13:59	SMPL
ICB019	ICB019	-333	-0.0459 -		SFAM01.1	PPB	10/16/2023 14:01	SMPL
CCV008	CCV008	26016	4.8767 -		SFAM01.1	PPB	10/16/2023 14:03	SMPL
CCB008	CCB008	-317	-0.0429 -		SFAM01.1	PPB	10/16/2023 14:05	SMPL
PB156406BL	PBW406	-77	0.002 -		SFAM01.1	PPB	10/16/2023 14:08	SMPL
O4901-01	MCOAT3	29	0.0218 -		SFAM01.1	PPB	10/16/2023 14:10	SMPL
O4901-02	MCOAT4	25	0.021 -		SFAM01.1	PPB	10/16/2023 14:12	SMPL
O4901-03	MCOAT5	46	0.0249 -		SFAM01.1	PPB	10/16/2023 14:14	SMPL
O4901-04	MCOAT6	38	0.0234 -		SFAM01.1	PPB	10/16/2023 14:17	SMPL
O4901-05	MCOAT7	20	0.0201 -		SFAM01.1	PPB	10/16/2023 14:19	SMPL
O4901-06	MCOAW2	64	0.0283 -		SFAM01.1	PPB	10/16/2023 14:21	SMPL
O4901-07	MCOAW3	32	0.0223 -		SFAM01.1	PPB	10/16/2023 14:23	SMPL
O4901-08	MCOAW3D	45	0.0247 -		SFAM01.1	PPB	10/16/2023 14:26	SMPL
O4901-09	MCOAW3S	6231	1.1804 -		SFAM01.1	PPB	10/16/2023 14:28	SMPL
PB156407BL	PBW407	-30	0.0107 -		SFAM01.1	PPB	10/16/2023 14:30	SMPL
O4902-01	MCOAT8	44	0.0246 -		SFAM01.1	PPB	10/16/2023 14:32	SMPL
O4902-02	MCOAT9	49	0.0255 -		SFAM01.1	PPB	10/16/2023 14:35	SMPL
O4902-03	MCOAW0	42	0.0242 -		SFAM01.1	PPB	10/16/2023 14:37	SMPL
O4902-04	MCOAW1	48	0.0253 -		SFAM01.1	PPB	10/16/2023 14:39	SMPL
O4902-05	MCOAW4	52	0.0261 -		SFAM01.1	PPB	10/16/2023 14:41	SMPL
O4902-06	MCOAW5	56	0.0268 -		SFAM01.1	PPB	10/16/2023 14:44	SMPL
O4902-07	MCOAW5D	66	0.0287 -		SFAM01.1	PPB	10/16/2023 14:46	SMPL
O4902-08	MCOAW5S	5614	1.0652 -		SFAM01.1	PPB	10/16/2023 14:51	SMPL
O4902-09	MCOAW6	-53	0.0064 -		SFAM01.1	PPB	10/16/2023 14:53	SMPL
CCV009	CCV009	24666	4.6245 -		SFAM01.1	PPB	10/16/2023 14:55	SMPL
CCB009	CCB009	-227	-0.0261 -		SFAM01.1	PPB	10/16/2023 14:58	SMPL