


LB127819

SFAM01.1

INSTRUMENT ID: CV1

 μ Abs.:

51421



Linear

A= 0.0000e+000

B= 1.9266e-004 slope

C= 3.1550e-002 $y = mx + c$

Rho= 0.9996591

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	O/P
0.00	0.000	0.030	0.030	-9	0.000	-9					0/01
0.05	0.050					473	0				19
0.20	0.200	0.237	0.037	1064	0.0 %	1064					-2
2.50	2.500	2.445	-0.055	12527	0.0 %	12527					-3
5.00	5.000	4.874	-0.126	25137	0.0 %	25137					2
7.50	7.500	7.676	0.176	39677	0.0 %	39677					-1
10.0	10.000	9.938	-0.062	51421	0.0 %	51421					

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	-9	-		0 SFAM01.1	PPB	10/16/2023 15:39	Std
	0.2 S01	1064	-		0.2 SFAM01.1	PPB	10/16/2023 15:41	Std
	2.5 S02	12527	-		2.5 SFAM01.1	PPB	10/16/2023 15:46	Std
	5 S03	25137	-		5 SFAM01.1	PPB	10/16/2023 15:49	Std
	7.5 S04	39677	-		7.5 SFAM01.1	PPB	10/16/2023 15:53	Std
	10 S05	51421	-		10 SFAM01.1	PPB	10/16/2023 15:55	Std
ICV020	ICV020	20494	3.9799 -		SFAM01.1	PPB	10/16/2023 15:58	SMPL
ICB020	ICB020	-304	-0.027 -		SFAM01.1	PPB	10/16/2023 16:00	SMPL
CCV010	CCV010	25656	4.9745 -		SFAM01.1	PPB	10/16/2023 16:03	SMPL
CCB010	CCB010	-334	-0.0328 -		SFAM01.1	PPB	10/16/2023 16:05	SMPL
PB156400BL	PBW400	-112	0.01 -		SFAM01.1	PPB	10/16/2023 16:07	SMPL
PB156400TB	LEB400	-48	0.0223 -		SFAM01.1	PPB	10/16/2023 16:10	SMPL
O4886-01	MCOM25	-19	0.0279 -		SFAM01.1	PPB	10/16/2023 16:12	SMPL
O4886-02	MCOM26	7	0.0329 -		SFAM01.1	PPB	10/16/2023 16:14	SMPL
O4886-03	MCOM27	31	0.0375 -		SFAM01.1	PPB	10/16/2023 16:16	SMPL
O4886-04	MCOM28	6	0.0327 -		SFAM01.1	PPB	10/16/2023 16:19	SMPL
O4886-05	MCOM28D	-1	0.0314 -		SFAM01.1	PPB	10/16/2023 16:21	SMPL
O4886-06	MCOM28S	5047	1.0039 -		SFAM01.1	PPB	10/16/2023 16:23	SMPL
PB156401BL	PBW401	-57	0.0206 -		SFAM01.1	PPB	10/16/2023 16:25	SMPL
PB156401TB	LEB401	1	0.0317 -		SFAM01.1	PPB	10/16/2023 16:28	SMPL
O4886-07	MCOM29	-6	0.0304 -		SFAM01.1	PPB	10/16/2023 16:30	SMPL
O4886-08	MCOM30	12	0.0339 -		SFAM01.1	PPB	10/16/2023 16:32	SMPL
O4886-09	MCOM31	2	0.0319 -		SFAM01.1	PPB	10/16/2023 16:34	SMPL
O4886-10	MCOM32	-7	0.0302 -		SFAM01.1	PPB	10/16/2023 16:37	SMPL
O4886-11	MCOM32D	-1	0.0314 -		SFAM01.1	PPB	10/16/2023 16:39	SMPL
O4886-12	MCOM32S	6139	1.2143 -		SFAM01.1	PPB	10/16/2023 16:41	SMPL
CCV011	CCV011	24134	4.6812 -		SFAM01.1	PPB	10/16/2023 16:43	SMPL
CCB011	CCB011	-251	-0.0168 -		SFAM01.1	PPB	10/16/2023 16:46	SMPL