



Analytical Summary Report

Analysis Method: 7196A

ANALYST: Rubina

Parameter: Hexavalent Chromium

SUPERVISOR REVIEW BY: Iwona

Run Number: LB127826

pH Meter ID: WC pH Meter-1

Reagent/Standard	Lot/Log #
hexavalent chromium color reagent	WP104678
5N sulfuric acid	WP102263
HNO3 Hex-Chrome, 5M	WP102262
Hexchrome Cleaning Solution	WP103556

Intercept: 0.0012

Slope: 0.7667

Regression: 0.999997

Seq	Lab ID	True Value (mg/l)	DF	Initial Vol (ml)	Final Vol (ml)	pH HNO3	pH H2SO4	Absorb.at 540nm		Absorbance Difference	Result (mg/L)	%D	Anal Date	Anal Time
								Backgrnd	Color					
1	CAL1	0	1	100	100	7.26	1.78	0.000	0.000	0.000	-0.00		10/17/2023	12:00
2	CAL2	0.01	1	100	100	7.37	1.88	0.000	0.009	0.009	0.010	0	10/17/2023	12:01
3	CAL3	0.025	1	100	100	7.36	1.85	0.000	0.020	0.020	0.024	-4	10/17/2023	12:02
4	CAL4	0.05	1	100	100	7.38	1.86	0.000	0.040	0.040	0.050	0	10/17/2023	12:03
5	CAL5	0.1	1	100	100	7.38	1.88	0.000	0.079	0.079	0.101	1	10/17/2023	12:04
6	CAL6	0.5	1	100	100	7.40	1.91	0.000	0.384	0.384	0.499	-0.2	10/17/2023	12:05
7	CAL7	1	1	100	100	7.37	1.87	0.000	0.768	0.768	1.000	0	10/17/2023	12:06



Analytical Summary Report

Analysis Method: 7196A

ANALYST: Rubina

Parameter: Hexavalent Chromium

SUPERVISOR REVIEW BY: Iwona

Run Number: LB127826

pH Meter ID: WC pH Meter-1

Seq	Lab ID	True Value	DF	Initial Vol (ml/gm)	Final Vol (ml)	pH HN03	pH H2SO4	Absorb.at540nm		Absorbance Difference	Intermediate Result (mg/L)	Anal Date	Anal Time
								Backgrnd	Color				
1	ICV	0.5	1	100	100	7.41	1.92	0.000	0.383	0.383	0.498	10/17/2023	12:07
2	ICB		1	100	100	7.25	1.77	0.000	0.001	0.001	0.000	10/17/2023	12:08
3	CCV1	0.5	1	100	100	7.44	1.93	0.000	0.385	0.385	0.501	10/17/2023	12:09
4	CCB1		1	100	100	7.28	1.78	0.000	0.001	0.001	0.000	10/17/2023	12:10
5	RL Check	0.01	1	100	100	7.40	1.90	0.000	0.009	0.009	0.010	10/17/2023	12:11
6	PB156388BL		1	2.50	100	7.26	1.76	0.000	0.001	0.001	0.000	10/17/2023	12:12
7	PB156388BS	20	1	2.50	100	7.41	1.92	0.000	0.384	0.384	0.499	10/17/2023	12:13
8	O4887-01		1	2.52	100	7.62	2.08	0.007	0.008	0.001	0.000	10/17/2023	12:14
9	O4888-01		1	2.54	100	7.60	2.12	0.010	0.072	0.062	0.079	10/17/2023	12:15
10	O4890-02		1	2.55	100	7.58	2.06	0.010	0.016	0.006	0.006	10/17/2023	12:16
11	O4892-01		1	2.53	100	7.60	2.18	0.006	0.006	0.000	-0.002	10/17/2023	12:17
12	O4892-03		1	2.54	100	7.52	2.10	0.010	0.012	0.002	0.001	10/17/2023	12:18
13	O4892-05		1	2.52	100	7.56	2.06	0.003	0.008	0.005	0.005	10/17/2023	12:19
14	O4895-01		1	2.54	100	7.62	2.16	0.009	0.010	0.001	0.000	10/17/2023	12:20
15	O4895-02		1	2.52	100	7.54	2.20	0.003	0.007	0.004	0.004	10/17/2023	12:21
16	CCV2	0.5	1	100	100	7.42	1.94	0.000	0.384	0.384	0.499	10/17/2023	12:22
17	CCB2		1	100	100	7.27	1.78	0.000	0.001	0.001	0.000	10/17/2023	12:23
18	O4895-03		1	2.54	100	7.52	2.08	0.003	0.008	0.005	0.005	10/17/2023	12:24
19	O4908-01		1	2.55	100	7.52	2.14	0.005	0.006	0.001	0.000	10/17/2023	12:25
20	O4908-04		1	2.53	100	7.58	2.10	0.010	0.010	0.000	-0.002	10/17/2023	12:26
21	O4908-07		1	2.52	100	7.54	2.06	0.002	0.004	0.002	0.001	10/17/2023	12:27
22	O4908-07DU		1	2.53	100	7.56	2.08	0.003	0.004	0.001	0.000	10/17/2023	12:28
23	O4908-07MS	40	2	2.52	100	7.54	2.08	0.003	0.335	0.332	0.431	10/17/2023	12:29
24	O4908-07MS	1284	40	2.53	100	7.52	2.06	0.000	0.567	0.567	0.738	10/17/2023	12:30
25	O4908-07MS	40	2	2.53	100	7.56	2.10	0.002	0.385	0.383	0.498	10/17/2023	12:31
26	CCV3	0.5	1	100	100	7.44	1.93	0.000	0.383	0.383	0.498	10/17/2023	12:32
27	CCB3		1	100	100	7.28	1.78	0.000	0.001	0.001	0.000	10/17/2023	12:33