

LB127827


7471B

INSTRUMENT ID: CVJ

Linear

 μ Abs.:

61181



A= 0.0000e+000

B= 1.6202e-004

C= 1.2925e-002

Rho= 0.9998456

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	1/11
0.0	0.000	0.012	0.012	-7	0.000	-7					
0.2	0.200	0.211	0.011	1225	0.0 %	1225					
2.5	2.500	2.440	-0.060	14982	0.0 %	14982					
5.0	5.000	4.986	-0.014	30696	0.0 %	30696					
7.5	7.500	7.625	0.125	46983	0.0 %	46983					
10.0	10.000	9.925	-0.075	61181	0.0 %	61181					

LB127837 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 17 Oct 2023 11:52:33

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-7	-	0.0000	7471B	PPB	17 Oct 2023 11:55:30	S	Std
0.2 - 1	51.2	1225	-	0.2000	7471B	PPB	17 Oct 2023 11:57:46	S	Std
2.5 - 1	52.5	14982	-	2.5000	7471B	PPB	17 Oct 2023 12:00:04	S	Std
5.0 - 1	55	30696	-	5.0000	7471B	PPB	17 Oct 2023 12:02:20	S	Std
7.5 - 1	57.5	46983	-	7.5000	7471B	PPB	17 Oct 2023 12:04:37	S	Std
10.0 - 1	510	61181	-	10.0000	7471B	PPB	17 Oct 2023 12:09:33	S	Std
ICV22 - 1	ICV22	26839	4.3613	-	7471B	PPB	17 Oct 2023 12:12:21	U	SMPL
ICB22 - 1	ICB22	-590	-0.0827	-	7471B	PPB	17 Oct 2023 12:14:37	U	SMPL
CCV15 - 1	CCV15	31590	5.1311	-	7471B	PPB	17 Oct 2023 12:16:53	U	SMPL
CCB15 - 1	CCB15	-573	-0.0799	-	7471B	PPB	17 Oct 2023 12:19:09	U	SMPL
CRA - 1	CRA	1090	0.1895	-	7471B	PPB	17 Oct 2023 12:21:26	U	SMPL
HighStd - 1	HighStd	62118	10.0771	-	7471B	PPB	17 Oct 2023 12:23:44	U	SMPL
ChkStd - 1	ChkStd	45255	7.3450	-	7471B	PPB	17 Oct 2023 12:25:59	U	SMPL
PB156399BL - 1	PBS	-226	-0.0237	-	7471B	PPB	17 Oct 2023 12:30:51	U	SMPL
PB156399BS - 1	LCSS	26642	4.3294	-	7471B	PPB	17 Oct 2023 12:33:06	U	SMPL
O4878-01 - 1	72-11995	6573	1.0779	-	7471B	PPB	17 Oct 2023 12:42:52	U	SMPL
O4878-01DUP - 1	72-11995DUP	5578	0.9167	-	7471B	PPB	17 Oct 2023 12:45:07	U	SMPL
O4878-01MS - 1	72-11995MS	33056	5.3686	-	7471B	PPB	17 Oct 2023 12:47:25	U	SMPL
O4878-01MSD - 1	72-11995MSD	33182	5.3890	-	7471B	PPB	17 Oct 2023 12:52:23	U	SMPL
O4878-03 - 1	222	1895	0.3199	-	7471B	PPB	17 Oct 2023 12:54:38	U	SMPL
CCV16 - 1	CCV16	31905	5.1821	-	7471B	PPB	17 Oct 2023 12:56:54	U	SMPL
CCB16 - 1	CCB16	-222	-0.0230	-	7471B	PPB	17 Oct 2023 12:59:09	U	SMPL
O4878-05 - 1	283	37973	6.1652	-	7471B	PPB	17 Oct 2023 13:01:25	U	SMPL
O4882-02 - 1	VNJ-218	3165	0.5257	-	7471B	PPB	17 Oct 2023 13:03:40	U	SMPL
O4882-04 - 1	VNJ-228	3243	0.5383	-	7471B	PPB	17 Oct 2023 13:05:58	U	SMPL
O4887-01 - 1	SU-622-COMP-04	2178	0.3658	-	7471B	PPB	17 Oct 2023 13:08:14	U	SMPL
O4888-01 - 1	TR-304-COMP-01	11375	1.8559	-	7471B	PPB	17 Oct 2023 13:10:30	U	SMPL
O4890-02 - 1	362	3697	0.6119	-	7471B	PPB	17 Oct 2023 13:12:46	U	SMPL
O4892-01 - 1	EO-02-10132023-A	7148	1.1710	-	7471B	PPB	17 Oct 2023 13:15:02	U	SMPL
O4892-03 - 1	EO-03-10132023-A	1646	0.2796	-	7471B	PPB	17 Oct 2023 13:17:18	U	SMPL
O4892-05 - 1	EO-03-10132023-B	1606	0.2731	-	7471B	PPB	17 Oct 2023 13:19:35	U	SMPL
O4895-01 - 1	A-1	4604	0.7589	-	7471B	PPB	17 Oct 2023 13:21:51	U	SMPL
CCV17 - 1	CCV17	32091	5.2122	-	7471B	PPB	17 Oct 2023 13:24:06	U	SMPL
CCB17 - 1	CCB17	-88	-0.0013	-	7471B	PPB	17 Oct 2023 13:26:22	U	SMPL
O4895-02 - 1	A-2	6265	1.0280	-	7471B	PPB	17 Oct 2023 13:28:37	U	SMPL
O4895-03 - 1	A-3	4516	0.7446	-	7471B	PPB	17 Oct 2023 13:30:52	U	SMPL
O4878-01LX5 - 1		880	0.1555	-	7471B	PPB	17 Oct 2023 13:33:08	U	SMPL
O4878-01A - 1		31105	5.0525	-	7471B	PPB	17 Oct 2023 13:35:23	U	SMPL
CCV18 - 1	CCV18	32599	5.2945	-	7471B	PPB	17 Oct 2023 13:37:39	U	SMPL
CCB18 - 1	CCB18	-130	-0.0081	-	7471B	PPB	17 Oct 2023 13:39:55	U	SMPL