

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : RM Instrument ID : Konelab

10/17/2023 11:47

Test: Ammonia-N

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	0.960	0.0	0.226	
ICB1	0.009	0.0	0.026	
CCV1	0.987	0.0	0.231	
CCB1	0.010	0.0	0.026	
RL CHECK	0.107	0.0	0.046	
PB156425BL	0.010	0.0	0.026	
PB156425BS	0.965	0.0	0.227	
O4894-01	2.515	0.0	0.552	Test limit high
O4894-01DU P	3.061	0.0	0.667	Test limit high
O4894-01MS	3.438	0.0	0.746	Test limit high
O4894-01MSD	3.468	0.0	0.753	Test limit high
O4894-02	2.584	0.0	0.567	Test limit high
O4909-01	14.330	0.0	3.034	Init abs., Test limit hig
O4911-01	27.685	0.0	5.840	Init abs., Test limit hig
CCV2	0.936	0.0	0.221	
CCB2	0.017	0.0	0.027	
O4911-02	0.015	0.0	0.027	
PB156424BL	0.019	0.0	0.028	
PB156424BS	0.989	0.0	0.232	
O4895-01	0.039	0.0	0.032	
O4895-01DUP	0.049	0.0	0.034	
O4895-01MS	0.959	0.0	0.225	
O4895-01MSD	0.956	0.0	0.225	
O4895-02	0.021	0.0	0.028	
O4895-03	0.022	0.0	0.028	
CCV3	1.010	0.0	0.236	
CCB3	0.019	0.0	0.028	
O4699-03	0.085	0.0	0.042	
O4894-01DLX2	1.249	0.0	0.286	
O4894-02DLX2	1.276	0.0	0.292	
O4909-01DLX10	1.609	0.0	0.362	
O4911-01DLX50	6.836	0.0	1.460	Test limit high
CCV4	1.016	0.0	0.237	
CCB4	0.013	0.0	0.027	
O4911-01DL2X250	1.470	0.0	0.333	
CCV5	0.993	0.0	0.233	
CCB5	0.013	0.0	0.027	

N 37
Mean 2.155
SD 5.0289
CV% 233.35

107°C 50-150

10/17/2023
RM

Aquakem v. 7.2AQ1

Results from time period:

Tue Oct 17 08:51:08 2023

Tue Oct 17 11:46:05 2023

Sample Id	Sam/Ctr/c/	Test short / Test type	Result	Result unit	Result date and time	Stat
0.0PPM	A	Ammonia-I P	0.0137	mg/l	10/17/2023 8:51:08	
0.1PPM	A	Ammonia-I P	0.1095	mg/l	10/17/2023 8:51:09	
0.2PPM	A	Ammonia-I P	0.2065	mg/l	10/17/2023 8:51:10	
0.4PPM	A	Ammonia-I P	0.3966	mg/l	10/17/2023 8:51:11	
1.0PPM	A	Ammonia-I P	0.9818	mg/l	10/17/2023 8:51:12	
1.3PPM	A	Ammonia-I P	1.2829	mg/l	10/17/2023 8:51:13	
2.0PPM	A	Ammonia-I P	2.0423	mg/l	10/17/2023 8:51:14	
ICV1	S	Ammonia-I P	0.9596	mg/l	10/17/2023 10:05:37	
ICB1	S	Ammonia-I P	0.0089	mg/l	10/17/2023 10:05:39	
CCV1	S	Ammonia-I P	0.9871	mg/l	10/17/2023 10:05:41	
CCB1	S	Ammonia-I P	0.0099	mg/l	10/17/2023 10:05:43	
RL CHECK	S	Ammonia-I P	0.1071	mg/l	10/17/2023 10:05:45	
PB156425BL	S	Ammonia-I P	0.0097	mg/l	10/17/2023 10:05:47	
PB156425BS	S	Ammonia-I P	0.9654	mg/l	10/17/2023 10:16:23	
O4894-01	S	Ammonia-I P	2.5146	mg/l	10/17/2023 10:16:24	
O4894-01DU P	S	Ammonia-I P	3.0612	mg/l	10/17/2023 10:16:25	
O4894-01MS	S	Ammonia-I P	3.4384	mg/l	10/17/2023 10:16:26	
O4894-01MSD	S	Ammonia-I P	3.4684	mg/l	10/17/2023 10:16:27	
O4894-02	S	Ammonia-I P	2.5845	mg/l	10/17/2023 10:16:30	
O4909-01	S	Ammonia-I P	14.3301	mg/l	10/17/2023 10:16:31	
O4911-01	S	Ammonia-I P	27.6851	mg/l	10/17/2023 10:16:32	
CCV2	S	Ammonia-I P	0.9358	mg/l	10/17/2023 10:27:06	
CCB2	S	Ammonia-I P	0.0168	mg/l	10/17/2023 10:27:08	
O4911-02	S	Ammonia-I P	0.0148	mg/l	10/17/2023 10:27:10	
PB156424BL	S	Ammonia-I P	0.0187	mg/l	10/17/2023 10:27:12	
PB156424BS	S	Ammonia-I P	0.9891	mg/l	10/17/2023 10:27:14	
O4895-01	S	Ammonia-I P	0.0386	mg/l	10/17/2023 10:37:46	
O4895-01DUP	S	Ammonia-I P	0.0489	mg/l	10/17/2023 10:37:47	
O4895-01MS	S	Ammonia-I P	0.9588	mg/l	10/17/2023 10:37:48	
O4895-01MSD	S	Ammonia-I P	0.9556	mg/l	10/17/2023 10:37:49	
O4895-02	S	Ammonia-I P	0.0214	mg/l	10/17/2023 10:37:52	
O4895-03	S	Ammonia-I P	0.0216	mg/l	10/17/2023 10:37:53	
CCV3	S	Ammonia-I P	1.0101	mg/l	10/17/2023 10:37:55	
CCB3	S	Ammonia-I P	0.0187	mg/l	10/17/2023 10:37:57	
O4699-03	S	Ammonia-I P	0.0852	mg/l	10/17/2023 11:21:03	
O4894-01DLX2	S	Ammonia-I P	1.249	mg/l	10/17/2023 11:21:05	
O4894-02DLX2	S	Ammonia-I P	1.2764	mg/l	10/17/2023 11:21:06	
O4909-01DLX10	S	Ammonia-I P	1.6086	mg/l	10/17/2023 11:21:07	
O4911-01DLX50	S	Ammonia-I P	6.8358	mg/l	10/17/2023 11:21:09	
CCV4	S	Ammonia-I P	1.016	mg/l	10/17/2023 11:21:11	
CCB4	S	Ammonia-I P	0.0133	mg/l	10/17/2023 11:21:13	
O4911-01DL2X250	S	Ammonia-I P	1.4696	mg/l	10/17/2023 11:46:01	
CCV5	S	Ammonia-I P	0.9929	mg/l	10/17/2023 11:46:03	
CCB5	S	Ammonia-I P	0.0129	mg/l	10/17/2023 11:46:05	

Calibration results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

10/17/2023 9:18

Reviewed by : RM

Instrument ID : Konelab


Test Ammonia-N

Accepted 10/17/2023 9:18

Factor 4.76
Bias 0.024

Coeff. of det. 0.998521

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.00PPM	0.027	0.0137	0.0000	-
2	NH3-2PPM	0.047	0.1095	0.1000	9.5
3	NH3-2PPM	0.067	0.2065	0.2000	3.2
4	NH3-2PPM	0.107	0.3966	0.4000	-0.9
5	NH3-2PPM	0.230	0.9818	1.0000	-1.8
6	NH3-2PPM	0.293	1.2829	1.3333	-1.3
7	NH3-2PPM	0.453	2.0423	2.0000	2.1

10/17/2023
RM