



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB127835

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.5	4.01	10/17/2023	13:25
2	CAL2	1	Water	NA	NA	20.6	7.00	10/17/2023	13:26
3	CAL3	1	Water	NA	NA	20.5	10.02	10/17/2023	13:28
4	ICV	1	Water	NA	NA	20.6	6.99	10/17/2023	13:30
5	CCV1	1	Water	NA	NA	20.5	2.01	10/17/2023	13:33
6	O4916-01	1	Water	NA	NA	20.1	5.28	10/17/2023	13:38
7	O4916-01DUP	1	Water	NA	NA	20.2	5.29	10/17/2023	13:39
8	O4931-01	1	Water	NA	NA	20.6	5.09	10/17/2023	13:43
9	CCV2	1	Water	NA	NA	20.3	12.02	10/17/2023	13:50

13127835

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH 04931

WorkList ID : 174647

Department : Wet-Chemistry

Date : 10-17-2023 13:10:31

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4916-01	JS960	Water	pH	Cool 4 deg C	STER17	I11	10/30/2023	9040C
O4931-01	POOL-WATER	Water	pH	Cool 4 deg C	STER11	E31	10/13/2023	9040C
O4931-02	CONTROL-SAMPLE	Water	pH	Cool 4 deg C	STER11	E31	10/13/2023	9040C

10/17/2023

763
10/17/2023

Date/Time 10-17-23 13:20

Raw Sample Received by: 10661

Raw Sample Relinquished by: [Signature]

Date/Time 10-17-23 16:00

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: 10661