


LB127837

SFAM01.1

INSTRUMENT 15: CV1

 μ Abs.:

56253



Linear

A= 0.0000e+000

B= 1.7632e-004 slope

C= -1.7175e-002 y-intercept

Rho= 0.9997558

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	6/6/1
0.00	0.000	-0.015	-0.015	13	0.000	13					1
0.05	0.050	0.055	0.005	410	0.0 %	410	0				10
0.20	0.200	0.203	0.003	1249	0.0 %	1249					2
2.50	2.500	2.510	0.010	14334	0.0 %	14334					0
5.00	5.000	4.920	-0.080	28003	0.0 %	28003					2
7.50	7.500	7.675	0.175	43627	0.0 %	43627					2
10.0	10.000	9.901	-0.099	56253	0.0 %	56253					1

LB127837
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	13	-		0 SFAM01.1	PPB	10/17/2023 14:54	Std
	0.05 S01	410	-	0.05	SFAM01.1	PPB	10/17/2023 14:57	Std
	0.2 S02	1249	-	0.2	SFAM01.1	PPB	10/17/2023 14:59	Std
	2.5 S03	14334	-	2.5	SFAM01.1	PPB	10/17/2023 15:01	Std
	5 S04	28003	-	5	SFAM01.1	PPB	10/17/2023 15:04	Std
	7.5 S05	43627	-	7.5	SFAM01.1	PPB	10/17/2023 15:06	Std
	10 S06	56253	-	10	SFAM01.1	PPB	10/17/2023 15:26	Std
ICV023	ICV023	23095	4.0549 -		SFAM01.1	PPB	10/17/2023 15:28	SMPL
ICB023	ICB023	-227	-0.0572 -		SFAM01.1	PPB	10/17/2023 15:31	SMPL
CCV019	CCV019	29043	5.1036 -		SFAM01.1	PPB	10/17/2023 15:33	SMPL
CCB019	CCB019	-269	-0.0646 -		SFAM01.1	PPB	10/17/2023 15:35	SMPL
PB156456BL	PBS456	-103	-0.0353 -		SFAM01.1	PPB	10/17/2023 15:37	SMPL
O4683-01	MJUBK3	2289	0.3864 -		SFAM01.1	PPB	10/17/2023 15:40	SMPL
O4683-02	MJUBK4	886	0.139 -		SFAM01.1	PPB	10/17/2023 15:42	SMPL
O4683-03	MJUBK5	921	0.1452 -		SFAM01.1	PPB	10/17/2023 15:44	SMPL
O4683-04	MJUBK6	658	0.0988 -		SFAM01.1	PPB	10/17/2023 15:46	SMPL
O4683-05	MJUBK7	380	0.0498 -		SFAM01.1	PPB	10/17/2023 15:49	SMPL
O4683-06	MJUBK8	451	0.0623 -		SFAM01.1	PPB	10/17/2023 15:51	SMPL
O4683-07	MJUBK8D	347	0.044 -		SFAM01.1	PPB	10/17/2023 15:53	SMPL
O4683-08	MJUBK8S	14861	2.6031 -		SFAM01.1	PPB	10/17/2023 15:55	SMPL
O4683-09	MJUBK9	240	0.0251 -		SFAM01.1	PPB	10/17/2023 15:58	SMPL
O4683-10	MJUBL0	507	0.0722 -		SFAM01.1	PPB	10/17/2023 16:00	SMPL
O4683-11	MJUBL1	498	0.0706 -		SFAM01.1	PPB	10/17/2023 16:02	SMPL
O4683-12	MJUBL2	358	0.0459 -		SFAM01.1	PPB	10/17/2023 16:05	SMPL
O4683-13	MJUBL3	470	0.0657 -		SFAM01.1	PPB	10/17/2023 16:07	SMPL
O4683-14	MJUBL4	378	0.0495 -		SFAM01.1	PPB	10/17/2023 16:09	SMPL
O4683-15	MJUBL5	1177	0.1904 -		SFAM01.1	PPB	10/17/2023 16:11	SMPL
O4683-16	MJUBL6	1419	0.233 -		SFAM01.1	PPB	10/17/2023 16:14	SMPL
O4683-17	MJUBL7	2453	0.4153 -		SFAM01.1	PPB	10/17/2023 16:16	SMPL
O4683-18	MJUBL8	2460	0.4166 -		SFAM01.1	PPB	10/17/2023 16:18	SMPL
O4683-19	MJUBL9	1634	0.2709 -		SFAM01.1	PPB	10/17/2023 16:20	SMPL
CCV020	CCV020	29940	5.2618 -		SFAM01.1	PPB	10/17/2023 16:23	SMPL
CCB020	CCB020	-82	-0.0316 -		SFAM01.1	PPB	10/17/2023 16:25	SMPL
O4683-20	MJUBM0	2115	0.3557 -		SFAM01.1	PPB	10/17/2023 16:27	SMPL
O4683-21	MJUBM2	568	0.083 -		SFAM01.1	PPB	10/17/2023 16:29	SMPL
O4683-22	MJUBM3	725	0.1107 -		SFAM01.1	PPB	10/17/2023 16:32	SMPL
PB156457BL	PBS457	-30	-0.0225 -		SFAM01.1	PPB	10/17/2023 16:34	SMPL
O4710-01	MJUBN5	549	0.0796 -		SFAM01.1	PPB	10/17/2023 16:36	SMPL
O4710-02	MJUBN6	342	0.0431 -		SFAM01.1	PPB	10/17/2023 16:39	SMPL
O4710-03	MJUBN7	15051	2.6366 -		SFAM01.1	PPB	10/17/2023 16:41	SMPL
O4710-04	MJUBN8	390	0.0516 -		SFAM01.1	PPB	10/17/2023 16:43	SMPL
O4710-05	MJUBN9	409	0.0549 -		SFAM01.1	PPB	10/17/2023 16:45	SMPL
O4710-06	MJUBP0	303	0.0362 -		SFAM01.1	PPB	10/17/2023 16:48	SMPL
O4710-07	MJUBP1	1184	0.1916 -		SFAM01.1	PPB	10/17/2023 16:50	SMPL
O4710-08	MJUBP3	508	0.0724 -		SFAM01.1	PPB	10/17/2023 16:52	SMPL
O4710-09	MJUBP4	309	0.0373 -		SFAM01.1	PPB	10/17/2023 16:54	SMPL

O4710-10	MJUBP5	707	0.1075 -	SFAM01.1	PPB	10/17/2023 16:57 SMPL
O4710-11	MJUBP7	4007	0.6893 -	SFAM01.1	PPB	10/17/2023 16:59 SMPL
O4710-12	MJUBP9	946	0.1496 -	SFAM01.1	PPB	10/17/2023 17:01 SMPL
O4710-13	MJUBQ0	872	0.1366 -	SFAM01.1	PPB	10/17/2023 17:03 SMPL
O4710-14	MJUBQ1	713	0.1085 -	SFAM01.1	PPB	10/17/2023 17:06 SMPL
O4710-15	MJUBQ2	814	0.1263 -	SFAM01.1	PPB	10/17/2023 17:08 SMPL
O4710-16	MJUBQ2D	815	0.1265 -	SFAM01.1	PPB	10/17/2023 17:10 SMPL
CCV021	CCV021	30691	5.3942 -	SFAM01.1	PPB	10/17/2023 17:12 SMPL
CCB021	CCB021	-31	-0.0226 -	SFAM01.1	PPB	10/17/2023 17:15 SMPL
O4710-17	MJUBQ2S	15902	2.7866 -	SFAM01.1	PPB	10/17/2023 17:17 SMPL
O4710-18	MJUBQ4	617	0.0916 -	SFAM01.1	PPB	10/17/2023 17:19 SMPL
O4710-19	MJUBQ5	779	0.1202 -	SFAM01.1	PPB	10/17/2023 17:22 SMPL
O4710-20	MJUBQ6	3607	0.6188 -	SFAM01.1	PPB	10/17/2023 17:24 SMPL
O4710-21	MJUBQ7	834	0.1299 -	SFAM01.1	PPB	10/17/2023 17:26 SMPL
O4710-22	MJUBQ8	1848	0.3087 -	SFAM01.1	PPB	10/17/2023 17:28 SMPL
PB156458BL	PBS458	-8	-0.0186 -	SFAM01.1	PPB	10/17/2023 17:31 SMPL
O4711-01	MJUBM4	349	0.0444 -	SFAM01.1	PPB	10/17/2023 17:33 SMPL
O4711-02	MJUBM5	306	0.0368 -	SFAM01.1	PPB	10/17/2023 17:35 SMPL
O4711-03	MJUBM6	1085	0.1741 -	SFAM01.1	PPB	10/17/2023 17:37 SMPL
O4711-04	MJUBM7	553	0.0803 -	SFAM01.1	PPB	10/17/2023 17:40 SMPL
O4711-05	MJUBM8	845	0.1318 -	SFAM01.1	PPB	10/17/2023 17:42 SMPL
O4711-06	MJUBM9	510	0.0727 -	SFAM01.1	PPB	10/17/2023 17:44 SMPL
O4711-07	MJUBN0	1258	0.2046 -	SFAM01.1	PPB	10/17/2023 17:46 SMPL
O4711-08	MJUBN1	971	0.154 -	SFAM01.1	PPB	10/17/2023 17:49 SMPL
O4711-09	MJUBN2	1977	0.3314 -	SFAM01.1	PPB	10/17/2023 17:51 SMPL
O4711-10	MJUBN3	625	0.093 -	SFAM01.1	PPB	10/17/2023 17:53 SMPL
O4711-11	MJUBN4	470	0.0657 -	SFAM01.1	PPB	10/17/2023 17:56 SMPL
O4711-12	MJUBQ9	810	0.1256 -	SFAM01.1	PPB	10/17/2023 17:58 SMPL
O4711-13	MJUBR0	1308	0.2134 -	SFAM01.1	PPB	10/17/2023 18:00 SMPL
CCV022	CCV022	31002	5.449 -	SFAM01.1	PPB	10/17/2023 18:02 SMPL
CCB022	CCB022	-14	-0.0196 -	SFAM01.1	PPB	10/17/2023 18:05 SMPL
O4711-14	MJUBR1	1183	0.1914 -	SFAM01.1	PPB	10/17/2023 18:07 SMPL
O4711-15	MJUBR2	762	0.1172 -	SFAM01.1	PPB	10/17/2023 18:09 SMPL
O4711-16	MJUBR3	2364	0.3996 -	SFAM01.1	PPB	10/17/2023 18:11 SMPL
O4711-17	MJUBR4	1809	0.3018 -	SFAM01.1	PPB	10/17/2023 18:14 SMPL
O4711-18	MJUBR5	4245	0.7313 -	SFAM01.1	PPB	10/17/2023 18:16 SMPL
O4711-19	MJUBR6	642	0.096 -	SFAM01.1	PPB	10/17/2023 18:18 SMPL
O4711-20	MJUBR7	242	0.0255 -	SFAM01.1	PPB	10/17/2023 18:21 SMPL
O4711-21	MJUBR7D	225	0.0225 -	SFAM01.1	PPB	10/17/2023 18:23 SMPL
O4711-22	MJUBR7S	17766	3.1153 -	SFAM01.1	PPB	10/17/2023 18:25 SMPL
CCV023	CCV023	30985	5.446 -	SFAM01.1	PPB	10/17/2023 18:27 SMPL
CCB023	CCB023	-31	-0.0226 -	SFAM01.1	PPB	10/17/2023 18:30 SMPL