



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB127843

Slope : 98.6

BalanceID: WC SC-6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	10/17/2023	13:00
2	CAL2	1	Water	NA	NA	20.6	7.01	10/17/2023	13:01
3	CAL3	1	Water	NA	NA	20.6	10.02	10/17/2023	13:02
4	ICV	1	Water	NA	NA	20.5	7.00	10/17/2023	13:05
5	CCV1	1	Water	NA	NA	20.6	2.01	10/17/2023	13:10
6	O4887-02	1	Solid	20.02	20	20.3	6.59	10/17/2023	13:11
7	O4887-02DUP	1	Solid	20.03	20	20.4	6.60	10/17/2023	13:12
8	O4888-02	1	Solid	20.05	20	20.6	6.81	10/17/2023	13:15
9	O4908-03	1	Solid	20.02	20	20.4	8.16	10/17/2023	13:25
10	O4908-06	1	Solid	20.01	20	20.4	8.06	10/17/2023	13:30
11	O4908-09	1	Solid	20.02	20	20.3	7.92	10/17/2023	13:35
12	CCV2	1	Water	NA	NA	20.6	12.02	10/17/2023	13:40