



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB127856

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	10/18/2023	08:30
2	O4924-03		1	14	100.00	>212.0	>212.0	10/18/2023	09:00
3	O4936-02		1	12	100.00	>212.0	>212.0	10/18/2023	09:30
4	O4936-03		1	13	100.00	>212.0	>212.0	10/18/2023	10:00
5	O4936-04		1	14	100.00	>212.0	>212.0	10/18/2023	10:30
6	O4936-05		1	12	100.00	>212.0	>212.0	10/18/2023	11:00
7	O4936-05DUP		1	12	100.00	>212.0	>212.0	10/18/2023	11:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB127856

WorkList Name : FP-10-17

WorkList ID : 174649

Department : Wet-Chemistry

Date : 10-17-2023 13:59:15

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4924-03 D	50102	Water	Flash Point	Cool 4 deg C	PSEG03	E31	10/17/2023	1010B
O4936-02 C	1026	Water	Flash Point	Cool 4 deg C	PSEG03	E31	10/17/2023	1010B
O4936-03 D	1027	Water	Flash Point	Cool 4 deg C	PSEG03	E31	10/17/2023	1010B
O4936-04 B	1030	Water	Flash Point	Cool 4 deg C	PSEG03	E31	10/17/2023	1010B
O4936-05 B	1695	Water	Flash Point	Cool 4 deg C	PSEG03	E31	10/17/2023	1010B

Date/Time 10/18/2023 08:20
 Raw Sample Received by: RM CWJ
 Raw Sample Relinquished by: JH/CWJ

Date/Time 10/18/2023 17:10
 Raw Sample Received by: JH/CWJ
 Raw Sample Relinquished by: RM CWJ