

LB 127859


SFAM01.1

INSTRUMENT ID: CV2

Linear

 μ Abs.:

58111



A= 0.0000e+000

B= 1.7073e-004 slope

C= -4.5472e-002 intercept

Rho= 0.9997291

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.033	-0.033	71	0.000	71					
0.05	0.050					410	0				
0.20	0.200	0.186	-0.014	1355	0.0 %	1355					
2.50	2.500	2.526	0.026	15064	0.0 %	15064					
5.00	5.000	4.982	-0.018	29447	0.0 %	29447					
7.50	7.500	7.663	0.163	45149	0.0 %	45149					
10.0	10.000	9.876	-0.124	58111	0.0 %	58111					

71
410
1355
15064
29447
45149
58111

INSTRUMENTY ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	71	-		0 SFAM01.1	PPB	10/18/2023 13:50	Std
	0.2 S01	1355	-		0.2 SFAM01.1	PPB	10/18/2023 13:52	Std
	2.5 S02	15064	-		2.5 SFAM01.1	PPB	10/18/2023 13:54	Std
	5 S03	29447	-		5 SFAM01.1	PPB	10/18/2023 13:56	Std
	7.5 S04	45149	-		7.5 SFAM01.1	PPB	10/18/2023 13:59	Std
	10 S05	58111	-		10 SFAM01.1	PPB	10/18/2023 14:01	Std
ICV024	ICV024	23565	3.9778 -		SFAM01.1	PPB	10/18/2023 14:04	SMPL
ICB024	ICB024	-129	-0.0675 -		SFAM01.1	PPB	10/18/2023 14:06	SMPL
CCV024	CCV024	29098	4.9225 -		SFAM01.1	PPB	10/18/2023 14:08	SMPL
CCB024	CCB024	-194	-0.0786 -		SFAM01.1	PPB	10/18/2023 14:11	SMPL
PB156481BL	PBS481	270	0.0006 -		SFAM01.1	PPB	10/18/2023 14:13	SMPL
O4755-01	YC5M2	6090	0.9943 -		SFAM01.1	PPB	10/18/2023 14:15	SMPL
O4755-03	YC5M2S	21456	3.6178 -		SFAM01.1	PPB	10/18/2023 14:20	SMPL
O4755-04	YC5M3	5493	0.8924 -		SFAM01.1	PPB	10/18/2023 14:22	SMPL
O4755-05	YC5M4	11836	1.9753 -		SFAM01.1	PPB	10/18/2023 14:24	SMPL
O4755-06	YC5M5	9852	1.6366 -		SFAM01.1	PPB	10/18/2023 14:26	SMPL
O4755-07	YC5M6	12256	2.047 -		SFAM01.1	PPB	10/18/2023 14:29	SMPL
O4755-08	YC5M7	17078	2.8703 -		SFAM01.1	PPB	10/18/2023 14:31	SMPL
O4755-09	YC5M8	8336	1.3778 -		SFAM01.1	PPB	10/18/2023 14:33	SMPL
O4755-10	YC5M9	10320	1.7165 -		SFAM01.1	PPB	10/18/2023 14:35	SMPL
O4755-11	YC5N0	2280	0.3438 -		SFAM01.1	PPB	10/18/2023 14:38	SMPL
O4755-12	YC5N1	1809	0.2634 -		SFAM01.1	PPB	10/18/2023 14:40	SMPL
O4755-13	YC5N2	2605	0.3993 -		SFAM01.1	PPB	10/18/2023 14:42	SMPL
O4755-14	YC5N3	3960	0.6306 -		SFAM01.1	PPB	10/18/2023 14:45	SMPL
O4755-15	YC5N4	2619	0.4017 -		SFAM01.1	PPB	10/18/2023 14:47	SMPL
O4755-16	YC5N5	2538	0.3878 -		SFAM01.1	PPB	10/18/2023 14:49	SMPL
O4755-17	YC5N6	2142	0.3202 -		SFAM01.1	PPB	10/18/2023 14:51	SMPL
O4755-18	YC5N7	2327	0.3518 -		SFAM01.1	PPB	10/18/2023 14:54	SMPL
O4755-19	YC5N8	2611	0.4003 -		SFAM01.1	PPB	10/18/2023 14:56	SMPL
O4755-02	YC5M2D	6095	0.9951 -		SFAM01.1	PPB	10/18/2023 14:58	SMPL
CCV025	CCV025	29920	5.0629 -		SFAM01.1	PPB	10/18/2023 15:01	SMPL
CCB025	CCB025	-5	-0.0463 -		SFAM01.1	PPB	10/18/2023 15:03	SMPL
O4755-20	YC5N9	3372	0.5302 -		SFAM01.1	PPB	10/18/2023 15:06	SMPL
O4755-21	YC5P2	20713	3.4909 -		SFAM01.1	PPB	10/18/2023 15:08	SMPL
O4755-22	YC5P9	25356	4.2836 -		SFAM01.1	PPB	10/18/2023 15:10	SMPL
PB156482BL	PBS482	281	0.0025 -		SFAM01.1	PPB	10/18/2023 15:12	SMPL
O4756-01	YC5P0	23736	4.007 -		SFAM01.1	PPB	10/18/2023 15:15	SMPL
O4756-02	YC5P1	11890	1.9845 -		SFAM01.1	PPB	10/18/2023 15:17	SMPL
O4756-03	YC5P3	24615	4.1571 -		SFAM01.1	PPB	10/18/2023 15:19	SMPL
O4756-04	YC5P4	21886	3.6912 -		SFAM01.1	PPB	10/18/2023 15:21	SMPL
O4756-05	YC5P5	23626	3.9883 -		SFAM01.1	PPB	10/18/2023 15:24	SMPL
O4756-06	YC5P6	21451	3.6169 -		SFAM01.1	PPB	10/18/2023 15:26	SMPL
O4756-07	YC5P7	27714	4.6862 -		SFAM01.1	PPB	10/18/2023 15:28	SMPL
O4756-08	YC5P8	17690	2.9748 -		SFAM01.1	PPB	10/18/2023 15:30	SMPL
O4756-09	YC5Q0	12587	2.1035 -		SFAM01.1	PPB	10/18/2023 15:33	SMPL
O4756-10	YC5Y8	20508	3.4559 -		SFAM01.1	PPB	10/18/2023 15:35	SMPL

O4756-11	YC5Q1	21243	3.5814 -	SFAM01.1	PPB	10/18/2023 15:37 SMPL
O4756-12	YC5Q2	15128	2.5374 -	SFAM01.1	PPB	10/18/2023 15:39 SMPL
O4756-13	YC5Q3	13138	2.1976 -	SFAM01.1	PPB	10/18/2023 15:42 SMPL
O4756-14	YC5Q4	56025	9.5198 -	SFAM01.1	PPB	10/18/2023 15:44 SMPL
O4756-15	YC5Q5	34521	5.8484 -	SFAM01.1	PPB	10/18/2023 15:46 SMPL
O4756-16	YC5Q6	24352	4.1122 -	SFAM01.1	PPB	10/18/2023 15:49 SMPL
CCV026	CCV026	30485	5.1593 -	SFAM01.1	PPB	10/18/2023 15:51 SMPL
CCB026	CCB026	-30	-0.0506 -	SFAM01.1	PPB	10/18/2023 15:53 SMPL
O4756-17	YC5Q6D	24681	4.1684 -	SFAM01.1	PPB	10/18/2023 15:58 SMPL
O4756-18	YC5Q6S	40619	6.8895 -	SFAM01.1	PPB	10/18/2023 16:00 SMPL
O4756-19	YC5Q7	21055	3.5493 -	SFAM01.1	PPB	10/18/2023 16:02 SMPL
O4756-20	YC5R4	3253	0.5099 -	SFAM01.1	PPB	10/18/2023 16:04 SMPL
O4756-21	YC5R7	22997	3.8809 -	SFAM01.1	PPB	10/18/2023 16:07 SMPL
O4756-22	YC5R8	21178	3.5703 -	SFAM01.1	PPB	10/18/2023 16:09 SMPL
PB156483BL	PBS483	-24	-0.0496 -	SFAM01.1	PPB	10/18/2023 16:11 SMPL
O4757-01	YC5R9	12220	2.0409 -	SFAM01.1	PPB	10/18/2023 16:13 SMPL
O4757-02	YC5S0	23420	3.9531 -	SFAM01.1	PPB	10/18/2023 16:16 SMPL
O4757-03	YC5S0D	25956	4.3861 -	SFAM01.1	PPB	10/18/2023 16:18 SMPL
O4757-05	YC5S1	3547	0.5601 -	SFAM01.1	PPB	10/18/2023 16:22 SMPL
O4757-06	YC5S2	2014	0.2984 -	SFAM01.1	PPB	10/18/2023 16:25 SMPL
O4757-07	YC5S3	4280	0.6853 -	SFAM01.1	PPB	10/18/2023 16:27 SMPL
O4757-08	YC5S4	2279	0.3436 -	SFAM01.1	PPB	10/18/2023 16:29 SMPL
O4757-09	YC5S5	4688	0.7549 -	SFAM01.1	PPB	10/18/2023 16:32 SMPL
O4757-10	YC5S6	2417	0.3672 -	SFAM01.1	PPB	10/18/2023 16:34 SMPL
O4757-11	YC5S7	3986	0.6351 -	SFAM01.1	PPB	10/18/2023 16:36 SMPL
O4757-12	YC5Y7	2876	0.4456 -	SFAM01.1	PPB	10/18/2023 16:38 SMPL
O4757-13	YC5Q8	8071	1.3325 -	SFAM01.1	PPB	10/18/2023 16:41 SMPL
CCV027	CCV027	29764	5.0362 -	SFAM01.1	PPB	10/18/2023 16:43 SMPL
CCB027	CCB027	-25	-0.0497 -	SFAM01.1	PPB	10/18/2023 16:45 SMPL
O4757-14	YC5Q9	19737	3.3243 -	SFAM01.1	PPB	10/18/2023 16:47 SMPL
O4757-15	YC5R0	15797	2.6516 -	SFAM01.1	PPB	10/18/2023 16:50 SMPL
O4757-16	YC5R1	15726	2.6395 -	SFAM01.1	PPB	10/18/2023 16:52 SMPL
O4757-17	YC5R2	20520	3.458 -	SFAM01.1	PPB	10/18/2023 16:54 SMPL
O4757-18	YC5R3	26320	4.4482 -	SFAM01.1	PPB	10/18/2023 16:57 SMPL
O4757-19	YC5R5	2935	0.4556 -	SFAM01.1	PPB	10/18/2023 16:59 SMPL
O4757-20	YC5R6	7496	1.2343 -	SFAM01.1	PPB	10/18/2023 17:01 SMPL
O4757-21	YC5S8	5716	0.9304 -	SFAM01.1	PPB	10/18/2023 17:03 SMPL
O4757-22	YC5S9	5853	0.9538 -	SFAM01.1	PPB	10/18/2023 17:06 SMPL
PB156484BL	PBS484	165	-0.0173 -	SFAM01.1	PPB	10/18/2023 17:08 SMPL
O4758-01	YC5T9	3900	0.6204 -	SFAM01.1	PPB	10/18/2023 17:10 SMPL
O4758-02	YC5W0	3208	0.5022 -	SFAM01.1	PPB	10/18/2023 17:12 SMPL
O4758-03	YC5W1	2442	0.3715 -	SFAM01.1	PPB	10/18/2023 17:15 SMPL
O4758-04	YC5W2	2564	0.3923 -	SFAM01.1	PPB	10/18/2023 17:17 SMPL
O4758-05	YC5W3	2271	0.3423 -	SFAM01.1	PPB	10/18/2023 17:19 SMPL
O4758-06	YC5W4	2096	0.3124 -	SFAM01.1	PPB	10/18/2023 17:22 SMPL
O4758-07	YC5W5	2296	0.3465 -	SFAM01.1	PPB	10/18/2023 17:24 SMPL
O4758-08	YC5W6	1894	0.2779 -	SFAM01.1	PPB	10/18/2023 17:26 SMPL

LB127859

INSTRUMENTY ID : CV1

O4758-09	YC5Y9	10415	1.7327 -	SFAM01.1	PPB	10/18/2023 17:28 SMPL
O4758-10	YC5Z0	7087	1.1645 -	SFAM01.1	PPB	10/18/2023 17:31 SMPL
CCV028	CCV028	29583	5.0053 -	SFAM01.1	PPB	10/18/2023 17:33 SMPL
CCB028	CCB028	-27	-0.0501 -	SFAM01.1	PPB	10/18/2023 17:35 SMPL
O4758-11	YC5W7	4166	0.6658 -	SFAM01.1	PPB	10/18/2023 17:37 SMPL
O4758-12	YC5W8	4586	0.7375 -	SFAM01.1	PPB	10/18/2023 17:40 SMPL
O4758-13	YC5W8D	4270	0.6836 -	SFAM01.1	PPB	10/18/2023 17:42 SMPL
O4758-14	YC5W8S	20465	3.4486 -	SFAM01.1	PPB	10/18/2023 17:48 SMPL
O4758-15	YC5W9	2641	0.4054 -	SFAM01.1	PPB	10/18/2023 17:50 SMPL
O4758-16	YC5X0	2018	0.2991 -	SFAM01.1	PPB	10/18/2023 17:52 SMPL
O4758-17	YC5X1	2770	0.4275 -	SFAM01.1	PPB	10/18/2023 17:55 SMPL
O4758-18	YC5X2	3122	0.4876 -	SFAM01.1	PPB	10/18/2023 17:57 SMPL
O4758-19	YC5X3	2325	0.3515 -	SFAM01.1	PPB	10/18/2023 17:59 SMPL
O4758-20	YC5X4	1304	0.1772 -	SFAM01.1	PPB	10/18/2023 18:01 SMPL
O4758-21	YC5X5	1416	0.1963 -	SFAM01.1	PPB	10/18/2023 18:04 SMPL
O4758-22	YC5X6	2729	0.4205 -	SFAM01.1	PPB	10/18/2023 18:06 SMPL
O4757-04	YC5S0S	35856	6.0763 -	SFAM01.1	PPB	10/18/2023 18:08 SMPL
CCV029	CCV029	30204	5.1113 -	SFAM01.1	PPB	10/18/2023 18:10 SMPL
CCB029	CCB029	-39	-0.0521 -	SFAM01.1	PPB	10/18/2023 18:13 SMPL