

Analysis Method: SM2510 BANALYST: IwonaParameter: ConductanceSUPERVISOR REVIEW BY: SohilRun Number: LB127860Meter ID: WC CONDUCTANCE METER-1Thermometer ID: WC PH METER-1

Reagent/Standard	Lot/Log #
conductivity standard	WP104779

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB127860BL		1	19.1	0.81	0.91	10/18/2023	13:40
2	LB127860BS	1413	1	19.4	1274.00	1426.59	10/18/2023	13:43
3	O4916-01		1	19.3	2.31	2.59	10/18/2023	13:46
4	O4916-01DUP		1	19.3	2.30	2.58	10/18/2023	13:49

Note : Conductivity result is corrected to 25°C

Result = $\frac{\text{(Reading)}}{1 + 0.0191 (\text{Temperature} - 25)}$
--

WORKLIST(Hardcopy Internal Chain)

LB 127060

WorkList Name : CONDUCTANCE-101823 **WorkList ID :** 174685 **Department :** Wet-Chemistry **Date :** 10-18-2023 13:28:49

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4916-01	JS960	Water	Conductance	Cool 4 deg C	STER17	I11	10/30/2023	SM2510 B

Date/Time 10/19/23 13:30
Raw Sample Received by: 12/01
Raw Sample Relinquished by: 12/01

Date/Time 10/18/23 13:55
Raw Sample Received by: 12/01
Raw Sample Relinquished by: 12/01