



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB127864

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	10/18/2023	14:20
2	O4936-09		1	2	5.00	41.0	41.0	10/18/2023	14:50
3	O4936-09DUP		1	2	5.00	41.0	41.0	10/18/2023	15:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66127864

WorkList Name : fp s-10-18

WorkList ID : 174668

Department : Wet-Chemistry

Date : 10-18-2023 08:29:09

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4936-09	1704	Solid	Flash Point	Cool 4 deg C	PSEG03	E31	10/17/2023	1010B

Date/Time 10/18/2023 14:15
 Raw Sample Received by: RM CWC
 Raw Sample Relinquished by: JAC CWC

Date/Time 10/18/2023 15:20
 Raw Sample Received by: JAC CWC
 Raw Sample Relinquished by: JAC CWC