

CB127868


7470A

INSTRUMENT ID: CV1

Linear

μ Abs.:

46567



A= 0.0000e+000

B= 2.1422e-004

C= -3.3933e-002

Rho= 0.9999164

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.004	-0.004	142	0.000	142	0				7.00
0.2	0.200	0.212	0.012	1148	0.0 %	1148					6
2.5	2.500	2.467	-0.033	11676	0.0 %	11676					1
5.0	5.000	4.988	-0.012	23443	0.0 %	23443					0
7.5	7.500	7.594	0.094	35609	0.0 %	35609					1
10.0	10.000	9.942	-0.058	46567	0.0 %	46567					1

LB127868 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 19 Oct 2023 09:56:19

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0.0 - 1	50	142	-	0.0000	7470A	PPB	19 Oct 2023 10:04:23	S
0.2 - 1	502	1148	-	0.2000	7470A	PPB	19 Oct 2023 10:12:09	S
2.5 - 1	52.5	11676	-	2.5000	7470A	PPB	19 Oct 2023 10:14:26	S
5.0 - 1	55	23443	-	5.0000	7470A	PPB	19 Oct 2023 10:16:43	S
7.5 - 1	57.5	35609	-	7.5000	7470A	PPB	19 Oct 2023 10:19:00	S
10.0 - 1	511	46567	-	10.0000	7470A	PPB	19 Oct 2023 10:21:17	S
ICV25 - 1	ICV25	17497	3.7143	-	7470A	PPB	19 Oct 2023 10:24:00	U S
ICB25 - 1	ICB25	-43	-0.0431	-	7470A	PPB	19 Oct 2023 10:29:13	U S
CCV30 - 1	CCV30	21816	4.6395	-	7470A	PPB	19 Oct 2023 10:31:28	U S
CCB30 - 1	CCB30	-297	-0.0976	-	7470A	PPB	19 Oct 2023 10:33:43	U S
CRA - 1	CRA	1007	0.1818	-	7470A	PPB	19 Oct 2023 10:36:00	U S
HighStd - 1	HighStd	46328	9.8906	-	7470A	PPB	19 Oct 2023 10:38:15	U S
ChkStd - 1	ChkStd	33313	7.1025	-	7470A	PPB	19 Oct 2023 10:40:30	U S
PB156473BL - 1	PBW	95	-0.0136	-	7470A	PPB	19 Oct 2023 10:44:55	U S
PB156473BS - 1	LCSW	16723	3.5485	-	7470A	PPB	19 Oct 2023 10:47:10	U S
O4930-01 - 1	SCFGRAB-1	-218	-0.0806	-	7470A	PPB	19 Oct 2023 10:49:26	U S
O4930-02 - 1	SCFGRAB-2	-26	-0.0395	-	7470A	PPB	19 Oct 2023 10:51:44	U S
O4930-03 - 1	SCFGRAB-3	0	-0.0339	-	7470A	PPB	19 Oct 2023 10:54:01	U S
O4930-04 - 1	SCFGRAB-4	34	-0.0266	-	7470A	PPB	19 Oct 2023 10:56:17	U S
O4930-05 - 1	SCFGRAB-5	32	-0.0271	-	7470A	PPB	19 Oct 2023 10:58:32	U S
CCV31 - 1	CCV31	22405	4.7657	-	7470A	PPB	19 Oct 2023 11:00:47	U S
CCB31 - 1	CCB31	-117	-0.0590	-	7470A	PPB	19 Oct 2023 11:03:03	U S
O4930-06 - 1	SCFGRAB-6	-33	-0.0410	-	7470A	PPB	19 Oct 2023 11:05:19	U S
O4930-07 - 1	SCFGRAB-7	308	0.0320	-	7470A	PPB	19 Oct 2023 11:07:34	U S
O4930-07DUP - 1	SCFGRAB-7DUP	18	-0.0301	-	7470A	PPB	19 Oct 2023 11:09:50	U S
O4930-07MS - 1	SCFGRAB-7MS	16724	3.5487	-	7470A	PPB	19 Oct 2023 11:12:06	U S
O4930-07MSD - 1	SCFGRAB-7MSD	17098	3.6288	-	7470A	PPB	19 Oct 2023 11:14:21	U S
PB156474BL - 1	PBW	22	-0.0292	-	7470A	PPB	19 Oct 2023 11:16:39	U S
PB156474BS - 1	LCSW	15613	3.3107	-	7470A	PPB	19 Oct 2023 11:18:58	U S
O4908-03 - 1	TP-1	-187	-0.0740	-	7470A	PPB	19 Oct 2023 11:21:15	U S
O4908-03DUP - 1	TP-1DUP	-11	-0.0363	-	7470A	PPB	19 Oct 2023 11:23:32	U S
O4908-03MS - 1	TP-1MS	15919	3.3763	-	7470A	PPB	19 Oct 2023 11:25:47	U S
CCV32 - 1	CCV32	23098	4.9142	-	7470A	PPB	19 Oct 2023 11:28:03	U S
CCB32 - 1	CCB32	-107	-0.0569	-	7470A	PPB	19 Oct 2023 11:30:18	U S
O4908-03MSD - 1	TP-1MSD	16334	3.4652	-	7470A	PPB	19 Oct 2023 11:36:59	U S
O4908-06 - 1	TP-2	-285	-0.0950	-	7470A	PPB	19 Oct 2023 11:39:14	U S
O4908-09 - 1	TP-3	-78	-0.0506	-	7470A	PPB	19 Oct 2023 11:41:31	U S
PB156473TB - 1		-43	-0.0431	-	7470A	PPB	19 Oct 2023 11:43:46	U S
PB156474TB - 1		-13	-0.0367	-	7470A	PPB	19 Oct 2023 11:46:02	U S
O4908-03LX5 - 1		47	-0.0239	-	7470A	PPB	19 Oct 2023 11:48:19	U S
O4908-03A - 1		17659	3.7490	-	7470A	PPB	19 Oct 2023 11:50:35	U S
O4930-07LX5 - 1		-115	-0.0586	-	7470A	PPB	19 Oct 2023 11:52:51	U S
O4930-07A - 1		19450	4.1327	-	7470A	PPB	19 Oct 2023 11:55:08	U S
CCV33 - 1	CCV33	21912	4.6601	-	7470A	PPB	19 Oct 2023 11:57:24	U S
CCB33 - 1	CCB33	-198	-0.0763	-	7470A	PPB	19 Oct 2023 11:59:41	U S