

L3127885


SFAM01.1

INSTRUMENT ID: EVJ

Linear

μ Abs.:

56467



A= 0.0000e+000

B= 1.7559e-004 slope

C= 6.0240e-003 y-intercept

Rho= 0.9995570

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.014	0.014	48	0.000	48					2
0.05	0.050					410	0				17
0.20	0.200	0.234	0.034	1299	0.0 %	1299					-1
2.50	2.500	2.480	-0.020	14087	0.0 %	14087					-3
5.00	5.000	4.848	-0.152	27575	0.0 %	27575					3
7.50	7.500	7.703	0.203	43835	0.0 %	43835					-1
10.0	10.000	9.921	-0.079	56467	0.0 %	56467					

LB127885

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	48	-		0 SFAM01.1	PPB	10/19/2023 15:22	Std
	0.2 S01	1299	-		0.2 SFAM01.1	PPB	10/19/2023 15:25	Std
	2.5 S02	14087	-		2.5 SFAM01.1	PPB	10/19/2023 15:27	Std
	5 S03	27575	-		5 SFAM01.1	PPB	10/19/2023 15:31	Std
	7.5 S04	43835	-		7.5 SFAM01.1	PPB	10/19/2023 15:33	Std
	10 S05	56467	-		10 SFAM01.1	PPB	10/19/2023 15:35	Std
ICV027	ICV027	22541	3.964 -		SFAM01.1	PPB	10/19/2023 15:43	SMPL
ICB027	ICB027	-184	-0.0263 -		SFAM01.1	PPB	10/19/2023 15:45	SMPL
CCV038	CCV038	28004	4.9232 -		SFAM01.1	PPB	10/19/2023 15:48	SMPL
CCB038	CCB038	-220	-0.0326 -		SFAM01.1	PPB	10/19/2023 15:50	SMPL
PB156510BL	PBW510	75	0.0192 -		SFAM01.1	PPB	10/19/2023 15:52	SMPL
O4746-01	MBH578	0	0.006 -		SFAM01.1	PPB	10/19/2023 15:54	SMPL
O4746-02	MBH579	89	0.0217 -		SFAM01.1	PPB	10/19/2023 15:57	SMPL
O4746-03	MBH580	48	0.0145 -		SFAM01.1	PPB	10/19/2023 15:59	SMPL
O4746-04	MBH582	42	0.0134 -		SFAM01.1	PPB	10/19/2023 16:01	SMPL
O4746-05	MBH583	22	0.0099 -		SFAM01.1	PPB	10/19/2023 16:03	SMPL
O4746-06	MBH583D	30	0.0113 -		SFAM01.1	PPB	10/19/2023 16:06	SMPL
O4746-07	MBH583S	6499	1.1472 -		SFAM01.1	PPB	10/19/2023 16:16	SMPL
O4759-09	YC5Y6	-22	0.0022 -		SFAM01.1	PPB	10/19/2023 16:19	SMPL
CCV039	CCV039	27533	4.8405 -		SFAM01.1	PPB	10/19/2023 16:21	SMPL
CCB039	CCB039	-187	-0.0268 -		SFAM01.1	PPB	10/19/2023 16:23	SMPL