



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB127888

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	10/19/2023	16:10
2	CAL2	1	Water	NA	NA	20.6	7.00	10/19/2023	16:11
3	CAL3	1	Water	NA	NA	20.6	10.02	10/19/2023	16:15
4	ICV	1	Water	NA	NA	20.6	7.01	10/19/2023	16:20
5	CCV1	1	Water	NA	NA	20.6	2.01	10/19/2023	16:25
6	O4957-01	1	Water	NA	NA	20.4	7.62	10/19/2023	16:33
7	O4957-01DUP	1	Water	NA	NA	20.5	7.63	10/19/2023	16:34
8	O4957-02	1	Water	NA	NA	20.6	7.30	10/19/2023	16:37
9	O4957-03	1	Water	NA	NA	20.6	7.31	10/19/2023	16:44
10	O4957-04	1	Water	NA	NA	20.6	7.64	10/19/2023	16:50
11	O4975-07	1	Water	NA	NA	20.3	9.80	10/19/2023	17:00
12	O4979-04	1	Water	NA	NA	20.3	3.14	10/19/2023	17:10
13	CCV2	1	Water	NA	NA	20.6	12.02	10/19/2023	17:13

VB121886

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH O4957

WorkList ID : 174740

Department : Wet-Chemistry

Date : 10-19-2023 15:45:31

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4957-01	MH-E	Water	pH	Cool 4 deg C	PSEG03	D31	10/18/2023	9040C
O4957-02	MH-F	Water	pH	Cool 4 deg C	PSEG03	D31	10/18/2023	9040C
O4957-03	MH-G	Water	pH	Cool 4 deg C	PSEG03	D31	10/18/2023	9040C
O4957-04	MH-H	Water	pH	Cool 4 deg C	PSEG03	D31	10/18/2023	9040C
O4975-07	1060	Water	pH	Cool 4 deg C	PSEG03	E21	10/19/2023	9040C
O4979-04	GAS-CONDENSATE	Water	pH	Cool 4 deg C	PSEG03	I11	10/19/2023	9040C

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

15:50

10/19/23

70 (WC)

IT (SM)

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

17:30

10/19/23

77 (SM)

70 (WC)