



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/20/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 17:15
In Date: 10/19/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:45
Out Date: 10/20/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB127891

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O4968-01	YC612	1	1.18	8.44	9.62	9.41	97.5	
O4968-02	YC613	2	1.19	8.42	9.61	9.54	99.2	
O4968-03	YC614	3	1.14	8.60	9.74	9.55	97.8	
O4968-04	YC614D	4	1.14	8.60	9.74	9.55	97.8	
O4968-05	YC614S	5	1.14	8.60	9.74	9.55	97.8	
O4968-06	YC615	6	1.16	8.42	9.58	9.49	98.9	
O4968-07	YC616	7	1.15	8.41	9.56	9.05	93.9	
O4968-08	YC617	8	1.19	8.63	9.82	9.42	95.4	
O4968-09	YC618	9	1.18	8.48	9.66	8.84	90.3	
O4968-10	YC619	10	1.18	8.54	9.72	8.99	91.5	
O4968-11	YC6E0	11	1.17	8.58	9.75	9.18	93.4	
O4968-12	YC6K6	12	1.17	8.56	9.73	9.00	91.5	
O4968-13	YC686	13	1.19	8.62	9.81	9.21	93.0	
O4968-14	YC687	14	1.16	8.61	9.77	8.95	90.5	
O4968-15	YC688	15	1.19	8.41	9.6	8.84	91.0	
O4968-16	YC689	16	1.15	8.67	9.82	9.1	91.7	
O4968-17	YC690	17	1.14	8.67	9.81	8.8	88.4	
O4968-18	YC6K4	18	1.15	8.83	9.98	9.28	92.1	
O4968-19	YC6K5	19	1.19	8.40	9.59	8.96	92.5	
O4968-20	YC6K7	20	1.16	8.62	9.78	8.99	90.8	
O4968-21	YC6K8	21	1.14	8.40	9.54	8.79	91.1	
O4968-22	YC6K9	22	1.19	8.63	9.82	9.06	91.2	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$