



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB127893

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	10/19/2023	12:45
2	O4957-01		1	12	100.00	>212.0	>212.0	10/19/2023	13:15
3	O4957-01DUP		1	12	100.00	>212.0	>212.0	10/19/2023	13:45
4	O4957-02		1	14	100.00	>212.0	>212.0	10/19/2023	14:15
5	O4957-03		1	12	100.00	>212.0	>212.0	10/19/2023	14:45
6	O4957-04		1	13	100.00	>212.0	>212.0	10/19/2023	15:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6127893

WorkList Name : FP-10-19

WorkList ID : 174745

Department : Wet-Chemistry

Date : 10-19-2023 10:40:23

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4957-01	MH-E	Water	Flash Point	Cool 4 deg C	PSEG03	D31	10/18/2023	1010B
O4957-02	MH-F	Water	Flash Point	Cool 4 deg C	PSEG03	D31	10/18/2023	1010B
O4957-03	MH-G	Water	Flash Point	Cool 4 deg C	PSEG03	D31	10/18/2023	1010B
O4957-04	MH-H	Water	Flash Point	Cool 4 deg C	PSEG03	D31	10/18/2023	1010B

Date/Time 10/19/2023 12:40
 Raw Sample Received by: RM gwc
 Raw Sample Relinquished by: gwc

Date/Time 10/19/2023 15:30
 Raw Sample Received by: gwc
 Raw Sample Relinquished by: RM cws