



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB127908

Ambient Barometric Pressure (mmHg): 750.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	83.0	10/20/2023	09:00
2	O4975-05		1	14	100.00	>212.0	>212.0	10/20/2023	09:30
3	O4975-06		1	12	100.00	>212.0	>212.0	10/20/2023	10:00
4	O4979-04		1	12	100.00	>212.0	>212.0	10/20/2023	10:30
5	O4979-04DUP		1	12	100.00	>212.0	>212.0	10/20/2023	11:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp-10-20

WorkList ID : 174752

Department : Wet-Chemistry

Date : 10-20-2023 08:07:27

65127908

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4975-05 5	52	Water	Flash Point	Cool 4 deg C	PSEG03	E21	10/19/2023	1010B
O4975-06 B	524	Water	Flash Point	Cool 4 deg C	PSEG03	E21	10/19/2023	1010B
O4979-04 E	GAS-CONDENSATE	Water	Flash Point	Cool 4 deg C	PSEG03	I11	10/19/2023	1010B

Date/Time 10/20/2023 08:50
 Raw Sample Received by: RMI GWD
 Raw Sample Relinquished by: JAGGOC

Date/Time 10/20/2023 11:35
 Raw Sample Received by: JAGGOC
 Raw Sample Relinquished by: JAGGOC