


LB127923

SFAM01.1

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 1.8836e-004 slope

C= -3.8861e-002 y-intercept

Rho= 0.9999062

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% of 1
0.00	0.000	0.017	0.017	297	0.000	297					1
0.05	0.050	0.040	-0.010	419	0.0 %	419	0				-20
0.20	0.200	0.199	-0.001	1261	0.0 %	1261					1
2.50	2.500	2.446	-0.054	13192	0.0 %	13192					-2
5.00	5.000	5.026	0.026	26887	0.0 %	26887					1
7.50	7.500	7.595	0.095	40526	0.0 %	40526					1
10.0	10.000	9.928	-0.072	52914	0.0 %	52914					1

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	297	-		0 SFAM01.1	PPB	10/23/2023 8:15	Std
	0.05 S01	419	-		0.05 SFAM01.1	PPB	10/23/2023 8:17	Std
	0.2 S02	1261	-		0.2 SFAM01.1	PPB	10/23/2023 8:19	Std
	2.5 S03	13192	-		2.5 SFAM01.1	PPB	10/23/2023 8:21	Std
	5 S04	26887	-		5 SFAM01.1	PPB	10/23/2023 8:24	Std
	7.5 S05	40526	-		7.5 SFAM01.1	PPB	10/23/2023 8:26	Std
	10 S06	52914	-		10 SFAM01.1	PPB	10/23/2023 8:28	Std
ICV030	ICV030	21652	4.0395 -		SFAM01.1	PPB	10/23/2023 8:31	SMPL
ICB030	ICB030	-41	-0.0466 -		SFAM01.1	PPB	10/23/2023 8:33	SMPL
CCV046	CCV046	27384	5.1192 -		SFAM01.1	PPB	10/23/2023 8:36	SMPL
CCB046	CCB046	-64	-0.0509 -		SFAM01.1	PPB	10/23/2023 8:38	SMPL
PB156547BL	PBS547	-1	-0.039 -		SFAM01.1	PPB	10/23/2023 8:40	SMPL
O4777-01	MJUBR8	319	0.0212 -		SFAM01.1	PPB	10/23/2023 8:42	SMPL
O4777-02	MJUBR9	1197	0.1866 -		SFAM01.1	PPB	10/23/2023 8:45	SMPL
O4777-03	MJUBS0	1778	0.296 -		SFAM01.1	PPB	10/23/2023 8:47	SMPL
O4777-04	MJUBS1	4355	0.7814 -		SFAM01.1	PPB	10/23/2023 8:49	SMPL
O4777-05	MJUBS2	1191	0.1855 -		SFAM01.1	PPB	10/23/2023 8:52	SMPL
O4777-06	MJUBS3	1206	0.1883 -		SFAM01.1	PPB	10/23/2023 8:54	SMPL
O4777-07	MJUBS4	968	0.1435 -		SFAM01.1	PPB	10/23/2023 8:56	SMPL
O4777-08	MJUBS5	1083	0.1651 -		SFAM01.1	PPB	10/23/2023 8:58	SMPL
O4777-09	MJUBS6	1896	0.3183 -		SFAM01.1	PPB	10/23/2023 9:01	SMPL
O4777-10	MJUBS7	676	0.0885 -		SFAM01.1	PPB	10/23/2023 9:03	SMPL
O4777-11	MJUBS8	2226	0.3804 -		SFAM01.1	PPB	10/23/2023 9:05	SMPL
O4777-12	MJUBS9	960	0.142 -		SFAM01.1	PPB	10/23/2023 9:07	SMPL
O4777-13	MJUBT0	590	0.0723 -		SFAM01.1	PPB	10/23/2023 9:10	SMPL
O4777-14	MJUBT1	630	0.0798 -		SFAM01.1	PPB	10/23/2023 9:12	SMPL
O4777-15	MJUBT2	1016	0.1525 -		SFAM01.1	PPB	10/23/2023 9:14	SMPL
O4777-16	MJUBT3	363	0.0295 -		SFAM01.1	PPB	10/23/2023 9:17	SMPL
O4777-17	MJUBT4	311	0.0197 -		SFAM01.1	PPB	10/23/2023 9:19	SMPL
O4777-18	MJUBT5	1165	0.1806 -		SFAM01.1	PPB	10/23/2023 9:21	SMPL
O4777-19	MJUBT5D	1236	0.194 -		SFAM01.1	PPB	10/23/2023 9:23	SMPL
CCV047	CCV047	28098	5.2537 -		SFAM01.1	PPB	10/23/2023 9:26	SMPL
CCB047	CCB047	149	-0.0108 -		SFAM01.1	PPB	10/23/2023 9:28	SMPL
O4777-20	MJUBT5S	13516	2.507 -		SFAM01.1	PPB	10/23/2023 9:35	SMPL
O4777-21	MJUBT6	1381	0.2213 -		SFAM01.1	PPB	10/23/2023 9:37	SMPL
O4777-22	MJUBT7	646	0.0828 -		SFAM01.1	PPB	10/23/2023 9:39	SMPL
PB156549BL	PBS549	17	-0.0357 -		SFAM01.1	PPB	10/23/2023 9:41	SMPL
O4797-01	MJUBT8	786	0.1092 -		SFAM01.1	PPB	10/23/2023 9:44	SMPL
O4797-02	MJUBT9	14477	2.688 -		SFAM01.1	PPB	10/23/2023 9:46	SMPL
O4797-03	MJUBW0	17792	3.3124 -		SFAM01.1	PPB	10/23/2023 9:48	SMPL
O4797-04	MJUBW1	3840	0.6844 -		SFAM01.1	PPB	10/23/2023 9:55	SMPL
O4797-05	MJUBW1D	3464	0.6136 -		SFAM01.1	PPB	10/23/2023 9:58	SMPL
O4797-06	MJUBW1S	16901	3.1446 -		SFAM01.1	PPB	10/23/2023 10:00	SMPL
CCV048	CCV048	28198	5.2725 -		SFAM01.1	PPB	10/23/2023 10:02	SMPL
CCB048	CCB048	106	-0.0189 -		SFAM01.1	PPB	10/23/2023 10:04	SMPL