



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB127926

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	10/23/2023	08:15
2	CAL2	1	Water	NA	NA	20.2	7.00	10/23/2023	08:16
3	CAL3	1	Water	NA	NA	20.2	10.02	10/23/2023	08:18
4	ICV	1	Water	NA	NA	20.6	7.00	10/23/2023	08:20
5	CCV1	1	Water	NA	NA	20.6	2.01	10/23/2023	08:25
6	O4991-01	1	Water	NA	NA	20.3	4.45	10/23/2023	08:30
7	O4991-01DUP	1	Water	NA	NA	20.4	4.46	10/23/2023	08:31
8	CCV2	1	Water	NA	NA	20.6	12.02	10/23/2023	08:35

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph o4991 WorkList ID : 174811 Department : Wet-Chemistry Date : 10-23-2023 08:01:29

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O4991-01	1021-1024	Water	pH	Cool 4 deg C	PSEG03	E31	10/20/2023	9040C

VB 12-7-926
VB 12-7-926

Date/Time 10/23/23 08:10
Raw Sample Received by: JO (QC)
Raw Sample Relinquished by: CP SM

Date/Time 10/23/23 11:00
Raw Sample Received by: CP SM
Raw Sample Relinquished by: JO (QC)