

LB127931


7471B

INSTRUMENT ID: CVJ

Linear

μ Abs.:

64199



A= 0.0000e+000

B= 1.5516e-004

C= -3.5395e-002

Rho= 0.9998335

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/sd
0.0	0.000	-0.034	-0.034	8	0.000	8					(
0.2	0.200	0.208	0.008	1568	0.0 %	1568					4
2.5	2.500	2.539	0.039	16592	0.0 %	16592					2
5.0	5.000	4.938	-0.062	32053	0.0 %	32053					-1
7.5	7.500	7.623	0.123	49357	0.0 %	49357					2
10.0	10.000	9.926	-0.074	64199	0.0 %	64199					-1

LB127931 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 23 Oct 2023 10:26:10

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SB	8	-	0.0000	7471B	PPB	23 Oct 2023 10:28:33	S	Std
0.2 - 1	SB-2	1568	-	0.2000	7471B	PPB	23 Oct 2023 10:30:50	S	Std
2.5 - 1	SB-5	16592	-	2.5000	7471B	PPB	23 Oct 2023 10:33:09	S	Std
5.0 - 1	SB	32053	-	5.0000	7471B	PPB	23 Oct 2023 10:35:27	S	Std
7.5 - 1	SB-5	49357	-	7.5000	7471B	PPB	23 Oct 2023 10:37:43	S	Std
10.0 - 1	SB-10	64199	-	10.0000	7471B	PPB	23 Oct 2023 10:42:37	S	Std
ICV31 - 1	ICV31	26433	4.0661	-	7471B	PPB	23 Oct 2023 10:45:22	U	SMPL
ICB31 - 1	ICB31	-403	-0.0979	-	7471B	PPB	23 Oct 2023 10:47:37	U	SMPL
CCV49 - 1	CCV49	32578	5.0196	-	7471B	PPB	23 Oct 2023 10:49:53	U	SMPL
CCB49 - 1	CCB49	-385	-0.0951	-	7471B	PPB	23 Oct 2023 10:52:08	U	SMPL
CRA - 1	CRA	1398	0.1815	-	7471B	PPB	23 Oct 2023 10:54:25	U	SMPL
HighStd - 1	HighStd	63895	9.8788	-	7471B	PPB	23 Oct 2023 10:56:40	U	SMPL
ChkStd - 1	ChkStd	46570	7.1906	-	7471B	PPB	23 Oct 2023 10:58:56	U	SMPL
PB156553BS - 1	LCSS	24068	3.6991	-	7471B	PPB	23 Oct 2023 11:03:32	U	SMPL
O4843-01 - 1	WC-1	20142	3.0899	-	7471B	PPB	23 Oct 2023 11:05:48	U	SMPL
O4879-01 - 1	TP-1	1456	0.1905	-	7471B	PPB	23 Oct 2023 11:08:06	U	SMPL
O4879-01DUP - 1	TP-1DUP	1265	0.1609	-	7471B	PPB	23 Oct 2023 11:10:23	U	SMPL
O4879-01MS - 1	TP-1MS	28747	4.4251	-	7471B	PPB	23 Oct 2023 11:12:41	U	SMPL
O4879-01MSD - 1	TP-1MSD	25171	3.8702	-	7471B	PPB	23 Oct 2023 11:14:56	U	SMPL
CCV50 - 1	CCV50	32388	4.9901	-	7471B	PPB	23 Oct 2023 11:17:11	U	SMPL
CCB50 - 1	CCB50	-156	-0.0596	-	7471B	PPB	23 Oct 2023 11:19:27	U	SMPL
O4879-03 - 1	TP-2	1429	0.1863	-	7471B	PPB	23 Oct 2023 11:21:43	U	SMPL
O4879-05 - 1	TP-3	968	0.1148	-	7471B	PPB	23 Oct 2023 11:23:57	U	SMPL
O4919-01 - 1	LSS-B2-COMP-101723	3438	0.4981	-	7471B	PPB	23 Oct 2023 11:26:13	U	SMPL
O4919-03 - 1	LSS-B5-COMP-101723	831	0.0935	-	7471B	PPB	23 Oct 2023 11:28:28	U	SMPL
O4941-01 - 1	213	2322	0.3249	-	7471B	PPB	23 Oct 2023 11:30:45	U	SMPL
O4941-03 - 1	72-12017	2933	0.4197	-	7471B	PPB	23 Oct 2023 11:33:01	U	SMPL
O4945-01 - 1	PAT-2-101823	4198	0.6160	-	7471B	PPB	23 Oct 2023 11:35:17	U	SMPL
O4960-01 - 1	VAUX-SLUDGE	2861	0.4085	-	7471B	PPB	23 Oct 2023 11:37:34	U	SMPL
O4974-01 - 1	TP-4	3758	0.5477	-	7471B	PPB	23 Oct 2023 11:39:50	U	SMPL
O4981-01 - 1	TR-04-10192023	3693	0.5376	-	7471B	PPB	23 Oct 2023 11:42:06	U	SMPL
CCV51 - 1	CCV51	33186	5.1139	-	7471B	PPB	23 Oct 2023 11:44:22	U	SMPL
CCB51 - 1	CCB51	-88	-0.0490	-	7471B	PPB	23 Oct 2023 11:46:38	U	SMPL
O4879-01LX5 - 1		174	-0.0084	-	7471B	PPB	23 Oct 2023 11:48:54	U	SMPL
O4879-01A - 1		27521	4.2349	-	7471B	PPB	23 Oct 2023 11:51:09	U	SMPL
PB156553BL - 1	PBS	-12	-0.0373	-	7471B	PPB	23 Oct 2023 11:57:39	U	SMPL
CCV62 - 1	CCV62	33040	5.0912	-	7471B	PPB	23 Oct 2023 12:00:38	U	SMPL
CCB62 - 1	CCB62	-148	-0.0584	-	7471B	PPB	23 Oct 2023 12:02:54	U	SMPL