

7470A

INSTRUMENT ID: CV1

Linear

$$\mu \text{ Abs.} :$$

49471

A= 0.0000e+000

B= 2.0351e-004

C= -2.7076e-003

Rho= 0.9998873

Accept=Accepted

Max. Conc:— 10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	avg
0.0	0.000	0.006	0.006	43	0.000	43	0				43
0.2	0.200	0.208	0.008	1035	0.0 %	1035					1035
2.5	2.500	2.497	-0.003	12285	0.0 %	12285					12285
5.0	5.000	5.036	0.036	24761	0.0 %	24761					24761
7.5	7.500	7.387	-0.113	36312	0.0 %	36312					36312
10.0	10.000	10.065	0.065	49471	0.0 %	49471					49471

LB127933 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 23 Oct 2023 12:44:54

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	43	-	0.0000	7470A	PPB	23 Oct 2023 13:38:34	S	Std
0.2 - 1	50.2	1035	-	0.2000	7470A	PPB	23 Oct 2023 13:40:50	S	Std
2.5 - 1	52	12285	-	2.5000	7470A	PPB	23 Oct 2023 13:43:08	S	Std
5.0 - 1	55	24761	-	5.0000	7470A	PPB	23 Oct 2023 13:45:25	S	Std
7.5 - 1	57.5	36312	-	7.5000	7470A	PPB	23 Oct 2023 13:47:41	S	Std
10.0 - 1	510	49471	-	10.0000	7470A	PPB	23 Oct 2023 13:49:58	S	Std
ICV32 - 1	ICV32	18148	3.6906	-	7470A	PPB	23 Oct 2023 13:52:45	U	SMPL
ICB32 - 1	ICB32	-382	-0.0804	-	7470A	PPB	23 Oct 2023 13:55:00	U	SMPL
CCV63 - 1	CCV63	23459	4.7714	-	7470A	PPB	23 Oct 2023 13:57:17	U	SMPL
CCB63 - 1	CCB63	-380	-0.0800	-	7470A	PPB	23 Oct 2023 13:59:32	U	SMPL
CRA - 1	CRA	927	0.1859	-	7470A	PPB	23 Oct 2023 14:01:49	U	SMPL
HighStd - 1	HighStd	48693	9.9068	-	7470A	PPB	23 Oct 2023 14:04:05	U	SMPL
ChkStd - 1	ChkStd	34120	6.9410	-	7470A	PPB	23 Oct 2023 14:06:20	U	SMPL
PB156567BS - 1	LCSW	19182	3.9010	-	7470A	PPB	23 Oct 2023 14:10:55	U	SMPL
O4960-02 - 1	VAUX-SLUDGE	-367	-0.0774	-	7470A	PPB	23 Oct 2023 14:13:11	U	SMPL
O4976-03 - 1	NB-426-COMP-01	-119	-0.0269	-	7470A	PPB	23 Oct 2023 14:15:29	U	SMPL
O4977-03 - 1	NB-406-COMP-01	-77	-0.0184	-	7470A	PPB	23 Oct 2023 14:17:46	U	SMPL
O4977-09 - 1	NB-406-COMP-02	-12	-0.0051	-	7470A	PPB	23 Oct 2023 14:20:03	U	SMPL
O4978-03 - 1	NB-406-COMP-03	-24	-0.0076	-	7470A	PPB	23 Oct 2023 14:22:18	U	SMPL
PB156567BL - 1	PBW	-44	-0.0117	-	7470A	PPB	23 Oct 2023 14:24:51	U	SMPL
CCV64 - 1	CCV64	23443	4.7682	-	7470A	PPB	23 Oct 2023 14:27:25	U	SMPL
CCB64 - 1	CCB64	-332	-0.0703	-	7470A	PPB	23 Oct 2023 14:29:39	U	SMPL
O4989-02 - 1	TRENCHWAY-SOIL-PILE	-93	-0.0216	-	7470A	PPB	23 Oct 2023 14:31:56	U	SMPL
O4989-02DUP - 1	TRENCHWAY-SOIL-PILEDUP	-39	-0.0106	-	7470A	PPB	23 Oct 2023 14:34:12	U	SMPL
O4989-02MS - 1	TRENCHWAY-SOIL-PILEMS	19520	3.9698	-	7470A	PPB	23 Oct 2023 14:36:27	U	SMPL
O4989-02MSD - 1	TRENCHWAY-SOIL-PILEMSD	21905	4.4552	-	7470A	PPB	23 Oct 2023 14:38:42	U	SMPL
PB156582BL - 1	PBW	-355	-0.0750	-	7470A	PPB	23 Oct 2023 14:40:59	U	SMPL
PB156582BS - 1	LCSW	19686	4.0036	-	7470A	PPB	23 Oct 2023 14:43:17	U	SMPL
O4917-11 - 1	LSS-B2-TW-101623	-337	-0.0713	-	7470A	PPB	23 Oct 2023 14:45:34	U	SMPL
O4917-12 - 1	LSS-B4-TW-101623	-128	-0.0288	-	7470A	PPB	23 Oct 2023 14:47:52	U	SMPL
O4936-04 - 1	1030	1208	0.2431	-	7470A	PPB	23 Oct 2023 14:50:09	U	SMPL
O4936-05 - 1	1695	3777	0.7659	-	7470A	PPB	23 Oct 2023 14:52:24	U	SMPL
CCV65 - 1	CCV65	22489	4.5740	-	7470A	PPB	23 Oct 2023 14:54:41	U	SMPL
CCB65 - 1	CCB65	-220	-0.0475	-	7470A	PPB	23 Oct 2023 14:56:59	U	SMPL
O4957-01 - 1	MH-E	292	0.0567	-	7470A	PPB	23 Oct 2023 14:59:15	U	SMPL
O4957-02 - 1	MH-F	636	0.1267	-	7470A	PPB	23 Oct 2023 15:01:30	U	SMPL
O4957-03 - 1	MH-G	379	0.0744	-	7470A	PPB	23 Oct 2023 15:03:45	U	SMPL
O4957-04 - 1	MH-H	12540	2.5493	-	7470A	PPB	23 Oct 2023 15:06:00	U	SMPL
O4957-04DUP - 1	MH-HDUP	12334	2.5074	-	7470A	PPB	23 Oct 2023 15:08:16	U	SMPL
O4957-04MS - 1	MH-HMS	30102	6.1233	-	7470A	PPB	23 Oct 2023 15:10:34	U	SMPL
O4957-04MSD - 1	MH-HMSD	30045	6.1117	-	7470A	PPB	23 Oct 2023 15:12:51	U	SMPL
O4991-01 - 1	1021-1024	2345	0.4745	-	7470A	PPB	23 Oct 2023 15:15:09	U	SMPL
CCV66 - 1	CCV66	23002	4.6784	-	7470A	PPB	23 Oct 2023 15:17:28	U	SMPL
CCB66 - 1	CCB66	-168	-0.0369	-	7470A	PPB	23 Oct 2023 15:19:44	U	SMPL
PB156507TB - 1		-16	-0.0060	-	7470A	PPB	23 Oct 2023 15:22:00	U	SMPL
O4989-02LX5 - 1		14	0.0001	-	7470A	PPB	23 Oct 2023 15:24:16	U	SMPL
O4989-02A - 1		18402	3.7423	-	7470A	PPB	23 Oct 2023 15:26:31	U	SMPL
O4957-04LX5 - 1		2882	0.5838	-	7470A	PPB	23 Oct 2023 15:28:46	U	SMPL
O4957-04A - 1		30603	6.2253	-	7470A	PPB	23 Oct 2023 15:31:01	U	SMPL
CCV67 - 1	CCV67	22926	4.6630	-	7470A	PPB	23 Oct 2023 15:35:54	U	SMPL
CCB67 - 1	CCB67	-291	-0.0619	-	7470A	PPB	23 Oct 2023 15:38:09	U	SMPL