



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB127940

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	10/23/2023	16:10
2	CAL2	1	Water	NA	NA	20.1	7.00	10/23/2023	16:11
3	CAL3	1	Water	NA	NA	20.3	10.02	10/23/2023	16:15
4	ICV	1	Water	NA	NA	20.4	7.01	10/23/2023	16:20
5	CCV1	1	Water	NA	NA	20.6	2.02	10/23/2023	16:25
6	O5020-01	1	Water	NA	NA	20.1	5.13	10/23/2023	16:33
7	O5020-01DUP	1	Water	NA	NA	20.2	5.14	10/23/2023	16:34
8	O5020-02	1	Water	NA	NA	20.2	4.82	10/23/2023	16:38
9	O5020-03	1	Water	NA	NA	20.2	5.11	10/23/2023	16:42
10	CCV2	1	Water	NA	NA	20.3	12.02	10/23/2023	16:45

VB 12-2940

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph 05021

WorkList ID : 174855

Department : Wet-Chemistry

Date : 10-23-2023 15:44:32

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5020-01	CELL-A	Water	pH	Cool 4 deg C	STER09	I41	10/16/2023	9040C
O5020-02	CELL-B/C	Water	pH	Cool 4 deg C	STER09	I41	10/16/2023	9040C
O5020-03	CELL-D	Water	pH	Cool 4 deg C	STER09	I41	10/16/2023	9040C

Date/Time 10/23/23 16:00

Raw Sample Received by: JH/ACC

Raw Sample Relinquished by: J.C. (un)

Date/Time 10/23/23

Raw Sample Received by: J.C. (un)

Raw Sample Relinquished by: JH/ACC

17:30

J.C. (un)
JH/ACC