



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB127943

Slope : 97.3

BalanceID: WC SC-6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	10/24/2023	08:25
2	CAL2	1	Water	NA	NA	20.2	7.00	10/24/2023	08:26
3	CAL3	1	Water	NA	NA	20.3	10.02	10/24/2023	08:30
4	ICV	1	Water	NA	NA	20.3	7.02	10/24/2023	08:35
5	CCV1	1	Water	NA	NA	20.2	2.01	10/24/2023	08:40
6	O5012-01	1	Solid	20.02	20	20.2	7.95	10/24/2023	08:45
7	O5012-01DUP	1	Solid	20.03	20	20.3	7.96	10/24/2023	08:46
8	CCV2	1	Water	NA	NA	20.5	12.02	10/24/2023	08:50

W3127943

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph 05012 WorkList ID : 174868 Department : Wet-Chemistry Date : 10-24-2023 08:17:57

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5012-01 E	BU-03-10232023	Solid	pH	Cool 4 deg C	PSEG05	E21	10/23/2023	9045D

Date/Time 10/24/23 08:20
Raw Sample Received by: SP CAC
Raw Sample Relinquished by: SP CAC

Date/Time 10/24/23 10:10:00
Raw Sample Received by: SP CAC
Raw Sample Relinquished by: SP CAC