


LB127945

7471B

/ INSTRUMENT ID: CY1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	4.1D
0.0	0.000	-0.021	-0.021	149	0.000	149					
0.2	0.200	0.186	-0.014	1544	0.0 %	1544					
2.5	2.500	2.481	-0.019	16981	0.0 %	16981					
5.0	5.000	5.120	0.120	34739	0.0 %	34739					
7.5	7.500	7.463	-0.037	50498	0.0 %	50498					
10.0	10.000	9.971	-0.029	67375	0.0 %	67375					

LB127945 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 24 Oct 2023 10:22:14

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SO	149	-	0.0000	7471B	PPB	24 Oct 2023 10:28:45	S	Std
0.2 - 1	51.2	1544	-	0.2000	7471B	PPB	24 Oct 2023 10:31:03	S	Std
2.5 - 1	52.5	16981	-	2.5000	7471B	PPB	24 Oct 2023 10:33:20	S	Std
5.0 - 1	59	34739	-	5.0000	7471B	PPB	24 Oct 2023 10:35:37	S	Std
7.5 - 1	57.5	50498	-	7.5000	7471B	PPB	24 Oct 2023 10:37:54	S	Std
10.0 - 1	510	67375	-	10.0000	7471B	PPB	24 Oct 2023 10:40:11	S	Std
ICV33 - 1	ICV33	26842	3.9465	-	7471B	PPB	24 Oct 2023 10:43:01	U	SMPL
ICB33 - 1	ICB33	-87	-0.0562	-	7471B	PPB	24 Oct 2023 10:47:52	U	SMPL
CCV68 - 1	CCV68	35214	5.1909	-	7471B	PPB	24 Oct 2023 10:50:07	U	SMPL
CCB68 - 1	CCB68	142	-0.0221	-	7471B	PPB	24 Oct 2023 10:54:57	U	SMPL
CRA - 1	CRA	1514	0.1818	-	7471B	PPB	24 Oct 2023 11:06:16	U	SMPL
HighStd - 1	HighStd	66701	9.8710	-	7471B	PPB	24 Oct 2023 11:08:30	U	SMPL
ChkStd - 1	ChkStd	47312	6.9891	-	7471B	PPB	24 Oct 2023 11:10:46	U	SMPL
PB156603BL - 1	PB156603BL	222	-0.0103	-	7471B	PPB	24 Oct 2023 11:16:05	U	SMPL
PB156603BS - 1	PB156603BS	29107	4.2831	-	7471B	PPB	24 Oct 2023 11:18:21	U	SMPL
O4976-01 - 1	NB-426-COMP-01	7639	1.0922	-	7471B	PPB	24 Oct 2023 11:20:37	U	SMPL
O4977-01 - 1	NB-406-COMP-01	4842	0.6764	-	7471B	PPB	24 Oct 2023 11:22:55	U	SMPL
O4977-07 - 1	NB-406-COMP-02	4071	0.5618	-	7471B	PPB	24 Oct 2023 11:25:13	U	SMPL
O4978-01 - 1	NB-406-COMP-03	1806	0.2252	-	7471B	PPB	24 Oct 2023 11:27:33	U	SMPL
O4984-02 - 1	WASTE	3868	0.5317	-	7471B	PPB	24 Oct 2023 11:29:48	U	SMPL
CCV69 - 1	CCV69	34153	5.0332	-	7471B	PPB	24 Oct 2023 11:32:04	U	SMPL
CCB69 - 1	CCB69	-106	-0.0590	-	7471B	PPB	24 Oct 2023 11:34:19	U	SMPL
O4989-01 - 1	TRENCHWAY-SOIL-PILE	1786	0.2222	-	7471B	PPB	24 Oct 2023 11:36:35	U	SMPL
O5001-01 - 1	486-WC-COMP	5385	0.7572	-	7471B	PPB	24 Oct 2023 11:38:50	U	SMPL
O5005-01 - 1	N-1(0-5)	28743	4.2290	-	7471B	PPB	24 Oct 2023 11:41:06	U	SMPL
O5005-02 - 1	N-4(0-5)	75168	11.1295	-	7471B	PPB	24 Oct 2023 11:43:23	U	SMPL
O5005-03 - 1	N-2(5-10)	3606	0.4927	-	7471B	PPB	24 Oct 2023 11:45:40	U	SMPL
O5005-04 - 1	N-3(0-5)	35614	5.2503	-	7471B	PPB	24 Oct 2023 11:47:58	U	SMPL
O5010-01 - 1	343	2034	0.2591	-	7471B	PPB	24 Oct 2023 11:50:14	U	SMPL
O5010-03 - 1	72-11985	4163	0.5755	-	7471B	PPB	24 Oct 2023 11:52:32	U	SMPL
O5011-01 - 1	RBR-200027	2486	0.3263	-	7471B	PPB	24 Oct 2023 11:54:50	U	SMPL
O5011-03 - 1	RBR-200010	1947	0.2461	-	7471B	PPB	24 Oct 2023 11:57:06	U	SMPL
CCV70 - 1	CCV70	35293	5.2026	-	7471B	PPB	24 Oct 2023 11:59:21	U	SMPL
CCB70 - 1	CCB70	-103	-0.0586	-	7471B	PPB	24 Oct 2023 12:01:37	U	SMPL
O5011-05 - 1	RBR-24313	10659	1.5411	-	7471B	PPB	24 Oct 2023 12:03:52	U	SMPL
O5012-01 - 1	BU-03-10232023	2250	0.2912	-	7471B	PPB	24 Oct 2023 12:06:08	U	SMPL
O5015-02 - 1	72-11991	2896	0.3872	-	7471B	PPB	24 Oct 2023 12:08:23	U	SMPL
O5021-02 - 1	350	11048	1.5989	-	7471B	PPB	24 Oct 2023 12:15:46	U	SMPL
O5021-02DUP - 1	350DUP	10456	1.5109	-	7471B	PPB	24 Oct 2023 12:18:01	U	SMPL
O5021-02MS - 1	350MS	35850	5.2854	-	7471B	PPB	24 Oct 2023 12:20:17	U	SMPL
O5021-02MSD - 1	350MSD	40307	5.9479	-	7471B	PPB	24 Oct 2023 12:22:33	U	SMPL
O5022-01 - 1	OILY-DEBRIS	-40	-0.0492	-	7471B	PPB	24 Oct 2023 12:24:50	U	SMPL
O5025-01 - 1	EO-01-10232023-A	8785	1.2625	-	7471B	PPB	24 Oct 2023 12:27:08	U	SMPL
CCV71 - 1	CCV71	37194	5.4852	-	7471B	PPB	24 Oct 2023 12:29:24	U	SMPL
CCB71 - 1	CCB71	22	-0.0400	-	7471B	PPB	24 Oct 2023 12:31:42	U	SMPL
O5021-02LX5 - 1		2339	0.3044	-	7471B	PPB	24 Oct 2023 12:33:59	U	SMPL
O5021-02A - 1		38065	5.6146	-	7471B	PPB	24 Oct 2023 12:36:15	U	SMPL
O5005-02DLX2 - 1		38889	5.7371	-	7471B	PPB	24 Oct 2023 12:38:31	U	SMPL
CCV72 - 1	CCV72	36402	5.3674	-	7471B	PPB	24 Oct 2023 12:40:47	U	SMPL
CCB72 - 1	CCB72	-49	-0.0505	-	7471B	PPB	24 Oct 2023 12:43:02	U	SMPL